

TYPE OF TEST: ISO 19798

When cartridges HP C9730A, HP C9731A, HP C9732A and HP C9733A are used in printers HP Color LaserJet 5500 and HP Color LaserJet 5550.

Toner Cartridge yield

Average Continuous Black Cartridge Yield 13500 standard pages
Average Continuous CMY Cartridge Yield 13200 standard pages

Declared Yield value in accordance with ISO/IEC 19798

90% Lower Confidence	Black Cartridge	13547 standard pages
90% Lower Confidence	Cyan Cartridge	13201 standard pages
90% Lower Confidence	Yellow Cartridge	14848 standard pages
90% Lower Confidence	Magenta Cartridge	14261 standard pages

Date Tested:

2014/2/20 - 2014/3/18

For questions concerning testing contact

Pelikan Hardcopy CZ s.r.o
Svatoborska 395
697 01, Kyjov
Czech Republic
+420 518 699 888

Number of Cartridges used in testing:

K = 10, C = 11, M = 9, Y = 9

Number of Cartridges used in calculations

K = 9, C = 9, M = 9, Y = 9

Type of Cartridge

HP C9730A, HP C9731A, HP C9732A, HP C9733A

Shake Procedure Used

No

Print mode:

Continuous

Number of engines used in testing:

3

Media Used:

Navigator Universal 80g/m²

Paper Size:

A4

Paper feed orientation:

Long edge feed

Application Software:

Adobe Reader 11.0

Test Page Version

ISO-IEC 24712 2007 Test Pages

Power (off/on) everyday?

Yes

Computer stations and printers:

	Station 1.	Station 2.	Station 3.
Computer Model:	Intel Pentium 4 - 2,4GHz	Intel Core Duo T5870 2,0GHz	Intel Core Duo T5870 2,0GHz
Operating System:	Windows XP Professional SP3	Windows XP Professional SP3	Windows XP Professional SP3
Printer Nr.	62686	62865	62864
Printer Type	HP Color LaserJet 5500	HP Color LaserJet 5500	HP Color LaserJet 5550
Serial Nr.	JPGF000317	JPSN81KKG0C	JPFN57PG08
Driver Version:	PCL6 - 61.071.661.41	PCL6 - 61.071.661.41	PCL6 - 61.071.661.41

Note:

Page yield of black cartridge rounded to 13500. It's 500 pages more than OEM.

Page yield of color cartridges rounded to 13200. It's 1200 pages more than OEM.



TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

Cartridges testing data:

Station 1	Color	Cartidge No.	Internal summary	Cartridge yield	Cartridge used in calculation
1st set	K	0002276238 - K001	PHI 14 I0091	14 934	Y
1st set	C	0002353302 - C001	PHI 14 I0101	14 046	Y
1st set	Y	0003008306 - Y001	PHI 14 I0121	17 839	Y
1st set	M	0002165249 - M001	PHI 14 I0112	17 256	Y
2nd set	K	0002163233 - K002	PHI 14 I0092	13 146	Y
2nd set	C	0003078315 - C002	PHI 14 I0102	14 524	Y
2nd set	Y	0003008306 - Y002	PHI 14 I0122	17 206	Y
2nd set	M	0002165249 - M002	PHI 14 I0113	17 414	Y
3rd set	K	0002276238 - K003	PHI 14 I0093	15 592	Y
3rd set	C	0002353302 - C003	PHI 14 I0103	11 889	Y
3rd set	Y	0003008306 - Y003	PHI 14 I0123	17 173	Y
3rd set	M	0002022215 - M003	PHI 14 I0114	14 766	Y
	K				
	C				
	Y				
	M				
	K				
	C				
	Y				
	M				

Color	Total yield	Avg yield	St Dev	90%LCB	Number cartridges used in calculation
K	43 672	14 557	1266	12 423	3
C	40 459	13 486	1404	11 120	3
Y	52 218	17 406	375	16 773	3
M	49 436	16 479	1485	13 975	3

Station 2	Color	Cartidge No.	Internal summary	Cartridge yield	Cartridge used in calculation
1st set	K	0002276238 - K004	PHI 14 I0094	13 900	Y
1st set	C	0002353302 - C004	PHI 14 I0104	5 500	N
1st set	Y	0002354302 - Y004	PHI 14 I0124	14 150	Y
1st set	M	0002165249 - M004	PHI 14 I0115	16 305	Y
2nd set	K	0002276238 - K005	PHI 14 I0095	15 123	Y
2nd set	C	0002353302 - C005	PHI 14 I0105	15 455	Y
2nd set	Y	0002354302 - Y005	PHI 14 I0125	14 365	Y
2nd set	M	0002165249 - M005	PHI 14 I0116	17 233	Y
3rd set	K	0002276238 - K006	PHI 14 I0096	15 494	Y
3rd set	C	0002353302 - C006	PHI 14 I0106	13 581	Y
3rd set	Y	0003008307 - Y006	PHI 14 I0126	18 035	Y
3rd set	M	0002022215 - M006	PHI 14 I0117	13 480	Y
4rd set	C	0003078315 - C007	PHI 14 I0107	13 542	Y
	C				
	Y				
	M				
	K				
	C				
	Y				
	M				

Color	Total yield	Avg yield	St Dev	90%LCB	Number cartridges used in calculation
K	44 517	14 839	834	13 433	3
C	42 578	14 193	1093	12 349	3
Y	46 550	15 517	2184	11 835	3
M	47 018	15 673	1955	12 377	3

Station 3	Color	Cartidge No.	Internal summary	Cartridge yield	Cartridge used in calculation
1st set	K	0002163233 - K007	PHI 14 I0097	13 080	Y
1st set	C	0003078315 - C008	PHI 14 I0108	14 286	Y
1st set	Y	0003008306 - Y007	PHI 14 I0127	16 153	Y
1st set	M	0002022215 - M007	PHI 14 I0118	12 684	Y
2nd set	K	0002163233 - K008	PHI 14 I0098	13 642	N
2nd set	C	0003078315 - C009	PHI 14 I0109	4 214	N
2nd set	Y	0002354302 - Y008	PHI 14 I0128	12 506	Y
2nd set	M	0002165249 - M008	PHI 14 I0119	17 190	Y
3rd set	K	0002163233 - K009	PHI 14 I0099	13 232	Y
3rd set	C	0003078315 - C010	PHI 14 I0110	12 757	Y
3rd set	Y	0003008307 - Y009	PHI 14 I0129	16 917	Y
3rd set	M	0002022215 - M009	PHI 14 I0120	13 099	Y
4rd set	K	0003009306 - K010	PHI 14 I0100	13 367	Y
4rd set	C	0003078313 - C011	PHI 14 I0111	14 710	Y
	Y				
	M				
	K				
	C				
	Y				
	M				

Color	Total yield	Avg yield	St Dev	90%LCB	Number cartridges used in calculation
K	39 679	13 226	144	12 984	3
C	41 753	13 918	1027	12 186	3
Y	45 576	15 192	2357	11 218	3
M	42 973	14 324	2490	10 126	3

Summary	Color	Cartidges No.	Total	Avg	St Dev	90%LCB	Number cartridges
	K	K001 - K010	127 868	14 208	1066	13 547	9
	C	C001 - C009	124 790	13 866	1073	13 201	9
	Y	Y001 - Y009	144 344	16 038	1921	14 848	9
	M	M001 - M009	139 427	15 492	1986	14 261	9

Note:

Statement Black OEM 13 000 pages.
Statement Cyan OEM 12 000 pages.
Statement Yellow OEM 12 000 pages.
Statement Magenta OEM 12 000 pages.

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

Temperature / Humidity

Thermometer hygrometer data logger
Measuring device Nr.

Comet S3120
12932752 - 12

Testing Room average temp. [°C]	22,3	Allowed 23
Testing Room - max. running average temp. [°C]	25,6	26
Testing Room - min. running average temp. [°C]	20,1	20

Testing Room average RH. [%]	50,6	Allowed 50
Testing Room - max. running average RH [%]	64,5	65
Testing Room - min. running average RH [%]	37,1	35

Note:

During tests was temperatures and humidity in accordance with the international standard ISO/IEC 19798.
Condition 8 hours before the start of the tests.

2014/2/20 - 2014/3/18

Date of testing: 19.2.2014 6:15 - 18.3.2014 11:00

Station 1	Station 2	Station 3	Date	Time	Testing Room average temp.	Testing Room average RH	Testing Room running average temp.	Testing Room running average RH	Value included in the test interval
					15-minute interval	15-minute interval			
N/A	N/A		19.2.2014 6:15	6:15:00	21,8	56,7	N/A	N/A	No
			19.2.2014 6:30	6:30:00	21,8	55,4	N/A	N/A	No
			19.2.2014 6:45	6:45:00	21,8	55,8	N/A	N/A	No
			19.2.2014 7:00	7:00:00	22	57,6	21,9	56,4	No
			19.2.2014 7:15	7:15:00	22,1	54,4	21,9	55,8	No
			19.2.2014 7:30	7:30:00	22,3	56,9	22,1	56,2	No
			19.2.2014 7:45	7:45:00	22,5	56	22,2	56,2	No
			19.2.2014 8:00	8:00:00	22,9	55,7	22,5	55,8	No
			19.2.2014 8:15	8:15:00	23,2	52,4	22,7	55,3	No
			19.2.2014 8:30	8:30:00	23,5	52,3	23,0	54,1	No
			19.2.2014 8:45	8:45:00	23,8	52,4	23,4	53,2	No
			19.2.2014 9:00	9:00:00	24	50,7	23,6	52,0	No
			19.2.2014 9:15	9:15:00	24	50,6	23,8	51,5	No
			19.2.2014 9:30	9:30:00	24,1	48,5	24,0	50,6	No
			19.2.2014 9:45	9:45:00	24,1	46,9	24,1	49,2	No
			19.2.2014 10:00	10:00:00	24,2	47,6	24,1	48,4	No
			19.2.2014 10:15	10:15:00	24,4	46,4	24,2	47,4	No
			19.2.2014 10:30	10:30:00	24,4	46	24,3	46,7	No
			19.2.2014 10:45	10:45:00	24,4	45,1	24,4	46,3	No
			19.2.2014 11:00	11:00:00	24,5	46,3	24,4	46,0	No
			19.2.2014 11:15	11:15:00	24,7	46,1	24,5	45,9	No
			19.2.2014 11:30	11:30:00	24,9	46,9	24,6	46,1	No
			19.2.2014 11:45	11:45:00	25,1	47,8	24,8	46,8	No
			19.2.2014 12:00	12:00:00	25,3	48	25,0	47,2	No
			19.2.2014 12:15	12:15:00	25,4	45,3	25,2	47,0	No
			19.2.2014 12:30	12:30:00	25,3	44	25,3	46,3	No
			19.2.2014 12:45	12:45:00	25,3	43,7	25,3	45,3	No
			19.2.2014 13:00	13:00:00	25,4	45,1	25,4	44,5	No
			19.2.2014 13:15	13:15:00	25,4	44,7	25,4	44,4	No
			19.2.2014 13:30	13:30:00	25,3	40,9	25,4	43,6	No
			19.2.2014 13:45	13:45:00	25,3	43,2	25,4	43,5	No
			19.2.2014 14:00	14:00:00	25,5	44,3	25,4	43,3	No
			19.2.2014 14:15	14:15:00	25,4	42,8	25,4	42,8	No
			19.2.2014 14:30	14:30:00	25,3	43,2	25,4	43,4	No
			19.2.2014 14:45	14:45:00	25,1	42,6	25,3	43,2	No
			19.2.2014 15:00	15:00:00	25,2	43,1	25,3	42,9	No
			19.2.2014 15:15	15:15:00	25,3	41,4	25,2	42,6	No
			19.2.2014 15:30	15:30:00	25,2	40,5	25,2	41,9	No
			19.2.2014 15:45	15:45:00	25	40,8	25,2	41,5	No
			19.2.2014 16:00	16:00:00	24,8	40,7	25,1	40,9	No
			19.2.2014 16:15	16:15:00	24,5	39,7	24,9	40,4	No
			19.2.2014 16:30	16:30:00	24	39,9	24,6	40,3	No
			19.2.2014 16:45	16:45:00	23,6	41,3	24,2	40,4	No
			19.2.2014 17:00	17:00:00	23,2	39,4	23,8	40,1	No
			19.2.2014 17:15	17:15:00	23	43,4	23,5	41,0	No
			19.2.2014 17:30	17:30:00	22,7	45,2	23,1	42,3	No
			19.2.2014 17:45	17:45:00	22,7	46,2	22,9	43,6	No
			19.2.2014 18:00	18:00:00	22,4	47,6	22,7	45,6	No
			19.2.2014 18:15	18:15:00	22,5	52,1	22,6	47,8	No
			19.2.2014 18:30	18:30:00	22,4	47,8	22,5	48,4	No
			19.2.2014 18:45	18:45:00	22,2	52,2	22,4	49,9	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_0091 PHI_14_0101 PHI_14_0112 PHI_14_0121	PHI_14_0094 PHI_14_0104 PHI_14_0115 PHI_14_0124	19.2.2014 19:00	19:00:00	22,1	47,5	22,3	49,9	No
		19.2.2014 19:15	19:15:00	22	52,7	22,2	50,1	No
		19.2.2014 19:30	19:30:00	21,9	51,4	22,1	51,0	No
		19.2.2014 19:45	19:45:00	21,9	53,4	22,0	51,3	No
		19.2.2014 20:00	20:00:00	22	55,5	22,0	53,3	No
		19.2.2014 20:15	20:15:00	21,8	53,7	21,9	53,5	No
		19.2.2014 20:30	20:30:00	21,9	54,6	21,9	54,3	No
		19.2.2014 20:45	20:45:00	21,9	51,3	21,9	53,8	No
		19.2.2014 21:00	21:00:00	21,7	54,4	21,8	53,5	No
		19.2.2014 21:15	21:15:00	21,8	55,5	21,8	54,0	No
		19.2.2014 21:30	21:30:00	21,6	53,6	21,8	53,7	No
		19.2.2014 21:45	21:45:00	21,7	54,9	21,7	54,6	No
		19.2.2014 22:00	22:00:00	21,7	55,1	21,7	54,8	No
		19.2.2014 22:15	22:15:00	21,8	55,1	21,7	54,7	Yes
		19.2.2014 22:30	22:30:00	21,6	54,3	21,7	54,9	Yes
		19.2.2014 22:45	22:45:00	21,6	55,1	21,7	54,9	Yes
		19.2.2014 23:00	23:00:00	21,7	55	21,7	54,9	Yes
		19.2.2014 23:15	23:15:00	21,7	56	21,7	55,1	Yes
		19.2.2014 23:30	23:30:00	21,7	55,4	21,7	55,4	Yes
		19.2.2014 23:45	23:45:00	21,7	53,7	21,7	55,0	Yes
		20.2.2014 0:00	0:00:00	21,7	54,6	21,7	54,9	Yes
		20.2.2014 0:15	0:15:00	21,7	53,5	21,7	54,3	Yes
		20.2.2014 0:30	0:30:00	21,7	53	21,7	53,7	Yes
		20.2.2014 0:45	0:45:00	21,7	54,9	21,7	54,0	Yes
		20.2.2014 1:00	1:00:00	21,7	54,7	21,7	54,0	Yes
		20.2.2014 1:15	1:15:00	21,7	54,7	21,7	54,3	Yes
		20.2.2014 1:30	1:30:00	21,7	54,1	21,7	54,6	Yes
		20.2.2014 1:45	1:45:00	21,6	56,9	21,7	55,1	Yes
		20.2.2014 2:00	2:00:00	21,7	54,7	21,7	55,1	Yes
		20.2.2014 2:15	2:15:00	21,7	55,7	21,7	55,4	Yes
		20.2.2014 2:30	2:30:00	21,6	52	21,7	54,8	Yes
		20.2.2014 2:45	2:45:00	21,5	55,6	21,6	54,5	Yes
		20.2.2014 3:00	3:00:00	21,5	57,2	21,6	55,1	Yes
		20.2.2014 3:15	3:15:00	21,6	56,2	21,6	55,3	Yes
		20.2.2014 3:30	3:30:00	21,6	55,2	21,6	56,1	Yes
		20.2.2014 3:45	3:45:00	21,6	56,1	21,6	56,2	Yes
		20.2.2014 4:00	4:00:00	21,6	56,1	21,6	55,9	Yes
		20.2.2014 4:15	4:15:00	21,6	52,5	21,6	55,0	Yes
		20.2.2014 4:30	4:30:00	21,4	55,1	21,6	55,0	Yes
		20.2.2014 4:45	4:45:00	21,6	54,6	21,6	54,6	Yes
		20.2.2014 5:00	5:00:00	21,4	53,6	21,5	54,0	Yes
		20.2.2014 5:15	5:15:00	21,4	55,2	21,5	54,6	Yes
		20.2.2014 5:30	5:30:00	21,5	55	21,5	54,6	Yes
		20.2.2014 5:45	5:45:00	21,3	54,6	21,4	54,6	Yes
		20.2.2014 6:00	6:00:00	21,4	55,2	21,4	55,0	Yes
		20.2.2014 6:15	6:15:00	21,5	54,5	21,4	54,8	Yes
		20.2.2014 6:30	6:30:00	21,3	54,7	21,4	54,8	Yes
		20.2.2014 6:45	6:45:00	21,4	55,4	21,4	55,0	Yes
		20.2.2014 7:00	7:00:00	21,6	52,1	21,5	54,2	Yes
		20.2.2014 7:15	7:15:00	21,9	53,3	21,6	53,9	Yes
		20.2.2014 7:30	7:30:00	22,2	53,7	21,8	53,6	Yes
		20.2.2014 7:45	7:45:00	22,5	51,6	22,1	52,7	Yes
		20.2.2014 8:00	8:00:00	22,8	52,4	22,4	52,8	Yes
		20.2.2014 8:15	8:15:00	23,1	52,5	22,7	52,6	Yes
		20.2.2014 8:30	8:30:00	23,4	51,4	23,0	52,0	Yes
		20.2.2014 8:45	8:45:00	23,7	51,1	23,3	51,9	Yes
		20.2.2014 9:00	9:00:00	24	51,5	23,6	51,6	Yes
		20.2.2014 9:15	9:15:00	24	42,2	23,8	49,1	Yes
		20.2.2014 9:30	9:30:00	24	44,2	23,9	47,3	Yes
		20.2.2014 9:45	9:45:00	24,2	46,5	24,1	46,1	Yes
		20.2.2014 10:00	10:00:00	24,6	47,4	24,2	45,1	Yes
		20.2.2014 10:15	10:15:00	24,8	46,6	24,4	46,2	Yes
		20.2.2014 10:30	10:30:00	25	44,6	24,7	46,3	Yes
		20.2.2014 10:45	10:45:00	25	42,9	24,9	45,4	Yes
		20.2.2014 11:00	11:00:00	25	43,4	25,0	44,4	Yes
		20.2.2014 11:15	11:15:00	24,9	41,6	25,0	43,1	Yes
		20.2.2014 11:30	11:30:00	25	43,1	25,0	42,8	Yes
		20.2.2014 11:45	11:45:00	25	41,7	25,0	42,5	Yes
		20.2.2014 12:00	12:00:00	25,2	43,1	25,0	42,4	Yes
		20.2.2014 12:15	12:15:00	25,2	40,9	25,1	42,2	Yes
		20.2.2014 12:30	12:30:00	25	38,8	25,1	41,1	Yes
		20.2.2014 12:45	12:45:00	25,1	40,6	25,1	40,9	Yes
		20.2.2014 13:00	13:00:00	25,1	40	25,1	40,1	Yes
		20.2.2014 13:15	13:15:00	25	39,6	25,1	39,8	Yes
		20.2.2014 13:30	13:30:00	25,1	40,4	25,1	40,2	Yes
		20.2.2014 13:45	13:45:00	25,5	43,5	25,2	40,9	Yes
		20.2.2014 14:00	14:00:00	25,6	42,3	25,3	41,5	Yes
		20.2.2014 14:15	14:15:00	25,5	41,7	25,4	42,0	Yes
		20.2.2014 14:30	14:30:00	25,7	42,2	25,6	42,4	Yes
		20.2.2014 14:45	14:45:00	25,6	39,9	25,6	41,5	No
		20.2.2014 15:00	15:00:00	25,5	39,4	25,6	40,8	No
		20.2.2014 15:15	15:15:00	25,6	41,2	25,6	40,7	No
		20.2.2014 15:30	15:30:00	25,8	40,8	25,6	40,3	No
		20.2.2014 15:45	15:45:00	25,5	40	25,6	40,4	No
		20.2.2014 16:00	16:00:00	25	39,3	25,5	40,3	No
		20.2.2014 16:15	16:15:00	24,6	40,2	25,2	40,1	No
		20.2.2014 16:30	16:30:00	24,2	39,9	24,8	39,9	No
		20.2.2014 16:45	16:45:00	23,9	41,5	24,4	40,2	No
		20.2.2014 17:00	17:00:00	23,6	41,7	24,1	40,8	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	20.2.2014 17:15	17:15:00	23,3	41,2	23,8	41,1	No
		20.2.2014 17:30	17:30:00	23	41,3	23,5	41,4	No
		20.2.2014 17:45	17:45:00	22,8	41,8	23,2	41,5	No
		20.2.2014 18:00	18:00:00	22,5	42,7	22,9	41,8	No
		20.2.2014 18:15	18:15:00	22,3	42	22,7	42,0	No
		20.2.2014 18:30	18:30:00	22,1	43,2	22,4	42,4	No
		20.2.2014 18:45	18:45:00	21,9	42,8	22,2	42,7	No
		20.2.2014 19:00	19:00:00	21,8	42,5	22,0	42,6	No
		20.2.2014 19:15	19:15:00	21,6	44,4	21,9	43,2	No
		20.2.2014 19:30	19:30:00	21,5	44,9	21,7	43,7	No
		20.2.2014 19:45	19:45:00	21,4	45	21,6	44,2	No
		20.2.2014 20:00	20:00:00	21,3	45,3	21,5	44,9	No
		20.2.2014 20:15	20:15:00	21,2	46	21,4	45,3	No
		20.2.2014 20:30	20:30:00	21,2	46,7	21,3	45,8	No
		20.2.2014 20:45	20:45:00	21,1	47,2	21,2	46,3	No
		20.2.2014 21:00	21:00:00	21	47,9	21,1	47,0	No
		20.2.2014 21:15	21:15:00	21	48,9	21,1	47,7	No
		20.2.2014 21:30	21:30:00	20,9	49,5	21,0	48,4	No
		20.2.2014 21:45	21:45:00	20,8	49,4	20,9	48,9	No
		20.2.2014 22:00	22:00:00	20,8	49,6	20,9	49,4	Yes
		20.2.2014 22:15	22:15:00	20,8	50,2	20,8	49,7	Yes
		20.2.2014 22:30	22:30:00	20,7	51,3	20,8	50,1	Yes
		20.2.2014 22:45	22:45:00	20,6	50,9	20,7	50,5	Yes
		20.2.2014 23:00	23:00:00	20,6	51,5	20,7	51,0	Yes
		20.2.2014 23:15	23:15:00	20,5	51,8	20,6	51,4	Yes
		20.2.2014 23:30	23:30:00	20,5	52,2	20,6	51,6	Yes
		20.2.2014 23:45	23:45:00	20,4	51,7	20,5	51,8	Yes
		21.2.2014 0:00	0:00:00	20,4	53,1	20,5	52,2	Yes
		21.2.2014 0:15	0:15:00	20,3	49,8	20,4	51,7	Yes
		21.2.2014 0:30	0:30:00	20,3	52,3	20,4	51,7	Yes
		21.2.2014 0:45	0:45:00	20,3	53,8	20,3	52,3	Yes
		21.2.2014 1:00	1:00:00	20,3	55,2	20,3	52,8	Yes
		21.2.2014 1:15	1:15:00	20,2	53,2	20,3	53,6	Yes
		21.2.2014 1:30	1:30:00	20,2	54	20,3	54,1	Yes
		21.2.2014 1:45	1:45:00	20,2	54,3	20,2	54,2	Yes
		21.2.2014 2:00	2:00:00	20,1	55,2	20,2	54,2	Yes
		21.2.2014 2:15	2:15:00	20,1	53,6	20,2	54,3	Yes
		21.2.2014 2:30	2:30:00	20,1	53,1	20,1	54,1	Yes
		21.2.2014 2:45	2:45:00	20,1	54,1	20,1	54,0	Yes
		21.2.2014 3:00	3:00:00	20,1	54,9	20,1	53,9	Yes
		21.2.2014 3:15	3:15:00	20,1	54,6	20,1	54,2	Yes
		21.2.2014 3:30	3:30:00	20,1	53,3	20,1	54,2	Yes
		21.2.2014 3:45	3:45:00	20	54,2	20,1	54,3	Yes
		21.2.2014 4:00	4:00:00	20	55,4	20,1	54,4	Yes
		21.2.2014 4:15	4:15:00	20,1	52,5	20,1	53,9	Yes
		21.2.2014 4:30	4:30:00	20,1	51,1	20,1	53,3	Yes
		21.2.2014 4:45	4:45:00	20,2	49,2	20,1	52,1	Yes
		21.2.2014 5:00	5:00:00	20,2	49,9	20,2	50,7	Yes
		21.2.2014 5:15	5:15:00	20,1	51,4	20,2	50,4	Yes
		21.2.2014 5:30	5:30:00	20,1	50	20,2	50,1	Yes
		21.2.2014 5:45	5:45:00	20,1	49,5	20,1	50,2	Yes
PHI_14_I0091 PHI_14_I0101 PHI_14_I0112 PHI_14_I0121 PHI_14_I0094 PHI_14_I0104 PHI_14_I0105 PHI_14_I0115 PHI_14_I0124		21.2.2014 6:00	6:00:00	20,1	50,6	20,1	50,4	Yes
		21.2.2014 6:15	6:15:00	20,1	50,8	20,1	50,2	Yes
		21.2.2014 6:30	6:30:00	20,1	50,3	20,1	50,3	Yes
		21.2.2014 6:45	6:45:00	20,2	49	20,1	50,2	Yes
		21.2.2014 7:00	7:00:00	20,3	49,5	20,2	49,9	Yes
		21.2.2014 7:15	7:15:00	20,6	48,2	20,3	49,3	Yes
		21.2.2014 7:30	7:30:00	20,8	48,1	20,5	48,7	Yes
		21.2.2014 7:45	7:45:00	21,1	48,4	20,7	48,6	Yes
		21.2.2014 8:00	8:00:00	21,4	48,2	21,0	48,2	Yes
		21.2.2014 8:15	8:15:00	21,8	49	21,3	48,4	Yes
		21.2.2014 8:30	8:30:00	22,2	48,8	21,6	48,6	Yes
		21.2.2014 8:45	8:45:00	22,5	47,7	22,0	48,4	Yes
		21.2.2014 9:00	9:00:00	22,8	47	22,3	48,1	Yes
		21.2.2014 9:15	9:15:00	23,2	48,5	22,7	48,0	Yes
		21.2.2014 9:30	9:30:00	23,4	47	23,0	47,6	Yes
		21.2.2014 9:45	9:45:00	23,5	44,9	23,2	46,9	Yes
		21.2.2014 10:00	10:00:00	23,6	45,4	23,4	46,5	Yes
		21.2.2014 10:15	10:15:00	23,8	45,2	23,6	45,6	Yes
		21.2.2014 10:30	10:30:00	23,9	43,4	23,7	44,7	Yes
		21.2.2014 10:45	10:45:00	24	42,7	23,8	44,2	Yes
		21.2.2014 11:00	11:00:00	24,1	44,6	24,0	44,0	Yes
		21.2.2014 11:15	11:15:00	24,3	45	24,1	43,9	Yes
		21.2.2014 11:30	11:30:00	24,4	44,3	24,2	44,2	Yes
		21.2.2014 11:45	11:45:00	24,5	44	24,3	44,5	Yes
		21.2.2014 12:00	12:00:00	24,4	40,3	24,4	43,4	Yes
		21.2.2014 12:15	12:15:00	24,2	39,9	24,4	42,1	Yes
		21.2.2014 12:30	12:30:00	23,9	36,8	24,3	40,3	Yes
		21.2.2014 12:45	12:45:00	23,8	37,6	24,1	38,7	Yes
		21.2.2014 13:00	13:00:00	23,6	38,1	23,9	38,1	Yes
		21.2.2014 13:15	13:15:00	23,5	39,1	23,7	37,9	Yes
		21.2.2014 13:30	13:30:00	23,6	41,4	23,6	39,1	Yes
		21.2.2014 13:45	13:45:00	23,7	39	23,6	39,4	Yes
		21.2.2014 14:00	14:00:00	23,7	41,1	23,6	40,2	Yes
		21.2.2014 14:15	14:15:00	23,8	40,7	23,7	40,6	Yes
		21.2.2014 14:30	14:30:00	23,7	39,5	23,7	40,1	Yes
		21.2.2014 14:45	14:45:00	23,8	40,8	23,8	40,5	No
		21.2.2014 15:00	15:00:00	23,6	37,8	23,7	39,7	No
		21.2.2014 15:15	15:15:00	23,7	40,8	23,7	39,7	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

	21.2.2014 15:30	15:30:00	23,8	40,2	23,7	39,9	No
	21.2.2014 15:45	15:45:00	23,9	38,1	23,8	39,2	No
	21.2.2014 16:00	16:00:00	23,2	37,2	23,7	39,1	No
	21.2.2014 16:15	16:15:00	22,5	34,8	23,4	37,6	No
	21.2.2014 16:30	16:30:00	22,4	43,4	23,0	38,4	No
	21.2.2014 16:45	16:45:00	21,9	37,9	22,5	38,3	No
	21.2.2014 17:00	17:00:00	21,6	42,1	22,1	39,6	No
	21.2.2014 17:15	17:15:00	21,8	45,8	21,9	42,3	No
	21.2.2014 17:30	17:30:00	22	47,2	21,8	43,3	No
	21.2.2014 17:45	17:45:00	21,9	43,6	21,8	44,7	No
	21.2.2014 18:00	18:00:00	21,8	49,5	21,9	46,5	No
	21.2.2014 18:15	18:15:00	22	49,9	21,9	47,6	No
	21.2.2014 18:30	18:30:00	22,1	49,7	22,0	48,2	No
	21.2.2014 18:45	18:45:00	22,1	50	22,0	49,8	No
	21.2.2014 19:00	19:00:00	22,2	50,5	22,1	50,0	No
	21.2.2014 19:15	19:15:00	22,2	50,2	22,2	50,1	No
	21.2.2014 19:30	19:30:00	22,2	50,8	22,2	50,4	No
	21.2.2014 19:45	19:45:00	22,2	51,4	22,2	50,7	No
	21.2.2014 20:00	20:00:00	22,2	51,8	22,2	51,1	No
	21.2.2014 20:15	20:15:00	22,2	52,2	22,2	51,6	No
	21.2.2014 20:30	20:30:00	22,2	51,7	22,2	51,8	No
	21.2.2014 20:45	20:45:00	22,2	51,8	22,2	51,9	No
	21.2.2014 21:00	21:00:00	22,2	51,7	22,2	51,9	No
	21.2.2014 21:15	21:15:00	22,1	52,1	22,2	51,8	No
	21.2.2014 21:30	21:30:00	22,1	52,8	22,2	52,1	No
	21.2.2014 21:45	21:45:00	22,1	53,2	22,1	52,5	No
	21.2.2014 22:00	22:00:00	22,1	53,7	22,1	53,0	No
	21.2.2014 22:15	22:15:00	22,1	53,1	22,1	53,2	No
	21.2.2014 22:30	22:30:00	22,1	53,8	22,1	53,5	No
	21.2.2014 22:45	22:45:00	22	54,2	22,1	53,7	No
	21.2.2014 23:00	23:00:00	22	54,7	22,1	54,0	No
	21.2.2014 23:15	23:15:00	22	54,1	22,0	54,2	No
	21.2.2014 23:30	23:30:00	22	54	22,0	54,3	No
	21.2.2014 23:45	23:45:00	21,9	54,8	22,0	54,4	No
	22.2.2014 0:00	0:00:00	21,9	54,8	22,0	54,4	No
	22.2.2014 0:15	0:15:00	21,9	54,5	21,9	54,5	No
	22.2.2014 0:30	0:30:00	21,9	55	21,9	54,8	No
	22.2.2014 0:45	0:45:00	21,9	55,6	21,9	55,0	No
	22.2.2014 1:00	1:00:00	21,9	55,6	21,9	55,2	No
	22.2.2014 1:15	1:15:00	21,9	55,6	21,9	55,5	No
	22.2.2014 1:30	1:30:00	21,8	56,4	21,9	55,8	No
	22.2.2014 1:45	1:45:00	21,8	56,6	21,9	56,1	No
	22.2.2014 2:00	2:00:00	21,8	56,5	21,8	56,3	No
	22.2.2014 2:15	2:15:00	21,8	55,6	21,8	56,3	No
	22.2.2014 2:30	2:30:00	21,8	56,4	21,8	56,3	No
	22.2.2014 2:45	2:45:00	21,8	56,6	21,8	56,3	No
	22.2.2014 3:00	3:00:00	21,8	55,7	21,8	56,1	No
	22.2.2014 3:15	3:15:00	21,7	56,3	21,8	56,3	No
	22.2.2014 3:30	3:30:00	21,7	55,6	21,8	56,1	No
	22.2.2014 3:45	3:45:00	21,7	56,5	21,7	56,0	No
	22.2.2014 4:00	4:00:00	21,7	56	21,7	56,1	No
	22.2.2014 4:15	4:15:00	21,7	56,7	21,7	56,2	No
	22.2.2014 4:30	4:30:00	21,7	56,6	21,7	56,5	No
	22.2.2014 4:45	4:45:00	21,7	56,6	21,7	56,5	No
	22.2.2014 5:00	5:00:00	21,6	56,8	21,7	56,7	No
	22.2.2014 5:15	5:15:00	21,7	56,7	21,7	56,7	No
	22.2.2014 5:30	5:30:00	21,6	52,4	21,7	55,6	No
	22.2.2014 5:45	5:45:00	21,5	56,3	21,6	55,6	No
	22.2.2014 6:00	6:00:00	21,6	56,4	21,6	55,5	No
	22.2.2014 6:15	6:15:00	21,6	53,4	21,6	54,6	No
	22.2.2014 6:30	6:30:00	21,4	56	21,5	55,5	No
	22.2.2014 6:45	6:45:00	21,5	56,4	21,5	55,6	No
	22.2.2014 7:00	7:00:00	21,6	57,1	21,5	55,7	No
	22.2.2014 7:15	7:15:00	21,4	53,7	21,5	55,8	No
	22.2.2014 7:30	7:30:00	21,4	56,6	21,5	56,0	No
	22.2.2014 7:45	7:45:00	21,5	56,9	21,5	56,1	No
	22.2.2014 8:00	8:00:00	21,4	51,8	21,4	54,8	No
	22.2.2014 8:15	8:15:00	21,4	56,9	21,4	55,6	No
	22.2.2014 8:30	8:30:00	21,5	53,5	21,5	54,8	No
	22.2.2014 8:45	8:45:00	21,4	56,4	21,4	54,7	No
	22.2.2014 9:00	9:00:00	21,3	54,6	21,4	55,4	No
	22.2.2014 9:15	9:15:00	21,4	56,6	21,4	55,3	No
	22.2.2014 9:30	9:30:00	21,6	55,9	21,4	55,9	No
	22.2.2014 9:45	9:45:00	21,5	52	21,5	54,8	No
	22.2.2014 10:00	10:00:00	21,4	57,3	21,5	55,5	No
	22.2.2014 10:15	10:15:00	21,6	56,1	21,5	55,3	No
	22.2.2014 10:30	10:30:00	21,7	54,2	21,6	54,9	No
	22.2.2014 10:45	10:45:00	21,4	57	21,5	56,2	No
	22.2.2014 11:00	11:00:00	21,6	56,3	21,6	55,9	No
	22.2.2014 11:15	11:15:00	21,7	55,8	21,6	55,8	No
	22.2.2014 11:30	11:30:00	21,4	56,7	21,5	56,5	No
	22.2.2014 11:45	11:45:00	21,6	56,3	21,6	56,3	No
	22.2.2014 12:00	12:00:00	21,4	56	21,5	56,2	No
	22.2.2014 12:15	12:15:00	21,6	56,4	21,5	56,4	No
	22.2.2014 12:30	12:30:00	21,5	52,2	21,5	55,2	No
	22.2.2014 12:45	12:45:00	21,5	56,8	21,5	55,4	No
	22.2.2014 13:00	13:00:00	21,7	55,6	21,6	55,3	No
	22.2.2014 13:15	13:15:00	21,4	56,7	21,5	55,3	No
	22.2.2014 13:30	13:30:00	21,6	55,9	21,6	56,3	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	22.2.2014 13:45	13:45:00	21.4	56.4	21.5	56.2	No
		22.2.2014 14:00	14:00:00	21.6	56	21.5	56.3	No
		22.2.2014 14:15	14:15:00	21.4	55.8	21.5	56.0	No
		22.2.2014 14:30	14:30:00	21.6	55.8	21.5	56.0	No
		22.2.2014 14:45	14:45:00	21.7	52.5	21.6	55.0	No
		22.2.2014 15:00	15:00:00	21.5	56.7	21.6	55.2	No
		22.2.2014 15:15	15:15:00	21.7	55	21.6	55.0	No
		22.2.2014 15:30	15:30:00	21.5	56.1	21.6	55.1	No
		22.2.2014 15:45	15:45:00	21.6	55.4	21.6	55.8	No
		22.2.2014 16:00	16:00:00	21.7	56.2	21.6	55.7	No
		22.2.2014 16:15	16:15:00	21.5	53.2	21.6	55.2	No
		22.2.2014 16:30	16:30:00	21.5	56.5	21.6	55.3	No
		22.2.2014 16:45	16:45:00	21.7	56.1	21.6	55.5	No
		22.2.2014 17:00	17:00:00	21.8	55.4	21.6	55.3	No
		22.2.2014 17:15	17:15:00	21.8	55.9	21.7	56.0	No
		22.2.2014 17:30	17:30:00	21.5	55.7	21.7	55.8	No
		22.2.2014 17:45	17:45:00	21.6	56	21.7	55.8	No
		22.2.2014 18:00	18:00:00	21.7	56.2	21.7	56.0	No
		22.2.2014 18:15	18:15:00	21.8	56.2	21.7	56.0	No
		22.2.2014 18:30	18:30:00	21.8	55.9	21.7	56.1	No
		22.2.2014 18:45	18:45:00	21.8	56.2	21.8	56.1	No
		22.2.2014 19:00	19:00:00	21.9	56.7	21.8	56.3	No
		22.2.2014 19:15	19:15:00	21.6	54.4	21.8	55.8	No
		22.2.2014 19:30	19:30:00	21.6	57.2	21.7	56.1	No
		22.2.2014 19:45	19:45:00	21.7	57	21.7	56.3	No
		22.2.2014 20:00	20:00:00	21.7	56.1	21.7	56.2	No
		22.2.2014 20:15	20:15:00	21.7	56.3	21.7	56.7	No
		22.2.2014 20:30	20:30:00	21.8	56.8	21.7	56.6	No
		22.2.2014 20:45	20:45:00	21.8	56	21.8	56.3	No
		22.2.2014 21:00	21:00:00	21.8	56.5	21.8	56.4	No
		22.2.2014 21:15	21:15:00	21.8	56	21.8	56.3	No
		22.2.2014 21:30	21:30:00	21.8	56.2	21.8	56.2	No
		22.2.2014 21:45	21:45:00	21.8	56.2	21.8	56.2	No
		22.2.2014 22:00	22:00:00	21.8	56.8	21.8	56.3	No
		22.2.2014 22:15	22:15:00	21.8	56.8	21.8	56.5	No
		22.2.2014 22:30	22:30:00	21.8	56.6	21.8	56.6	No
		22.2.2014 22:45	22:45:00	21.8	56.3	21.8	56.6	No
		22.2.2014 23:00	23:00:00	21.8	56.2	21.8	56.5	No
		22.2.2014 23:15	23:15:00	21.8	56.7	21.8	56.5	No
		22.2.2014 23:30	23:30:00	21.8	56.7	21.8	56.5	No
		22.2.2014 23:45	23:45:00	21.8	56.6	21.8	56.6	No
		23.2.2014 0:00	0:00:00	21.7	56.3	21.8	56.6	No
		23.2.2014 0:15	0:15:00	21.7	55.4	21.8	56.3	No
		23.2.2014 0:30	0:30:00	21.7	55.7	21.7	56.0	No
		23.2.2014 0:45	0:45:00	21.7	56.6	21.7	56.0	No
		23.2.2014 1:00	1:00:00	21.6	56.2	21.7	56.0	No
		23.2.2014 1:15	1:15:00	21.6	56.5	21.7	56.3	No
		23.2.2014 1:30	1:30:00	21.6	56.8	21.6	56.5	No
		23.2.2014 1:45	1:45:00	21.6	57	21.6	56.6	No
		23.2.2014 2:00	2:00:00	21.6	56.7	21.6	56.8	No
		23.2.2014 2:15	2:15:00	21.6	55.9	21.6	56.6	No
		23.2.2014 2:30	2:30:00	21.6	56	21.6	56.4	No
		23.2.2014 2:45	2:45:00	21.6	56.3	21.6	56.2	No
		23.2.2014 3:00	3:00:00	21.6	56.3	21.6	56.1	No
		23.2.2014 3:15	3:15:00	21.6	56.4	21.6	56.3	No
		23.2.2014 3:30	3:30:00	21.6	55.5	21.6	56.1	No
		23.2.2014 3:45	3:45:00	21.6	55.6	21.6	56.0	No
		23.2.2014 4:00	4:00:00	21.6	55.7	21.6	55.8	No
		23.2.2014 4:15	4:15:00	21.6	56	21.6	55.7	No
		23.2.2014 4:30	4:30:00	21.6	56.5	21.6	56.0	No
		23.2.2014 4:45	4:45:00	21.6	56.1	21.6	56.1	No
		23.2.2014 5:00	5:00:00	21.6	55.5	21.6	56.0	No
		23.2.2014 5:15	5:15:00	21.6	53.8	21.6	55.5	No
		23.2.2014 5:30	5:30:00	21.3	52.6	21.5	54.5	No
		23.2.2014 5:45	5:45:00	21.5	56.8	21.5	54.7	No
		23.2.2014 6:00	6:00:00	21.2	54.2	21.4	54.4	No
		23.2.2014 6:15	6:15:00	21.4	56.4	21.4	55.0	No
		23.2.2014 6:30	6:30:00	21.2	54.2	21.3	55.4	No
		23.2.2014 6:45	6:45:00	21.4	55.3	21.3	55.0	No
		23.2.2014 7:00	7:00:00	21.4	56.6	21.4	55.6	No
		23.2.2014 7:15	7:15:00	21.5	56.6	21.4	55.7	No
		23.2.2014 7:30	7:30:00	21.3	55.2	21.4	55.9	No
		23.2.2014 7:45	7:45:00	21.4	56.6	21.4	56.3	No
		23.2.2014 8:00	8:00:00	21.5	57.1	21.4	56.4	No
		23.2.2014 8:15	8:15:00	21.4	51.7	21.4	55.2	No
		23.2.2014 8:30	8:30:00	21.4	56.2	21.4	55.4	No
		23.2.2014 8:45	8:45:00	21.5	56.2	21.5	55.3	No
		23.2.2014 9:00	9:00:00	21.5	51.7	21.5	54.0	No
		23.2.2014 9:15	9:15:00	21.3	54.5	21.4	54.7	No
		23.2.2014 9:30	9:30:00	21.4	55.7	21.4	54.5	No
		23.2.2014 9:45	9:45:00	21.4	55.5	21.4	54.4	No
		23.2.2014 10:00	10:00:00	21.5	55.5	21.4	55.3	No
		23.2.2014 10:15	10:15:00	21.5	56	21.5	55.7	No
		23.2.2014 10:30	10:30:00	21.5	54.9	21.5	55.5	No
		23.2.2014 10:45	10:45:00	21.5	55.4	21.5	55.5	No
		23.2.2014 11:00	11:00:00	21.5	55.4	21.5	55.4	No
		23.2.2014 11:15	11:15:00	21.5	55.1	21.5	55.2	No
		23.2.2014 11:30	11:30:00	21.5	55.8	21.5	55.4	No
		23.2.2014 11:45	11:45:00	21.6	55	21.5	55.3	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

		23.2.2014 12:00	12:00:00	21,5	55,3	21,5	55,3	No
		23.2.2014 12:15	12:15:00	21,6	55,1	21,6	55,3	No
		23.2.2014 12:30	12:30:00	21,6	55,5	21,6	55,2	No
		23.2.2014 12:45	12:45:00	21,6	55,2	21,6	55,3	No
		23.2.2014 13:00	13:00:00	21,6	55,5	21,6	55,3	No
		23.2.2014 13:15	13:15:00	21,6	55,3	21,6	55,4	No
		23.2.2014 13:30	13:30:00	21,6	54,3	21,6	55,1	No
		23.2.2014 13:45	13:45:00	21,6	55,8	21,6	55,2	No
		23.2.2014 14:00	14:00:00	21,7	55,1	21,6	55,1	No
		23.2.2014 14:15	14:15:00	21,7	55,3	21,7	55,1	No
		23.2.2014 14:30	14:30:00	21,7	55,7	21,7	55,5	No
		23.2.2014 14:45	14:45:00	21,4	52,6	21,6	54,7	No
		23.2.2014 15:00	15:00:00	21,5	54,8	21,6	54,6	No
		23.2.2014 15:15	15:15:00	21,6	54,3	21,6	54,4	No
		23.2.2014 15:30	15:30:00	21,6	53,9	21,5	53,9	No
		23.2.2014 15:45	15:45:00	21,7	50,8	21,6	53,5	No
		23.2.2014 16:00	16:00:00	21,4	53,8	21,6	53,2	No
		23.2.2014 16:15	16:15:00	21,5	53,9	21,6	53,1	No
		23.2.2014 16:30	16:30:00	21,6	54	21,6	53,1	No
		23.2.2014 16:45	16:45:00	21,7	53,8	21,6	53,9	No
		23.2.2014 17:00	17:00:00	21,4	50,8	21,6	53,1	No
		23.2.2014 17:15	17:15:00	21,5	54,9	21,6	53,4	No
		23.2.2014 17:30	17:30:00	21,6	53,1	21,6	53,2	No
		23.2.2014 17:45	17:45:00	21,7	54,1	21,6	53,2	No
		23.2.2014 18:00	18:00:00	21,5	48,8	21,6	52,7	No
		23.2.2014 18:15	18:15:00	21,4	53,6	21,6	52,4	No
		23.2.2014 18:30	18:30:00	21,6	54	21,6	52,6	No
		23.2.2014 18:45	18:45:00	21,6	55,1	21,5	52,9	No
		23.2.2014 19:00	19:00:00	21,4	53,3	21,5	54,0	No
		23.2.2014 19:15	19:15:00	21,5	54,7	21,5	54,3	No
		23.2.2014 19:30	19:30:00	21,6	55,5	21,5	54,7	No
		23.2.2014 19:45	19:45:00	21,6	54,7	21,5	54,6	No
		23.2.2014 20:00	20:00:00	21,7	55,3	21,6	55,1	No
		23.2.2014 20:15	20:15:00	21,8	56	21,7	55,4	No
		23.2.2014 20:30	20:30:00	21,5	51,5	21,7	54,4	No
		23.2.2014 20:45	20:45:00	21,5	53,9	21,6	54,2	No
		23.2.2014 21:00	21:00:00	21,6	52,4	21,6	53,5	No
		23.2.2014 21:15	21:15:00	21,6	51,7	21,6	52,4	No
		23.2.2014 21:30	21:30:00	21,3	49,6	21,5	51,9	No
		23.2.2014 21:45	21:45:00	21,3	46,9	21,5	50,2	No
		23.2.2014 22:00	22:00:00	21,4	51,2	21,4	49,9	Yes
		23.2.2014 22:15	22:15:00	21,2	49,4	21,3	49,3	Yes
		23.2.2014 22:30	22:30:00	21,4	48,8	21,3	49,1	Yes
		23.2.2014 22:45	22:45:00	21,2	50,1	21,3	49,9	Yes
		23.2.2014 23:00	23:00:00	21,4	48,2	21,3	49,1	Yes
		23.2.2014 23:15	23:15:00	21,2	49,6	21,3	49,2	Yes
		23.2.2014 23:30	23:30:00	21,4	48,7	21,3	49,2	Yes
		23.2.2014 23:45	23:45:00	21,2	47,7	21,3	48,6	Yes
		24.2.2014 0:00	0:00:00	21,2	44,4	21,3	47,6	Yes
		24.2.2014 0:15	0:15:00	21,2	47,2	21,3	47,0	Yes
		24.2.2014 0:30	0:30:00	21,3	44,1	21,2	45,9	Yes
		24.2.2014 0:45	0:45:00	21,3	47,3	21,3	45,8	Yes
		24.2.2014 1:00	1:00:00	21,3	44,1	21,3	45,7	Yes
		24.2.2014 1:15	1:15:00	21,2	47,3	21,3	45,7	Yes
		24.2.2014 1:30	1:30:00	21,4	44,2	21,3	45,7	Yes
		24.2.2014 1:45	1:45:00	21,3	47	21,3	45,7	Yes
		24.2.2014 2:00	2:00:00	21,4	43,8	21,3	45,6	Yes
		24.2.2014 2:15	2:15:00	21,3	46,8	21,4	45,5	Yes
		24.2.2014 2:30	2:30:00	21,3	43,3	21,3	45,2	Yes
		24.2.2014 2:45	2:45:00	21,3	46,4	21,3	45,1	Yes
		24.2.2014 3:00	3:00:00	21,4	43,1	21,3	44,9	Yes
		24.2.2014 3:15	3:15:00	21,3	46,2	21,3	44,8	Yes
		24.2.2014 3:30	3:30:00	21,5	44	21,4	44,9	Yes
		24.2.2014 3:45	3:45:00	21,3	46	21,4	44,8	Yes
		24.2.2014 4:00	4:00:00	21,2	44,8	21,3	45,3	Yes
		24.2.2014 4:15	4:15:00	21,3	45,7	21,3	45,1	Yes
		24.2.2014 4:30	4:30:00	21,2	43,8	21,3	45,1	Yes
		24.2.2014 4:45	4:45:00	21,3	45,5	21,3	45,0	Yes
		24.2.2014 5:00	5:00:00	21,3	43,3	21,3	44,6	Yes
		24.2.2014 5:15	5:15:00	21,4	44,6	21,3	44,3	Yes
		24.2.2014 5:30	5:30:00	21,2	45,3	21,3	44,7	Yes
		24.2.2014 5:45	5:45:00	21,1	44,9	21,3	44,5	Yes
		24.2.2014 6:00	6:00:00	21,4	44,7	21,3	44,9	Yes
		24.2.2014 6:15	6:15:00	21,2	44,3	21,2	44,8	Yes
		24.2.2014 6:30	6:30:00	21,4	43	21,3	44,2	Yes
		24.2.2014 6:45	6:45:00	21,3	44	21,3	44,0	Yes
		24.2.2014 7:00	7:00:00	21,1	38	21,3	42,3	Yes
		24.2.2014 7:15	7:15:00	21,5	43,4	21,3	42,1	Yes
		24.2.2014 7:30	7:30:00	21,8	44,8	21,4	42,6	Yes
		24.2.2014 7:45	7:45:00	21,8	44,1	21,6	42,6	Yes
		24.2.2014 8:00	8:00:00	21,9	42,5	21,8	43,7	Yes
		24.2.2014 8:15	8:15:00	22,2	43,5	21,9	43,7	Yes
		24.2.2014 8:30	8:30:00	22,4	43,2	22,1	43,3	Yes
		24.2.2014 8:45	8:45:00	22,6	43,5	22,3	43,2	Yes
		24.2.2014 9:00	9:00:00	22,8	43,8	22,5	43,5	Yes
		24.2.2014 9:15	9:15:00	23	41	22,7	42,9	Yes
		24.2.2014 9:30	9:30:00	23	40,5	22,9	42,2	Yes
		24.2.2014 9:45	9:45:00	23,1	41,7	23,0	41,8	Yes
		24.2.2014 10:00	10:00:00	23,2	40,3	23,1	40,9	Yes

J0091

J0101

J0112

J0121

J0094

J0105

J0115

J0124

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14	24.2.2014 10:15	10:15:00	23,5	41,7	23,2	41,1	Yes
PHI_14	24.2.2014 10:30	10:30:00	23,5	39,3	23,3	40,8	Yes
PHI_14	24.2.2014 10:45	10:45:00	23,7	40,9	23,5	40,6	Yes
PHI_14	24.2.2014 11:00	11:00:00	23,6	40,1	23,6	40,5	Yes
PHI_14	24.2.2014 11:15	11:15:00	23,7	40,2	23,6	40,1	Yes
PHI_14	24.2.2014 11:30	11:30:00	23,5	38,1	23,6	39,8	Yes
PHI_14	24.2.2014 11:45	11:45:00	23,7	38,9	23,6	39,3	Yes
PHI_14	24.2.2014 12:00	12:00:00	23,8	40,8	23,7	39,5	Yes
PHI_14	24.2.2014 12:15	12:15:00	23,8	39,3	23,7	39,3	Yes
PHI_14	24.2.2014 12:30	12:30:00	23,8	37,9	23,8	39,2	Yes
PHI_14	24.2.2014 12:45	12:45:00	23,8	37,3	23,8	38,8	Yes
PHI_14	24.2.2014 13:00	13:00:00	24	40	23,9	38,6	Yes
PHI_14	24.2.2014 13:15	13:15:00	23,9	38,2	23,9	38,4	Yes
PHI_14	24.2.2014 13:30	13:30:00	23,9	38,6	23,9	38,5	Yes
PHI_14	24.2.2014 13:45	13:45:00	24,3	42,4	24,0	39,8	Yes
PHI_14	24.2.2014 14:00	14:00:00	24,7	46,1	24,2	41,3	Yes
PHI_14	24.2.2014 14:15	14:15:00	24,7	43,9	24,4	42,8	Yes
PHI_14	24.2.2014 14:30	14:30:00	24,6	42,4	24,6	43,7	Yes
PHI_14	24.2.2014 14:45	14:45:00	24,6	41,2	24,7	43,4	No
PHI_14	24.2.2014 15:00	15:00:00	24,5	41,6	24,6	42,3	No
PHI_14	24.2.2014 15:15	15:15:00	24,4	41,3	24,5	41,6	No
PHI_14	24.2.2014 15:30	15:30:00	24,4	40,9	24,5	41,3	No
PHI_14	24.2.2014 15:45	15:45:00	23,9	37,3	24,3	40,3	No
PHI_14	24.2.2014 16:00	16:00:00	23,3	36,5	24,0	39,0	No
PHI_14	24.2.2014 16:15	16:15:00	22,8	36,3	23,6	37,8	No
PHI_14	24.2.2014 16:30	16:30:00	22,3	37	23,1	36,8	No
PHI_14	24.2.2014 16:45	16:45:00	22,1	39,7	22,6	37,4	No
PHI_14	24.2.2014 17:00	17:00:00	22,2	42,4	22,4	38,9	No
PHI_14	24.2.2014 17:15	17:15:00	21,8	39,2	22,1	39,6	No
PHI_14	24.2.2014 17:30	17:30:00	21,8	45,2	22,0	41,6	No
PHI_14	24.2.2014 17:45	17:45:00	22	46,2	22,0	43,3	No
PHI_14	24.2.2014 18:00	18:00:00	21,5	40,5	21,8	42,8	No
PHI_14	24.2.2014 18:15	18:15:00	21,3	44,9	21,7	44,2	No
PHI_14	24.2.2014 18:30	18:30:00	21,6	50,3	21,6	45,5	No
PHI_14	24.2.2014 18:45	18:45:00	21,9	52,4	21,6	47,0	No
PHI_14	24.2.2014 19:00	19:00:00	21,8	46,1	21,7	48,4	No
PHI_14	24.2.2014 19:15	19:15:00	21,7	52,9	21,8	50,4	No
PHI_14	24.2.2014 19:30	19:30:00	22	54,5	21,9	51,5	No
PHI_14	24.2.2014 19:45	19:45:00	22,1	53,5	21,9	51,8	No
PHI_14	24.2.2014 20:00	20:00:00	21,6	47,8	21,9	52,2	No
PHI_14	24.2.2014 20:15	20:15:00	21,7	55,7	21,9	52,9	No
PHI_14	24.2.2014 20:30	20:30:00	21,9	56,2	21,8	53,3	No
PHI_14	24.2.2014 20:45	20:45:00	22	56,2	21,8	54,0	No
PHI_14	24.2.2014 21:00	21:00:00	22,1	55,2	21,9	55,8	No
PHI_14	24.2.2014 21:15	21:15:00	22,2	56,3	22,1	56,0	No
PHI_14	24.2.2014 21:30	21:30:00	22,1	52,9	22,1	55,2	No
PHI_14	24.2.2014 21:45	21:45:00	21,6	53,4	22,0	54,5	No
PHI_14	24.2.2014 22:00	22:00:00	21,8	58,3	21,9	55,2	Yes
PHI_14	24.2.2014 22:15	22:15:00	21,9	56,7	21,9	55,3	Yes
PHI_14	24.2.2014 22:30	22:30:00	21,9	57,4	21,8	56,5	Yes
PHI_14	24.2.2014 22:45	22:45:00	22	58,3	21,9	57,7	Yes
PHI_14	24.2.2014 23:00	23:00:00	22,1	59,3	22,0	57,9	Yes
PHI_14	24.2.2014 23:15	23:15:00	22,1	57,2	22,0	58,1	Yes
PHI_14	24.2.2014 23:30	23:30:00	22	55,4	22,1	57,6	Yes
PHI_14	24.2.2014 23:45	23:45:00	22	57,2	22,1	57,3	Yes
PHI_14	25.2.2014 0:00	0:00:00	22,1	58	22,1	57,0	Yes
PHI_14	25.2.2014 0:15	0:15:00	22,1	56,1	22,1	56,7	Yes
PHI_14	25.2.2014 0:30	0:30:00	22,1	57,1	22,1	57,1	Yes
PHI_14	25.2.2014 0:45	0:45:00	22	57,8	22,1	57,3	Yes
PHI_14	25.2.2014 1:00	1:00:00	22,1	58,6	22,1	57,4	Yes
PHI_14	25.2.2014 1:15	1:15:00	22,1	58,8	22,1	58,1	Yes
PHI_14	25.2.2014 1:30	1:30:00	22,1	59,6	22,1	58,7	Yes
PHI_14	25.2.2014 1:45	1:45:00	22,1	60,1	22,1	59,3	Yes
PHI_14	25.2.2014 2:00	2:00:00	22,1	60,7	22,1	59,8	Yes
PHI_14	25.2.2014 2:15	2:15:00	22,1	61,4	22,1	60,5	Yes
PHI_14	25.2.2014 2:30	2:30:00	22,1	58,8	22,1	60,3	Yes
PHI_14	25.2.2014 2:45	2:45:00	21,6	55,4	22,0	59,1	Yes
PHI_14	25.2.2014 3:00	3:00:00	21,6	61,6	21,9	59,3	Yes
PHI_14	25.2.2014 3:15	3:15:00	21,7	61,1	21,8	59,2	Yes
PHI_14	25.2.2014 3:30	3:30:00	21,8	61,1	21,7	59,8	Yes
PHI_14	25.2.2014 3:45	3:45:00	21,8	62,5	21,7	61,6	Yes
PHI_14	25.2.2014 4:00	4:00:00	21,8	63,3	21,8	62,0	Yes
PHI_14	25.2.2014 4:15	4:15:00	21,6	55,6	21,8	60,6	Yes
PHI_14	25.2.2014 4:30	4:30:00	21,4	54,6	21,7	59,0	Yes
PHI_14	25.2.2014 4:45	4:45:00	21,4	62,6	21,6	59,0	Yes
PHI_14	25.2.2014 5:00	5:00:00	21,4	54,7	21,5	56,9	Yes
PHI_14	25.2.2014 5:15	5:15:00	21,1	55,2	21,3	56,8	Yes
PHI_14	25.2.2014 5:30	5:30:00	21,2	62,1	21,3	58,7	Yes
PHI_14	25.2.2014 5:45	5:45:00	21,2	56,5	21,2	57,1	Yes
PHI_14	25.2.2014 6:00	6:00:00	21,2	61,2	21,2	58,8	Yes
PHI_14	25.2.2014 6:15	6:15:00	21,3	63,9	21,2	60,9	Yes
PHI_14	25.2.2014 6:30	6:30:00	21,1	53,7	21,2	58,8	Yes
PHI_14	25.2.2014 6:45	6:45:00	21	59,8	21,2	59,7	Yes
PHI_14	25.2.2014 7:00	7:00:00	21,3	57,4	21,2	58,7	Yes
PHI_14	25.2.2014 7:15	7:15:00	21,2	53,8	21,2	56,2	Yes
PHI_14	25.2.2014 7:30	7:30:00	21,5	52,7	21,3	55,9	Yes
PHI_14	25.2.2014 7:45	7:45:00	21,6	51,4	21,4	53,8	Yes
PHI_14	25.2.2014 8:00	8:00:00	21,7	48,4	21,5	51,6	Yes
PHI_14	25.2.2014 8:15	8:15:00	21,8	47,5	21,7	50,0	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0091 PHI_14_I0101 PHI_14_I0112 PHI_14_I0121	PHI_14_I0094 PHI_14_I0095 PHI_14_I0105 PHI_14_I0115 PHI_14_I0124 PHI_14_I0125	25.2.2014 8:30	8:30:00	22	46,7	21,8	48,5	Yes
		25.2.2014 8:45	8:45:00	22,1	45,4	21,9	47,0	Yes
		25.2.2014 9:00	9:00:00	22,1	45,7	22,0	46,3	Yes
		25.2.2014 9:15	9:15:00	22,2	43,5	22,1	45,3	Yes
		25.2.2014 9:30	9:30:00	22,2	42,6	22,2	44,3	Yes
		25.2.2014 9:45	9:45:00	22,2	40,9	22,2	43,2	Yes
		25.2.2014 10:00	10:00:00	22,3	42,4	22,2	42,4	Yes
		25.2.2014 10:15	10:15:00	22,4	42,4	22,3	42,1	Yes
		25.2.2014 10:30	10:30:00	22,4	40,5	22,3	41,6	Yes
		25.2.2014 10:45	10:45:00	22,4	41,5	22,4	41,7	Yes
		25.2.2014 11:00	11:00:00	22,7	41,2	22,5	41,4	Yes
		25.2.2014 11:15	11:15:00	22,8	41,8	22,6	41,3	Yes
		25.2.2014 11:30	11:30:00	22,9	39,8	22,7	41,1	Yes
		25.2.2014 11:45	11:45:00	23	40,8	22,9	40,9	Yes
		25.2.2014 12:00	12:00:00	22,9	38,3	22,9	40,2	Yes
		25.2.2014 12:15	12:15:00	22,9	41,4	22,9	40,1	Yes
		25.2.2014 12:30	12:30:00	22,8	40,1	22,9	40,2	Yes
		25.2.2014 12:45	12:45:00	22,7	38,9	22,8	39,7	Yes
		25.2.2014 13:00	13:00:00	22,4	37,6	22,7	39,5	Yes
		25.2.2014 13:15	13:15:00	22,5	40,6	22,6	39,3	Yes
		25.2.2014 13:30	13:30:00	22,4	39,8	22,5	39,2	Yes
		25.2.2014 13:45	13:45:00	22,5	40,3	22,5	39,6	Yes
		25.2.2014 14:00	14:00:00	22,5	39,6	22,5	40,1	Yes
		25.2.2014 14:15	14:15:00	22,5	40,1	22,5	40,0	Yes
		25.2.2014 14:30	14:30:00	22,5	40,7	22,5	40,2	Yes
		25.2.2014 14:45	14:45:00	22,5	40,2	22,5	40,2	No
		25.2.2014 15:00	15:00:00	22,6	40,5	22,5	40,4	No
		25.2.2014 15:15	15:15:00	22,7	41,3	22,6	40,7	No
		25.2.2014 15:30	15:30:00	23,1	42,1	22,7	41,0	No
		25.2.2014 15:45	15:45:00	23,1	38,7	22,9	40,7	No
		25.2.2014 16:00	16:00:00	22,7	37,8	22,9	40,0	No
		25.2.2014 16:15	16:15:00	22,3	39	22,8	39,4	No
		25.2.2014 16:30	16:30:00	22,1	38	22,6	38,4	No
		25.2.2014 16:45	16:45:00	22,2	46	22,3	40,2	No
		25.2.2014 17:00	17:00:00	22,4	48,2	22,3	42,8	No
		25.2.2014 17:15	17:15:00	22,5	49,4	22,3	45,4	No
		25.2.2014 17:30	17:30:00	22,6	50,1	22,4	48,4	No
		25.2.2014 17:45	17:45:00	22,7	51,6	22,6	49,8	No
		25.2.2014 18:00	18:00:00	22,7	52,2	22,6	50,8	No
		25.2.2014 18:15	18:15:00	22,8	51,4	22,7	51,3	No
		25.2.2014 18:30	18:30:00	22,8	52,1	22,8	51,8	No
		25.2.2014 18:45	18:45:00	22,8	51,8	22,8	51,9	No
		25.2.2014 19:00	19:00:00	22,8	51,6	22,8	51,7	No
		25.2.2014 19:15	19:15:00	22,7	52,7	22,8	52,1	No
		25.2.2014 19:30	19:30:00	22,7	53,1	22,8	52,3	No
		25.2.2014 19:45	19:45:00	22,7	53,5	22,7	52,7	No
		25.2.2014 20:00	20:00:00	22,7	53,1	22,7	53,1	No
		25.2.2014 20:15	20:15:00	22,7	52,9	22,7	53,2	No
		25.2.2014 20:30	20:30:00	22,6	53,3	22,7	53,2	No
		25.2.2014 20:45	20:45:00	22,5	53,6	22,6	53,2	No
		25.2.2014 21:00	21:00:00	22,5	52,9	22,6	53,2	No
		25.2.2014 21:15	21:15:00	22,4	54,1	22,5	53,5	No
		25.2.2014 21:30	21:30:00	22,4	53,6	22,5	53,6	No
		25.2.2014 21:45	21:45:00	22,4	53,5	22,4	53,5	No
		25.2.2014 22:00	22:00:00	22,4	53,9	22,4	53,8	Yes
		25.2.2014 22:15	22:15:00	22,3	52,1	22,4	53,3	Yes
		25.2.2014 22:30	22:30:00	22,3	54,1	22,4	53,4	Yes
		25.2.2014 22:45	22:45:00	22,2	53,6	22,3	53,4	Yes
		25.2.2014 23:00	23:00:00	22,2	53,7	22,3	53,4	Yes
		25.2.2014 23:15	23:15:00	22,1	52,1	22,2	53,4	Yes
		25.2.2014 23:30	23:30:00	22	52,2	22,1	52,9	Yes
		25.2.2014 23:45	23:45:00	22	54,4	22,1	53,1	Yes
		26.2.2014 0:00	0:00:00	22	53,3	22,0	53,0	Yes
		26.2.2014 0:15	0:15:00	21,9	52,3	22,0	53,1	Yes
		26.2.2014 0:30	0:30:00	21,9	54,2	22,0	53,6	Yes
		26.2.2014 0:45	0:45:00	21,9	53,8	21,9	53,4	Yes
		26.2.2014 1:00	1:00:00	21,8	53,8	21,9	53,5	Yes
		26.2.2014 1:15	1:15:00	21,8	55,3	21,9	54,3	Yes
		26.2.2014 1:30	1:30:00	21,8	54,2	21,8	54,3	Yes
		26.2.2014 1:45	1:45:00	21,8	54,3	21,8	54,4	Yes
		26.2.2014 2:00	2:00:00	21,8	54,6	21,8	54,6	Yes
		26.2.2014 2:15	2:15:00	21,8	54,9	21,8	54,5	Yes
		26.2.2014 2:30	2:30:00	21,8	55,1	21,8	54,7	Yes
		26.2.2014 2:45	2:45:00	21,8	55,6	21,8	55,1	Yes
		26.2.2014 3:00	3:00:00	21,8	55,5	21,8	55,3	Yes
		26.2.2014 3:15	3:15:00	21,8	55,2	21,8	55,4	Yes
		26.2.2014 3:30	3:30:00	21,8	55,5	21,8	55,5	Yes
		26.2.2014 3:45	3:45:00	21,8	55,7	21,8	55,5	Yes
		26.2.2014 4:00	4:00:00	21,8	55,9	21,8	55,6	Yes
		26.2.2014 4:15	4:15:00	21,8	56,1	21,8	55,8	Yes
		26.2.2014 4:30	4:30:00	21,8	55,6	21,8	55,8	Yes
		26.2.2014 4:45	4:45:00	21,9	55,3	21,8	55,7	Yes
		26.2.2014 5:00	5:00:00	21,9	56,1	21,9	55,8	Yes
		26.2.2014 5:15	5:15:00	22	56,6	21,9	55,9	Yes
		26.2.2014 5:30	5:30:00	22,1	56,4	22,0	56,1	Yes
		26.2.2014 5:45	5:45:00	22,1	55,7	22,0	56,2	Yes
		26.2.2014 6:00	6:00:00	22,2	55,6	22,1	56,1	Yes
		26.2.2014 6:15	6:15:00	22,2	55	22,2	55,7	Yes
		26.2.2014 6:30	6:30:00	22,2	55,1	22,2	55,4	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0091 PHI_14_I0101 PHI_14_I0112 PHI_14_I0121	PHI_14_I0095 PHI_14_I0105 PHI_14_I0115 PHI_14_I0125	N/A	26.2.2014 6:45	6:45:00	22,4	56,6	22,3	55,6	Yes
			26.2.2014 7:00	7:00:00	23	55,1	22,5	55,5	Yes
			26.2.2014 7:15	7:15:00	23,3	56,5	22,7	55,8	Yes
			26.2.2014 7:30	7:30:00	23,2	53,7	23,0	55,5	Yes
N/A	N/A	N/A	26.2.2014 7:45	7:45:00	23,5	50,8	23,3	54,0	Yes
			26.2.2014 8:00	8:00:00	23,4	47,9	23,4	52,2	Yes
			26.2.2014 8:15	8:15:00	23,4	43,7	23,4	49,0	Yes
			26.2.2014 8:30	8:30:00	23,3	44,1	23,4	46,6	Yes
N/A	N/A	N/A	26.2.2014 8:45	8:45:00	23,3	49	23,4	46,2	Yes
			26.2.2014 9:00	9:00:00	23,7	51,8	23,4	47,2	Yes
			26.2.2014 9:15	9:15:00	23,8	48,3	23,5	48,3	Yes
			26.2.2014 9:30	9:30:00	23,6	44,7	23,6	48,5	Yes
N/A	N/A	N/A	26.2.2014 9:45	9:45:00	23,8	49	23,7	48,5	Yes
			26.2.2014 10:00	10:00:00	24,1	48,5	23,8	47,6	Yes
			26.2.2014 10:15	10:15:00	24,5	50,9	24,0	48,3	Yes
			26.2.2014 10:30	10:30:00	24,7	43,4	24,3	48,0	Yes
N/A	N/A	N/A	26.2.2014 10:45	10:45:00	24,5	45,4	24,5	47,1	Yes
			26.2.2014 11:00	11:00:00	24,4	43,6	24,5	45,8	Yes
			26.2.2014 11:15	11:15:00	24,2	42,8	24,5	43,8	Yes
			26.2.2014 11:30	11:30:00	24,1	42,2	24,3	43,5	Yes
N/A	N/A	N/A	26.2.2014 11:45	11:45:00	24,3	43,3	24,3	43,0	Yes
			26.2.2014 12:00	12:00:00	24,7	46,1	24,3	43,6	Yes
			26.2.2014 12:15	12:15:00	25,1	45,9	24,6	44,4	Yes
			26.2.2014 12:30	12:30:00	25,2	45,6	24,8	45,2	Yes
N/A	N/A	N/A	26.2.2014 12:45	12:45:00	25,4	44,8	25,1	45,6	Yes
			26.2.2014 13:00	13:00:00	25,3	38	25,3	43,6	Yes
			26.2.2014 13:15	13:15:00	24,9	38,8	25,2	41,8	Yes
			26.2.2014 13:30	13:30:00	23,1	43,1	24,7	41,2	Yes
N/A	N/A	N/A	26.2.2014 13:45	13:45:00	23,3	44,2	24,2	41,0	Yes
			26.2.2014 14:00	14:00:00	23,3	42,2	23,7	42,1	Yes
			26.2.2014 14:15	14:15:00	23,4	42,4	23,3	43,0	Yes
			26.2.2014 14:30	14:30:00	23,5	42	23,4	42,7	Yes
N/A	N/A	N/A	26.2.2014 14:45	14:45:00	23	38,9	23,3	41,4	No
			26.2.2014 15:00	15:00:00	22,7	43	23,2	41,6	No
			26.2.2014 15:15	15:15:00	22,6	47	23,0	42,7	No
			26.2.2014 15:30	15:30:00	22,5	46,8	22,7	43,9	No
N/A	N/A	N/A	26.2.2014 15:45	15:45:00	22,4	49,6	22,6	46,6	No
			26.2.2014 16:00	16:00:00	22,3	49	22,5	48,1	No
			26.2.2014 16:15	16:15:00	22,2	50,5	22,4	49,0	No
			26.2.2014 16:30	16:30:00	22,1	52,5	22,3	50,4	No
N/A	N/A	N/A	26.2.2014 16:45	16:45:00	22	53,4	22,2	51,4	No
			26.2.2014 17:00	17:00:00	22	54,3	22,1	52,7	No
			26.2.2014 17:15	17:15:00	21,9	55	22,0	53,8	No
			26.2.2014 17:30	17:30:00	21,8	54,8	21,9	54,4	No
N/A	N/A	N/A	26.2.2014 17:45	17:45:00	21,8	55,1	21,9	54,8	No
			26.2.2014 18:00	18:00:00	21,8	53,7	21,8	54,7	No
			26.2.2014 18:15	18:15:00	21,8	55,2	21,8	54,7	No
			26.2.2014 18:30	18:30:00	21,7	54,6	21,8	54,7	No
N/A	N/A	N/A	26.2.2014 18:45	18:45:00	21,7	55,3	21,8	54,7	No
			26.2.2014 19:00	19:00:00	21,7	55,7	21,7	55,2	No
			26.2.2014 19:15	19:15:00	21,7	55,2	21,7	55,2	No
			26.2.2014 19:30	19:30:00	21,6	55,4	21,7	55,4	No
N/A	N/A	N/A	26.2.2014 19:45	19:45:00	21,6	55,3	21,7	55,4	No
			26.2.2014 20:00	20:00:00	21,7	53,9	21,7	55,0	No
			26.2.2014 20:15	20:15:00	21,5	56,4	21,6	55,3	No
			26.2.2014 20:30	20:30:00	21,6	56,7	21,6	55,6	No
N/A	N/A	N/A	26.2.2014 20:45	20:45:00	21,5	55,6	21,6	55,7	No
			26.2.2014 21:00	21:00:00	21,6	56,4	21,6	56,3	No
			26.2.2014 21:15	21:15:00	21,6	56,2	21,6	56,2	No
			26.2.2014 21:30	21:30:00	21,5	55,9	21,6	56,0	No
N/A	N/A	N/A	26.2.2014 21:45	21:45:00	21,6	56,2	21,6	56,2	No
			26.2.2014 22:00	22:00:00	21,5	56,9	21,6	56,3	Yes
			26.2.2014 22:15	22:15:00	21,5	54,4	21,5	55,9	Yes
			26.2.2014 22:30	22:30:00	21,5	54,2	21,5	55,4	Yes
N/A	N/A	N/A	26.2.2014 22:45	22:45:00	21,5	55,4	21,5	55,2	Yes
			26.2.2014 23:00	23:00:00	21,4	56	21,5	55,0	Yes
			26.2.2014 23:15	23:15:00	21,4	54,4	21,5	55,0	Yes
			26.2.2014 23:30	23:30:00	21,4	55,6	21,4	55,4	Yes
N/A	N/A	N/A	26.2.2014 23:45	23:45:00	21,4	55,9	21,4	55,5	Yes
			27.2.2014 0:00	0:00:00	21,4	56,6	21,4	55,6	Yes
			27.2.2014 0:15	0:15:00	21,4	57,2	21,4	56,3	Yes
			27.2.2014 0:30	0:30:00	21,4	57,4	21,4	56,8	Yes
N/A	N/A	N/A	27.2.2014 0:45	0:45:00	21,5	58,1	21,4	57,3	Yes
			27.2.2014 1:00	1:00:00	21,4	57,4	21,4	57,5	Yes
			27.2.2014 1:15	1:15:00	21,5	57,4	21,5	57,6	Yes
			27.2.2014 1:30	1:30:00	21,5	56,9	21,5	57,5	Yes
N/A	N/A	N/A	27.2.2014 1:45	1:45:00	21,4	57,2	21,5	57,2	Yes
			27.2.2014 2:00	2:00:00	21,4	58,1	21,5	57,4	Yes
			27.2.2014 2:15	2:15:00	21,5	57,5	21,5	57,4	Yes
			27.2.2014 2:30	2:30:00	21,5	57,2	21,5	57,5	Yes
N/A	N/A	N/A	27.2.2014 2:45	2:45:00	21,4	57,2	21,5	57,5	Yes
			27.2.2014 3:00	3:00:00	21,4	57,9	21,5	57,5	Yes
			27.2.2014 3:15	3:15:00	21,4	55,8	21,4	57,0	Yes
			27.2.2014 3:30	3:30:00	21,5	57,3	21,4	57,1	Yes
N/A	N/A	N/A	27.2.2014 3:45	3:45:00	21,5	58,2	21,5	57,3	Yes
			27.2.2014 4:00	4:00:00	21,6	58,1	21,5	57,4	Yes
			27.2.2014 4:15	4:15:00	21,6	56,3	21,6	57,5	Yes
			27.2.2014 4:30	4:30:00	21,6	55,3	21,6	57,0	Yes
N/A	N/A	N/A	27.2.2014 4:45	4:45:00	21,5	56,9	21,6	56,7	Yes



TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0091 PHI_14_I0092 PHI_14_I0101 PHI_14_I0102 PHI_14_I0112 PHI_14_I0121 PHI_14_I0125	PHI_14_I0095 PHI_14_I0106 PHI_14_I0105 PHI_14_I0116 PHI_14_I0125	27.2.2014 5:00	5:00:00	21,6	57,1	21,6	56,4	Yes
		27.2.2014 5:15	5:15:00	21,5	58	21,6	56,8	Yes
		27.2.2014 5:30	5:30:00	21,5	58	21,5	57,5	Yes
		27.2.2014 5:45	5:45:00	21,7	57,7	21,6	57,7	Yes
		27.2.2014 6:00	6:00:00	21,9	53,5	21,7	56,8	Yes
		27.2.2014 6:15	6:15:00	22,1	51,8	21,8	55,3	Yes
		27.2.2014 6:30	6:30:00	22,3	50,9	22,0	53,5	Yes
		27.2.2014 6:45	6:45:00	22,4	48,2	22,2	51,1	Yes
		27.2.2014 7:00	7:00:00	22,6	48	22,4	49,7	Yes
		27.2.2014 7:15	7:15:00	22,7	47,1	22,5	48,6	Yes
		27.2.2014 7:30	7:30:00	22,8	45,4	22,6	47,2	Yes
		27.2.2014 7:45	7:45:00	22,9	44,8	22,8	46,3	Yes
		27.2.2014 8:00	8:00:00	23	44,9	22,9	45,6	Yes
		27.2.2014 8:15	8:15:00	23,1	44,6	23,0	44,9	Yes
		27.2.2014 8:30	8:30:00	23,1	43,5	23,0	44,5	Yes
		27.2.2014 8:45	8:45:00	23,2	43	23,1	44,0	Yes
		27.2.2014 9:00	9:00:00	23,1	41,1	23,1	43,1	Yes
		27.2.2014 9:15	9:15:00	23	41,4	23,1	42,3	Yes
		27.2.2014 9:30	9:30:00	23	42,5	23,1	42,0	Yes
		27.2.2014 9:45	9:45:00	23	41,7	23,0	41,7	Yes
		27.2.2014 10:00	10:00:00	23,2	42,2	23,1	42,0	Yes
		27.2.2014 10:15	10:15:00	23,1	40,5	23,1	41,7	Yes
		27.2.2014 10:30	10:30:00	23,2	41,4	23,1	41,5	Yes
		27.2.2014 10:45	10:45:00	23,2	40,1	23,2	41,1	Yes
		27.2.2014 11:00	11:00:00	23,2	41,2	23,2	40,8	Yes
		27.2.2014 11:15	11:15:00	23,2	39,3	23,2	40,5	Yes
		27.2.2014 11:30	11:30:00	23,1	39,6	23,2	40,1	Yes
		27.2.2014 11:45	11:45:00	23	38,9	23,1	39,8	Yes
		27.2.2014 12:00	12:00:00	23,1	39,6	23,1	39,4	Yes
		27.2.2014 12:15	12:15:00	23,1	39,1	23,1	39,3	Yes
		27.2.2014 12:30	12:30:00	23,1	40,2	23,1	39,5	Yes
		27.2.2014 12:45	12:45:00	23,1	39,4	23,1	39,6	Yes
		27.2.2014 13:00	13:00:00	23,1	39,8	23,1	39,6	Yes
		27.2.2014 13:15	13:15:00	22,9	38,2	23,1	39,4	Yes
		27.2.2014 13:30	13:30:00	23	39,8	23,0	39,3	Yes
		27.2.2014 13:45	13:45:00	22,8	40,1	23,0	39,5	Yes
		27.2.2014 14:00	14:00:00	22,7	41,4	22,9	39,9	Yes
		27.2.2014 14:15	14:15:00	22,6	41,1	22,8	40,6	Yes
		27.2.2014 14:30	14:30:00	22,6	39,9	22,7	40,6	Yes
		27.2.2014 14:45	14:45:00	22,2	37,7	22,5	40,0	No
		27.2.2014 15:00	15:00:00	22,1	39,2	22,4	39,5	No
		27.2.2014 15:15	15:15:00	21,8	37,9	22,2	38,7	No
		27.2.2014 15:30	15:30:00	21,6	38,6	21,9	38,4	No
		27.2.2014 15:45	15:45:00	21,7	46,5	21,8	40,6	No
		27.2.2014 16:00	16:00:00	21,6	46,6	21,7	42,4	No
		27.2.2014 16:15	16:15:00	21,8	49,7	21,7	45,4	No
		27.2.2014 16:30	16:30:00	21,8	46,8	21,7	47,4	No
		27.2.2014 16:45	16:45:00	21,7	51,7	21,7	48,7	No
		27.2.2014 17:00	17:00:00	22	52,1	21,8	50,1	No
		27.2.2014 17:15	17:15:00	22,1	52,5	21,9	50,8	No
		27.2.2014 17:30	17:30:00	21,8	49,5	21,9	51,5	No
		27.2.2014 17:45	17:45:00	21,8	53,8	21,9	52,0	No
		27.2.2014 18:00	18:00:00	22	54	21,9	52,5	No
		27.2.2014 18:15	18:15:00	22,1	53,9	21,9	52,8	No
		27.2.2014 18:30	18:30:00	22,2	54,3	22,0	54,0	No
		27.2.2014 18:45	18:45:00	22,2	54,6	22,1	54,2	No
		27.2.2014 19:00	19:00:00	22,3	55	22,2	54,5	No
		27.2.2014 19:15	19:15:00	22,3	53,2	22,3	54,3	No
		27.2.2014 19:30	19:30:00	21,7	53,4	22,1	54,1	No
		27.2.2014 19:45	19:45:00	21,8	55,6	22,0	54,3	No
		27.2.2014 20:00	20:00:00	21,9	54,8	21,9	54,3	No
		27.2.2014 20:15	20:15:00	21,9	55	21,8	54,7	No
		27.2.2014 20:30	20:30:00	21,9	55,3	21,9	55,2	No
		27.2.2014 20:45	20:45:00	22	55,8	21,9	55,2	No
		27.2.2014 21:00	21:00:00	22	55,8	22,0	55,5	No
		27.2.2014 21:15	21:15:00	21,9	53,7	22,0	55,2	No
		27.2.2014 21:30	21:30:00	21,9	54,7	22,0	55,0	No
		27.2.2014 21:45	21:45:00	21,9	55,4	21,9	54,9	No
		27.2.2014 22:00	22:00:00	21,9	55,8	21,9	54,9	Yes
		27.2.2014 22:15	22:15:00	21,9	51,6	21,9	54,4	Yes
		27.2.2014 22:30	22:30:00	21,8	52,8	21,9	53,9	Yes
		27.2.2014 22:45	22:45:00	21,8	54,4	21,9	53,7	Yes
		27.2.2014 23:00	23:00:00	21,8	54,7	21,8	53,4	Yes
		27.2.2014 23:15	23:15:00	21,7	52,8	21,8	53,7	Yes
		27.2.2014 23:30	23:30:00	21,6	54,3	21,7	54,1	Yes
		27.2.2014 23:45	23:45:00	21,6	54,9	21,7	54,2	Yes
N/A	N/A	28.2.2014 0:00	0:00:00	21,6	55,5	21,6	54,4	Yes
		28.2.2014 0:15	0:15:00	21,6	55,8	21,6	55,1	Yes
		28.2.2014 0:30	0:30:00	21,6	55,8	21,6	55,5	Yes
		28.2.2014 0:45	0:45:00	21,6	55,8	21,6	55,7	Yes
		28.2.2014 1:00	1:00:00	21,6	55,7	21,6	55,8	Yes
		28.2.2014 1:15	1:15:00	21,7	55,5	21,6	55,7	Yes
		28.2.2014 1:30	1:30:00	21,7	55,6	21,7	55,7	Yes
		28.2.2014 1:45	1:45:00	21,7	55,2	21,7	55,5	Yes
		28.2.2014 2:00	2:00:00	21,7	55,4	21,7	55,4	Yes
		28.2.2014 2:15	2:15:00	21,7	56,1	21,7	55,6	Yes
		28.2.2014 2:30	2:30:00	21,7	55,3	21,7	55,5	Yes
		28.2.2014 2:45	2:45:00	21,7	55	21,7	55,5	Yes
		28.2.2014 3:00	3:00:00	21,7	55,9	21,7	55,6	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0092	28.2.2014 3:15	3:15:00	21.4	56.5	21.6	55.7	Yes
	28.2.2014 3:30	3:30:00	21.5	56.2	21.6	55.9	Yes
	28.2.2014 3:45	3:45:00	21.3	57	21.5	56.4	Yes
	28.2.2014 4:00	4:00:00	21.5	54.3	21.4	56.0	Yes
	28.2.2014 4:15	4:15:00	21.2	56.7	21.4	56.1	Yes
	28.2.2014 4:30	4:30:00	21.4	54.6	21.4	55.7	Yes
	28.2.2014 4:45	4:45:00	21.2	56.3	21.3	55.5	Yes
	28.2.2014 5:00	5:00:00	21.4	55.7	21.3	55.8	Yes
	28.2.2014 5:15	5:15:00	21.2	56.2	21.3	55.7	Yes
	28.2.2014 5:30	5:30:00	21.4	54.7	21.3	55.7	Yes
PHI_14_I0112	28.2.2014 5:45	5:45:00	21.3	48.9	21.3	53.9	Yes
	28.2.2014 6:00	6:00:00	21.4	52.2	21.3	53.0	Yes
	28.2.2014 6:15	6:15:00	21.7	53.4	21.5	52.3	Yes
	28.2.2014 6:30	6:30:00	21.6	48.1	21.5	50.7	Yes
	28.2.2014 6:45	6:45:00	21.8	50.7	21.6	51.1	Yes
	28.2.2014 7:00	7:00:00	21.8	46.8	21.7	49.8	Yes
	28.2.2014 7:15	7:15:00	22	47.4	21.8	48.3	Yes
	28.2.2014 7:30	7:30:00	22.1	45.6	21.9	47.6	Yes
	28.2.2014 7:45	7:45:00	22.3	45.2	22.1	46.3	Yes
	28.2.2014 8:00	8:00:00	22.4	45.5	22.2	45.9	Yes
PHI_14_I0121	28.2.2014 8:15	8:15:00	22.5	43.7	22.3	45.0	Yes
	28.2.2014 8:30	8:30:00	22.6	43.1	22.5	44.4	Yes
	28.2.2014 8:45	8:45:00	22.5	41.8	22.5	43.5	Yes
	28.2.2014 9:00	9:00:00	22.6	41.9	22.6	42.6	Yes
	28.2.2014 9:15	9:15:00	22.7	42.9	22.6	42.4	Yes
	28.2.2014 9:30	9:30:00	22.7	40.9	22.6	41.9	Yes
	28.2.2014 9:45	9:45:00	22.8	41	22.7	41.7	Yes
	28.2.2014 10:00	10:00:00	22.8	40.4	22.8	41.3	Yes
	28.2.2014 10:15	10:15:00	22.8	41.2	22.8	40.9	Yes
	28.2.2014 10:30	10:30:00	22.9	42	22.8	41.2	Yes
PHI_14_I0096	28.2.2014 10:45	10:45:00	22.9	39.9	22.9	40.9	Yes
	28.2.2014 11:00	11:00:00	23	41.2	22.9	41.1	Yes
	28.2.2014 11:15	11:15:00	23	39.8	23.0	40.7	Yes
	28.2.2014 11:30	11:30:00	23	39.3	23.0	40.1	Yes
	28.2.2014 11:45	11:45:00	22.9	37.3	23.0	39.4	Yes
	28.2.2014 12:00	12:00:00	22.9	38.8	23.0	38.8	Yes
	28.2.2014 12:15	12:15:00	23	38.1	23.0	38.4	Yes
	28.2.2014 12:30	12:30:00	23	38.9	23.0	38.3	Yes
	28.2.2014 12:45	12:45:00	23.1	39	23.0	38.7	Yes
	28.2.2014 13:00	13:00:00	23.1	40.5	23.1	39.1	Yes
PHI_14_I0106	28.2.2014 13:15	13:15:00	23.1	41.4	23.1	40.0	Yes
	28.2.2014 13:30	13:30:00	23.1	42.3	23.1	40.8	Yes
	28.2.2014 13:45	13:45:00	23.1	42	23.1	41.6	Yes
	28.2.2014 14:00	14:00:00	23.1	43.3	23.1	42.3	Yes
	28.2.2014 14:15	14:15:00	23.1	43.5	23.1	42.8	Yes
	28.2.2014 14:30	14:30:00	23.2	42.7	23.1	42.9	Yes
	28.2.2014 14:45	14:45:00	22.6	38.5	23.0	42.0	No
	28.2.2014 15:00	15:00:00	22.1	37.7	22.8	40.6	No
	28.2.2014 15:15	15:15:00	22.1	42.8	22.5	40.4	No
	28.2.2014 15:30	15:30:00	21.7	46.4	22.1	41.4	No
PHI_14_I0116	28.2.2014 15:45	15:45:00	21.7	44.4	21.9	42.8	No
	28.2.2014 16:00	16:00:00	21.8	50.6	21.8	46.1	No
	28.2.2014 16:15	16:15:00	21.6	48.5	21.7	47.5	No
	28.2.2014 16:30	16:30:00	21.8	52.2	21.7	48.9	No
	28.2.2014 16:45	16:45:00	22	52.4	21.8	50.9	No
	28.2.2014 17:00	17:00:00	21.6	51.5	21.8	51.2	No
	28.2.2014 17:15	17:15:00	21.8	53	21.8	52.3	No
	28.2.2014 17:30	17:30:00	21.5	53.5	21.7	52.6	No
	28.2.2014 17:45	17:45:00	21.7	54.3	21.7	53.1	No
	28.2.2014 18:00	18:00:00	21.8	54.4	21.7	53.8	No
PHI_14_I0126	28.2.2014 18:15	18:15:00	21.9	52.3	21.7	53.6	No
	28.2.2014 18:30	18:30:00	21.9	53.6	21.8	53.7	No
	28.2.2014 18:45	18:45:00	22	54	21.9	53.6	No
	28.2.2014 19:00	19:00:00	22	54.6	22.0	53.6	No
	28.2.2014 19:15	19:15:00	22.1	55	22.0	54.3	No
	28.2.2014 19:30	19:30:00	22	51.2	22.0	53.7	No
	28.2.2014 19:45	19:45:00	22	52.4	22.0	53.3	No
	28.2.2014 20:00	20:00:00	22	51.5	22.0	52.5	No
	28.2.2014 20:15	20:15:00	21.9	53	22.0	52.0	No
	28.2.2014 20:30	20:30:00	21.9	53.7	22.0	52.7	No
	28.2.2014 20:45	20:45:00	21.9	54.4	21.9	53.2	No
	28.2.2014 21:00	21:00:00	21.9	54.8	21.9	54.0	No
	28.2.2014 21:15	21:15:00	21.9	53	21.9	54.0	No
	28.2.2014 21:30	21:30:00	21.9	54.1	21.9	54.1	No
	28.2.2014 21:45	21:45:00	21.9	54.7	21.9	54.2	No
	28.2.2014 22:00	22:00:00	21.9	55.4	21.9	54.3	No
	28.2.2014 22:15	22:15:00	21.8	51.9	21.9	54.0	No
	28.2.2014 22:30	22:30:00	21.8	52.7	21.9	53.7	No
	28.2.2014 22:45	22:45:00	21.7	53.7	21.8	53.4	No
	28.2.2014 23:00	23:00:00	21.7	54.4	21.8	53.2	No
	28.2.2014 23:15	23:15:00	21.7	52.8	21.7	53.4	No
	28.2.2014 23:30	23:30:00	21.6	53.8	21.7	53.7	No
	28.2.2014 23:45	23:45:00	21.6	54.6	21.7	53.9	No
	1.3.2014 0:00	0:00:00	21.6	55	21.6	54.1	No
	1.3.2014 0:15	0:15:00	21.6	55.2	21.6	54.7	No
	1.3.2014 0:30	0:30:00	21.6	55.7	21.6	55.1	No
	1.3.2014 0:45	0:45:00	21.5	55.9	21.6	55.5	No
	1.3.2014 1:00	1:00:00	21.5	56.2	21.6	55.8	No
	1.3.2014 1:15	1:15:00	21.5	56.4	21.5	56.1	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	1.3.2014 1:30	1:30:00	21,5	56,7	21,5	56,3	No
		1.3.2014 1:45	1:45:00	21,5	56,5	21,5	56,5	No
		1.3.2014 2:00	2:00:00	21,5	57	21,5	56,7	No
		1.3.2014 2:15	2:15:00	21,4	57,1	21,5	56,8	No
		1.3.2014 2:30	2:30:00	21,4	56	21,5	56,7	No
		1.3.2014 2:45	2:45:00	21,4	56,8	21,4	56,7	No
		1.3.2014 3:00	3:00:00	21,4	56,8	21,4	56,7	No
		1.3.2014 3:15	3:15:00	21,4	55,9	21,4	56,4	No
		1.3.2014 3:30	3:30:00	21,4	57,2	21,4	56,7	No
		1.3.2014 3:45	3:45:00	21,4	56,6	21,4	56,6	No
		1.3.2014 4:00	4:00:00	21,3	56,1	21,4	56,5	No
		1.3.2014 4:15	4:15:00	21,3	57,1	21,4	56,8	No
		1.3.2014 4:30	4:30:00	21,3	56,5	21,3	56,6	No
		1.3.2014 4:45	4:45:00	21,3	56,6	21,3	56,6	No
		1.3.2014 5:00	5:00:00	21,3	56,9	21,3	56,8	No
		1.3.2014 5:15	5:15:00	21,2	56,4	21,3	56,6	No
		1.3.2014 5:30	5:30:00	21,2	56,6	21,3	56,6	No
		1.3.2014 5:45	5:45:00	21,2	56,8	21,2	56,7	No
		1.3.2014 6:00	6:00:00	21,2	56,3	21,2	56,5	No
		1.3.2014 6:15	6:15:00	21,2	56,7	21,2	56,6	No
		1.3.2014 6:30	6:30:00	21,1	54,7	21,2	56,1	No
		1.3.2014 6:45	6:45:00	21	55,8	21,1	55,9	No
		1.3.2014 7:00	7:00:00	21	56,7	21,1	56,0	No
		1.3.2014 7:15	7:15:00	21	57	21,0	56,1	No
		1.3.2014 7:30	7:30:00	21	57,3	21,0	56,7	No
		1.3.2014 7:45	7:45:00	20,9	56,6	21,0	56,9	No
		1.3.2014 8:00	8:00:00	20,9	57,1	21,0	57,0	No
		1.3.2014 8:15	8:15:00	20,9	56,6	20,9	56,9	No
		1.3.2014 8:30	8:30:00	20,9	56,4	20,9	56,7	No
		1.3.2014 8:45	8:45:00	20,8	55,8	20,9	56,5	No
		1.3.2014 9:00	9:00:00	20,8	56,6	20,9	56,4	No
		1.3.2014 9:15	9:15:00	20,8	55,7	20,8	56,1	No
		1.3.2014 9:30	9:30:00	20,8	56,3	20,8	56,1	No
		1.3.2014 9:45	9:45:00	20,8	56,2	20,8	56,2	No
		1.3.2014 10:00	10:00:00	20,8	56	20,8	56,1	No
		1.3.2014 10:15	10:15:00	20,8	56,5	20,8	56,3	No
		1.3.2014 10:30	10:30:00	20,7	55,6	20,8	56,1	No
		1.3.2014 10:45	10:45:00	20,7	54,6	20,8	55,7	No
		1.3.2014 11:00	11:00:00	20,7	56,1	20,7	55,7	No
		1.3.2014 11:15	11:15:00	20,7	56,1	20,7	55,6	No
		1.3.2014 11:30	11:30:00	20,7	55,7	20,7	55,6	No
		1.3.2014 11:45	11:45:00	20,7	55,9	20,7	56,0	No
		1.3.2014 12:00	12:00:00	20,7	53,2	20,7	55,2	No
		1.3.2014 12:15	12:15:00	20,6	54	20,7	54,7	No
		1.3.2014 12:30	12:30:00	20,6	55,9	20,7	54,8	No
		1.3.2014 12:45	12:45:00	20,6	56,8	20,6	55,0	No
		1.3.2014 13:00	13:00:00	20,7	55,3	20,6	55,5	No
		1.3.2014 13:15	13:15:00	20,7	56,5	20,7	56,1	No
		1.3.2014 13:30	13:30:00	20,8	55,4	20,7	56,0	No
		1.3.2014 13:45	13:45:00	20,8	56,5	20,8	55,9	No
		1.3.2014 14:00	14:00:00	20,8	55,6	20,8	56,0	No
		1.3.2014 14:15	14:15:00	20,9	53,8	20,8	55,3	No
		1.3.2014 14:30	14:30:00	20,9	55,7	20,9	55,4	No
		1.3.2014 14:45	14:45:00	20,9	55	20,9	55,0	No
		1.3.2014 15:00	15:00:00	20,9	56,1	20,9	55,2	No
		1.3.2014 15:15	15:15:00	21	55	20,9	55,5	No
		1.3.2014 15:30	15:30:00	21	55,7	21,0	55,5	No
		1.3.2014 15:45	15:45:00	21,1	55,5	21,0	55,6	No
		1.3.2014 16:00	16:00:00	21,1	55,4	21,1	55,4	No
		1.3.2014 16:15	16:15:00	21,1	55,1	21,1	55,4	No
		1.3.2014 16:30	16:30:00	21,2	56,5	21,1	55,6	No
		1.3.2014 16:45	16:45:00	21,2	55,7	21,2	55,7	No
		1.3.2014 17:00	17:00:00	21,3	55,6	21,2	55,7	No
		1.3.2014 17:15	17:15:00	21,3	55,7	21,3	55,9	No
		1.3.2014 17:30	17:30:00	21,3	55,8	21,3	55,7	No
		1.3.2014 17:45	17:45:00	21,3	56	21,3	55,8	No
		1.3.2014 18:00	18:00:00	21,3	55,4	21,3	55,7	No
		1.3.2014 18:15	18:15:00	21,3	56,6	21,3	56,0	No
		1.3.2014 18:30	18:30:00	21,3	55,2	21,3	55,8	No
		1.3.2014 18:45	18:45:00	21,3	56,4	21,3	55,9	No
		1.3.2014 19:00	19:00:00	21,3	55,3	21,3	55,9	No
		1.3.2014 19:15	19:15:00	21,3	55,7	21,3	55,7	No
		1.3.2014 19:30	19:30:00	21,3	56,1	21,3	55,9	No
		1.3.2014 19:45	19:45:00	21,3	55,2	21,3	55,6	No
		1.3.2014 20:00	20:00:00	21,2	53,9	21,3	55,2	No
		1.3.2014 20:15	20:15:00	21,2	55,8	21,3	55,3	No
		1.3.2014 20:30	20:30:00	21,1	56,8	21,2	55,4	No
		1.3.2014 20:45	20:45:00	21,1	55,6	21,2	55,5	No
		1.3.2014 21:00	21:00:00	21,1	56,1	21,1	56,1	No
		1.3.2014 21:15	21:15:00	21,1	57	21,1	56,4	No
		1.3.2014 21:30	21:30:00	21,1	55	21,1	55,9	No
		1.3.2014 21:45	21:45:00	21	56,8	21,1	56,2	No
		1.3.2014 22:00	22:00:00	21	55,5	21,1	56,1	No
		1.3.2014 22:15	22:15:00	21	56,2	21,0	55,9	No
		1.3.2014 22:30	22:30:00	21	55,3	21,0	56,0	No
		1.3.2014 22:45	22:45:00	21	55,9	21,0	55,7	No
		1.3.2014 23:00	23:00:00	20,9	55,1	21,0	55,6	No
		1.3.2014 23:15	23:15:00	20,9	53,9	21,0	55,1	No
		1.3.2014 23:30	23:30:00	20,8	55,3	20,9	55,1	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

			1.3.2014 23:45	23:45:00	20,8	56,2	20,9	55,1	No
			2.3.2014 0:00	0:00:00	20,8	55,1	20,8	55,1	No
			2.3.2014 0:15	0:15:00	20,7	56,1	20,8	55,7	No
			2.3.2014 0:30	0:30:00	20,7	55,6	20,8	55,8	No
			2.3.2014 0:45	0:45:00	20,7	55	20,7	55,5	No
			2.3.2014 1:00	1:00:00	20,7	55,8	20,7	55,6	No
			2.3.2014 1:15	1:15:00	20,7	55,6	20,7	55,5	No
			2.3.2014 1:30	1:30:00	20,7	56,1	20,7	55,6	No
			2.3.2014 1:45	1:45:00	20,7	55,2	20,7	55,7	No
			2.3.2014 2:00	2:00:00	20,7	56,6	20,7	55,9	No
			2.3.2014 2:15	2:15:00	20,7	55,4	20,7	55,8	No
			2.3.2014 2:30	2:30:00	20,7	55,4	20,7	55,7	No
			2.3.2014 2:45	2:45:00	20,7	55,8	20,7	55,8	No
			2.3.2014 3:00	3:00:00	20,7	55,8	20,7	55,6	No
			2.3.2014 3:15	3:15:00	20,7	56,3	20,7	55,8	No
			2.3.2014 3:30	3:30:00	20,6	55,6	20,7	55,9	No
			2.3.2014 3:45	3:45:00	20,6	56,2	20,7	56,0	No
			2.3.2014 4:00	4:00:00	20,7	55,6	20,7	55,9	No
			2.3.2014 4:15	4:15:00	20,7	56,2	20,7	55,9	No
			2.3.2014 4:30	4:30:00	20,8	56,9	20,7	56,2	No
			2.3.2014 4:45	4:45:00	20,9	57,3	20,8	56,5	No
			2.3.2014 5:00	5:00:00	21	55,1	20,9	56,4	No
			2.3.2014 5:15	5:15:00	20,8	57,4	20,9	56,7	No
			2.3.2014 5:30	5:30:00	21	56,9	20,9	56,7	No
			2.3.2014 5:45	5:45:00	21,1	57,2	21,0	56,7	No
			2.3.2014 6:00	6:00:00	21,2	57,4	21,0	57,2	No
			2.3.2014 6:15	6:15:00	21,2	56,5	21,1	57,0	No
			2.3.2014 6:30	6:30:00	21	56,4	21,1	56,9	No
			2.3.2014 6:45	6:45:00	21,1	56,2	21,1	56,6	No
			2.3.2014 7:00	7:00:00	21,2	56,7	21,1	56,5	No
			2.3.2014 7:15	7:15:00	21,2	55	21,1	56,1	No
			2.3.2014 7:30	7:30:00	21	57,7	21,1	56,4	No
			2.3.2014 7:45	7:45:00	21,1	56,9	21,1	56,6	No
			2.3.2014 8:00	8:00:00	21,2	57,2	21,1	56,7	No
			2.3.2014 8:15	8:15:00	21,2	57,2	21,1	57,3	No
			2.3.2014 8:30	8:30:00	21,2	57,5	21,2	57,2	No
			2.3.2014 8:45	8:45:00	21,2	57,6	21,2	57,4	No
			2.3.2014 9:00	9:00:00	21,3	56,6	21,2	57,2	No
			2.3.2014 9:15	9:15:00	21,2	56,9	21,2	57,2	No
			2.3.2014 9:30	9:30:00	21,2	56	21,2	56,8	No
			2.3.2014 9:45	9:45:00	21,2	56,8	21,2	56,6	No
			2.3.2014 10:00	10:00:00	21,2	56,6	21,2	56,6	No
			2.3.2014 10:15	10:15:00	21,2	56,8	21,2	56,6	No
			2.3.2014 10:30	10:30:00	21,2	56,8	21,2	56,8	No
			2.3.2014 10:45	10:45:00	21,2	55,6	21,2	56,5	No
			2.3.2014 11:00	11:00:00	21,2	56,7	21,2	56,5	No
			2.3.2014 11:15	11:15:00	21,2	57,3	21,2	56,6	No
			2.3.2014 11:30	11:30:00	21,2	57,5	21,2	56,8	No
			2.3.2014 11:45	11:45:00	21,2	57,4	21,2	57,2	No
			2.3.2014 12:00	12:00:00	21,2	56,6	21,2	57,2	No
			2.3.2014 12:15	12:15:00	21,2	57,3	21,2	57,2	No
			2.3.2014 12:30	12:30:00	21,3	56,8	21,2	57,0	No
			2.3.2014 12:45	12:45:00	21,3	57,5	21,3	57,1	No
			2.3.2014 13:00	13:00:00	21,3	56,5	21,3	57,0	No
			2.3.2014 13:15	13:15:00	21,3	56,5	21,3	56,8	No
			2.3.2014 13:30	13:30:00	21,4	56,7	21,3	56,8	No
			2.3.2014 13:45	13:45:00	21,4	56,3	21,4	56,5	No
			2.3.2014 14:00	14:00:00	21,4	57,1	21,4	56,7	No
			2.3.2014 14:15	14:15:00	21,4	51,9	21,4	55,5	No
			2.3.2014 14:30	14:30:00	21,2	56,3	21,4	55,4	No
			2.3.2014 14:45	14:45:00	21,3	55,9	21,3	55,3	No
			2.3.2014 15:00	15:00:00	21,4	55,7	21,3	55,0	No
			2.3.2014 15:15	15:15:00	21,4	55,7	21,3	55,9	No
			2.3.2014 15:30	15:30:00	21,5	54,9	21,4	55,6	No
			2.3.2014 15:45	15:45:00	21,5	54,8	21,5	55,3	No
			2.3.2014 16:00	16:00:00	21,6	54,6	21,5	55,0	No
			2.3.2014 16:15	16:15:00	21,6	55,3	21,6	54,9	No
			2.3.2014 16:30	16:30:00	21,6	56,2	21,6	55,2	No
			2.3.2014 16:45	16:45:00	21,6	55	21,6	55,3	No
			2.3.2014 17:00	17:00:00	21,7	55,8	21,6	55,6	No
			2.3.2014 17:15	17:15:00	21,7	54,9	21,7	55,5	No
			2.3.2014 17:30	17:30:00	21,3	55,4	21,6	55,3	No
			2.3.2014 17:45	17:45:00	21,4	55,9	21,5	55,5	No
			2.3.2014 18:00	18:00:00	21,5	56,2	21,5	55,6	No
			2.3.2014 18:15	18:15:00	21,5	56,9	21,4	56,1	No
			2.3.2014 18:30	18:30:00	21,5	55,9	21,5	56,2	No
			2.3.2014 18:45	18:45:00	21,5	56	21,5	56,3	No
			2.3.2014 19:00	19:00:00	21,6	57,4	21,5	56,6	No
			2.3.2014 19:15	19:15:00	21,6	56	21,6	56,3	No
			2.3.2014 19:30	19:30:00	21,6	56,5	21,6	56,5	No
			2.3.2014 19:45	19:45:00	21,6	55,8	21,6	56,4	No
			2.3.2014 20:00	20:00:00	21,5	55,5	21,6	56,0	No
			2.3.2014 20:15	20:15:00	21,4	51,6	21,5	54,9	No
			2.3.2014 20:30	20:30:00	21,2	57,4	21,4	55,1	No
			2.3.2014 20:45	20:45:00	21,1	55,2	21,3	54,9	No
			2.3.2014 21:00	21:00:00	21,2	57	21,2	55,3	No
			2.3.2014 21:15	21:15:00	21,4	55,7	21,2	56,3	No
			2.3.2014 21:30	21:30:00	21,1	55,1	21,2	55,8	No
			2.3.2014 21:45	21:45:00	21,2	54,5	21,2	55,6	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0092 PHI_14_I0102 PHI_14_I0113 PHI_14_I0122	PHI_14_I0096 PHI_14_I0106 PHI_14_I0116 PHI_14_I0126	2.3.2014 22:00	22:00:00	21,4	53,8	21,3	54,8	Yes
		2.3.2014 22:15	22:15:00	21,1	52,4	21,2	54,0	Yes
		2.3.2014 22:30	22:30:00	21,2	52,9	21,2	53,4	Yes
		2.3.2014 22:45	22:45:00	21,3	51,9	21,3	52,8	Yes
		2.3.2014 23:00	23:00:00	21,4	51,1	21,3	52,1	Yes
		2.3.2014 23:15	23:15:00	21	50,6	21,2	51,6	Yes
		2.3.2014 23:30	23:30:00	21,2	50,1	21,2	50,9	Yes
		2.3.2014 23:45	23:45:00	21,3	46,6	21,2	49,6	Yes
		3.3.2014 0:00	0:00:00	21,1	49,7	21,2	49,3	Yes
		3.3.2014 0:15	0:15:00	21,3	48,8	21,2	48,8	Yes
		3.3.2014 0:30	0:30:00	21,4	47,6	21,3	48,2	Yes
		3.3.2014 0:45	0:45:00	21,1	49	21,2	48,8	Yes
		3.3.2014 1:00	1:00:00	21,3	48,2	21,3	48,4	Yes
		3.3.2014 1:15	1:15:00	21,4	47,2	21,3	48,0	Yes
		3.3.2014 1:30	1:30:00	21,1	48,2	21,2	48,2	Yes
		3.3.2014 1:45	1:45:00	21,3	47,4	21,3	47,8	Yes
		3.3.2014 2:00	2:00:00	21,3	44,8	21,3	46,9	Yes
		3.3.2014 2:15	2:15:00	21,1	47,3	21,2	46,9	Yes
		3.3.2014 2:30	2:30:00	21,3	46,4	21,3	46,5	Yes
		3.3.2014 2:45	2:45:00	21,2	42,9	21,2	45,4	Yes
		3.3.2014 3:00	3:00:00	21,2	46,3	21,2	45,7	Yes
		3.3.2014 3:15	3:15:00	21,3	44,5	21,3	45,0	Yes
		3.3.2014 3:30	3:30:00	21,1	46	21,2	44,9	Yes
		3.3.2014 3:45	3:45:00	21,3	45,2	21,2	45,5	Yes
		3.3.2014 4:00	4:00:00	21,1	43,5	21,2	44,8	Yes
		3.3.2014 4:15	4:15:00	21,2	45,2	21,2	45,0	Yes
		3.3.2014 4:30	4:30:00	21	43,8	21,2	44,4	Yes
		3.3.2014 4:45	4:45:00	21,2	45	21,1	44,4	Yes
		3.3.2014 5:00	5:00:00	21,2	42	21,2	44,0	Yes
		3.3.2014 5:15	5:15:00	21,2	44,4	21,2	43,8	Yes
		3.3.2014 5:30	5:30:00	21	44	21,2	43,9	Yes
		3.3.2014 5:45	5:45:00	21,2	45,1	21,2	43,9	Yes
		3.3.2014 6:00	6:00:00	21,3	45,3	21,2	44,7	Yes
		3.3.2014 6:15	6:15:00	21,5	47,4	21,3	45,5	Yes
		3.3.2014 6:30	6:30:00	21,4	44,5	21,4	45,6	Yes
		3.3.2014 6:45	6:45:00	21,6	45,7	21,5	45,7	Yes
		3.3.2014 7:00	7:00:00	21,6	47,5	21,5	46,3	Yes
		3.3.2014 7:15	7:15:00	21,7	44	21,6	45,4	Yes
		3.3.2014 7:30	7:30:00	21,7	42,3	21,7	44,9	Yes
		3.3.2014 7:45	7:45:00	21,8	42,8	21,7	44,2	Yes
		3.3.2014 8:00	8:00:00	21,9	42,6	21,8	42,9	Yes
		3.3.2014 8:15	8:15:00	22	42,7	21,9	42,6	Yes
		3.3.2014 8:30	8:30:00	22,1	42,6	22,0	42,7	Yes
		3.3.2014 8:45	8:45:00	22,1	41,6	22,0	42,4	Yes
		3.3.2014 9:00	9:00:00	22	40	22,1	41,7	Yes
		3.3.2014 9:15	9:15:00	22,3	42,5	22,1	41,7	Yes
		3.3.2014 9:30	9:30:00	22,6	42,6	22,3	41,7	Yes
		3.3.2014 9:45	9:45:00	22,6	41,5	22,4	41,7	Yes
		3.3.2014 10:00	10:00:00	22,6	42,3	22,5	42,2	Yes
		3.3.2014 10:15	10:15:00	22,5	40,6	22,6	41,8	Yes
		3.3.2014 10:30	10:30:00	22,6	42,3	22,6	41,7	Yes
		3.3.2014 10:45	10:45:00	22,5	40,1	22,6	41,3	Yes
		3.3.2014 11:00	11:00:00	22,7	41,1	22,6	41,0	Yes
		3.3.2014 11:15	11:15:00	22,6	39,3	22,6	40,7	Yes
		3.3.2014 11:30	11:30:00	22,6	40,9	22,6	40,4	Yes
		3.3.2014 11:45	11:45:00	22,5	38,9	22,6	40,1	Yes
		3.3.2014 12:00	12:00:00	22,6	41,2	22,6	40,1	Yes
		3.3.2014 12:15	12:15:00	22,6	39,1	22,6	40,0	Yes
		3.3.2014 12:30	12:30:00	22,7	41,7	22,6	40,2	Yes
		3.3.2014 12:45	12:45:00	22,6	39,3	22,6	40,3	Yes
		3.3.2014 13:00	13:00:00	22,7	41,9	22,7	40,5	Yes
		3.3.2014 13:15	13:15:00	22,7	40,2	22,7	40,8	Yes
		3.3.2014 13:30	13:30:00	22,8	39,8	22,7	40,3	Yes
		3.3.2014 13:45	13:45:00	22,8	39,9	22,8	40,5	Yes
		3.3.2014 14:00	14:00:00	22,7	39,2	22,8	39,8	Yes
		3.3.2014 14:15	14:15:00	22,7	41,9	22,8	40,2	Yes
		3.3.2014 14:30	14:30:00	22,4	37,9	22,7	39,7	Yes
		3.3.2014 14:45	14:45:00	22	36,2	22,5	38,8	No
		3.3.2014 15:00	15:00:00	21,9	44,6	22,3	40,2	No
		3.3.2014 15:15	15:15:00	21,6	45	22,0	40,9	No
		3.3.2014 15:30	15:30:00	21,8	47,8	21,8	43,4	No
		3.3.2014 15:45	15:45:00	21,5	49,1	21,7	46,6	No
		3.3.2014 16:00	16:00:00	21,7	50,2	21,7	48,0	No
		3.3.2014 16:15	16:15:00	21,9	50,5	21,7	49,4	No
		3.3.2014 16:30	16:30:00	22	51,5	21,8	50,3	No
		3.3.2014 16:45	16:45:00	22	52	21,9	51,1	No
		3.3.2014 17:00	17:00:00	21,9	48,5	22,0	50,6	No
		3.3.2014 17:15	17:15:00	21,6	53,8	21,9	51,5	No
		3.3.2014 17:30	17:30:00	21,8	52,3	21,8	51,7	No
		3.3.2014 17:45	17:45:00	21,8	52,9	21,8	51,9	No
		3.3.2014 18:00	18:00:00	21,9	53,2	21,8	53,1	No
		3.3.2014 18:15	18:15:00	21,9	53,1	21,9	52,9	No
		3.3.2014 18:30	18:30:00	21,9	53,6	21,9	53,2	No
		3.3.2014 18:45	18:45:00	21,9	53,8	21,9	53,4	No
		3.3.2014 19:00	19:00:00	21,9	54,3	21,9	53,7	No
		3.3.2014 19:15	19:15:00	21,9	54,6	21,9	54,1	No
		3.3.2014 19:30	19:30:00	21,9	54,6	21,9	54,3	No
		3.3.2014 19:45	19:45:00	21,9	54,7	21,9	54,6	No
		3.3.2014 20:00	20:00:00	21,9	55,1	21,9	54,8	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	3.3.2014 20:15	20:15:00	21,9	55,8	21,9	55,1	No
		3.3.2014 20:30	20:30:00	21,8	52,8	21,9	54,6	No
		3.3.2014 20:45	20:45:00	21,5	55,8	21,8	54,9	No
		3.3.2014 21:00	21:00:00	21,6	55,6	21,7	55,0	No
		3.3.2014 21:15	21:15:00	21,6	55,3	21,6	54,9	No
		3.3.2014 21:30	21:30:00	21,6	55,4	21,6	55,5	No
		3.3.2014 21:45	21:45:00	21,6	55,4	21,6	55,4	No
		3.3.2014 22:00	22:00:00	21,6	55,7	21,6	55,5	Yes
		3.3.2014 22:15	22:15:00	21,6	56,3	21,6	55,7	Yes
		3.3.2014 22:30	22:30:00	21,6	56	21,6	55,9	Yes
		3.3.2014 22:45	22:45:00	21,6	55,1	21,6	55,8	Yes
		3.3.2014 23:00	23:00:00	21,6	56,1	21,6	55,9	Yes
		3.3.2014 23:15	23:15:00	21,6	55,4	21,6	55,7	Yes
		3.3.2014 23:30	23:30:00	21,6	55,6	21,6	55,6	Yes
		3.3.2014 23:45	23:45:00	21,6	56,1	21,6	55,8	Yes
		4.3.2014 0:00	0:00:00	21,6	55,3	21,6	55,6	Yes
		4.3.2014 0:15	0:15:00	21,6	56,4	21,6	55,9	Yes
		4.3.2014 0:30	0:30:00	21,6	55,9	21,6	55,9	Yes
		4.3.2014 0:45	0:45:00	21,6	55,9	21,6	55,9	Yes
		4.3.2014 1:00	1:00:00	21,6	55,7	21,6	56,0	Yes
		4.3.2014 1:15	1:15:00	21,6	55,6	21,6	55,8	Yes
		4.3.2014 1:30	1:30:00	21,6	55,5	21,6	55,7	Yes
		4.3.2014 1:45	1:45:00	21,6	55,8	21,6	55,7	Yes
		4.3.2014 2:00	2:00:00	21,6	55,4	21,6	55,6	Yes
		4.3.2014 2:15	2:15:00	21,6	55,8	21,6	55,6	Yes
		4.3.2014 2:30	2:30:00	21,6	55,9	21,6	55,7	Yes
		4.3.2014 2:45	2:45:00	21,6	55,8	21,6	55,7	Yes
		4.3.2014 3:00	3:00:00	21,6	56,6	21,6	56,0	Yes
		4.3.2014 3:15	3:15:00	21,5	52,4	21,6	55,2	Yes
		4.3.2014 3:30	3:30:00	21,3	57,5	21,5	55,6	Yes
		4.3.2014 3:45	3:45:00	21,4	52,7	21,5	54,8	Yes
		4.3.2014 4:00	4:00:00	21,2	57,5	21,4	55,0	Yes
PHI_14_0092 PHI_14_0102 PHI_14_0113 PHI_14_0122	PHI_14_0096 PHI_14_0107 PHI_14_0117 PHI_14_0126	4.3.2014 4:15	4:15:00	21,4	55,9	21,3	55,9	Yes
		4.3.2014 4:30	4:30:00	21,2	55,6	21,3	55,4	Yes
		4.3.2014 4:45	4:45:00	21,3	56,1	21,3	56,3	Yes
		4.3.2014 5:00	5:00:00	21,5	55	21,4	55,7	Yes
		4.3.2014 5:15	5:15:00	21,2	55,8	21,3	55,6	Yes
		4.3.2014 5:30	5:30:00	21,3	55	21,3	55,5	Yes
		4.3.2014 5:45	5:45:00	21,2	55,5	21,3	55,3	Yes
		4.3.2014 6:00	6:00:00	21,2	49,8	21,2	54,0	Yes
		4.3.2014 6:15	6:15:00	21	46,7	21,2	51,8	Yes
		4.3.2014 6:30	6:30:00	21,4	55,1	21,2	51,8	Yes
		4.3.2014 6:45	6:45:00	21,5	48,6	21,3	50,1	Yes
		4.3.2014 7:00	7:00:00	21,7	49,4	21,4	50,0	Yes
		4.3.2014 7:15	7:15:00	21,7	50,8	21,6	51,0	Yes
		4.3.2014 7:30	7:30:00	21,8	47,7	21,7	49,1	Yes
		4.3.2014 7:45	7:45:00	21,9	49,3	21,8	49,3	Yes
		4.3.2014 8:00	8:00:00	21,9	46,1	21,8	48,5	Yes
		4.3.2014 8:15	8:15:00	22	47,4	21,9	47,6	Yes
		4.3.2014 8:30	8:30:00	22	44,8	22,0	46,9	Yes
		4.3.2014 8:45	8:45:00	22,1	47,2	22,0	46,4	Yes
		4.3.2014 9:00	9:00:00	22,2	44,2	22,1	45,9	Yes
		4.3.2014 9:15	9:15:00	22,2	43,5	22,1	44,9	Yes
		4.3.2014 9:30	9:30:00	22,3	45,1	22,2	45,0	Yes
		4.3.2014 9:45	9:45:00	22,3	43	22,3	44,0	Yes
		4.3.2014 10:00	10:00:00	22,4	48	22,3	44,9	Yes
		4.3.2014 10:15	10:15:00	22,4	44,4	22,4	45,1	Yes
		4.3.2014 10:30	10:30:00	22,4	43,4	22,4	44,7	Yes
		4.3.2014 10:45	10:45:00	22,5	44,2	22,4	45,0	Yes
		4.3.2014 11:00	11:00:00	22,7	45,2	22,5	44,3	Yes
		4.3.2014 11:15	11:15:00	22,6	42,6	22,6	43,9	Yes
		4.3.2014 11:30	11:30:00	22,7	45,5	22,6	44,4	Yes
		4.3.2014 11:45	11:45:00	22,7	42,2	22,7	43,9	Yes
		4.3.2014 12:00	12:00:00	22,6	41,4	22,7	42,9	Yes
		4.3.2014 12:15	12:15:00	22,7	44,2	22,7	43,3	Yes
		4.3.2014 12:30	12:30:00	22,6	46,2	22,7	43,5	Yes
		4.3.2014 12:45	12:45:00	22,6	42,7	22,6	43,6	Yes
		4.3.2014 13:00	13:00:00	22,5	43,8	22,6	44,2	Yes
		4.3.2014 13:15	13:15:00	22,8	45,7	22,6	44,6	Yes
		4.3.2014 13:30	13:30:00	22,9	49,1	22,7	45,3	Yes
		4.3.2014 13:45	13:45:00	23	50,4	22,8	47,3	Yes
		4.3.2014 14:00	14:00:00	23	50,7	22,9	49,0	Yes
		4.3.2014 14:15	14:15:00	23	45,2	23,0	48,9	Yes
		4.3.2014 14:30	14:30:00	23,1	49,2	23,0	48,9	Yes
		4.3.2014 14:45	14:45:00	23	45,3	23,0	47,6	Yes
		4.3.2014 15:00	15:00:00	23,1	49,2	23,1	47,2	Yes
		4.3.2014 15:15	15:15:00	22,9	44,7	23,0	47,1	Yes
		4.3.2014 15:30	15:30:00	23,1	45,5	23,0	46,2	Yes
		4.3.2014 15:45	15:45:00	22,8	42,1	23,0	45,4	Yes
		4.3.2014 16:00	16:00:00	22,4	40,3	22,8	43,2	Yes
		4.3.2014 16:15	16:15:00	22	44,7	22,6	43,2	No
		4.3.2014 16:30	16:30:00	22,2	50	22,4	44,3	No
		4.3.2014 16:45	16:45:00	22,1	51	22,2	46,5	No
		4.3.2014 17:00	17:00:00	22,3	53,4	22,2	49,8	No
		4.3.2014 17:15	17:15:00	22,5	54,1	22,3	52,1	No
		4.3.2014 17:30	17:30:00	22,6	53,8	22,4	53,1	No
		4.3.2014 17:45	17:45:00	22,7	54,6	22,5	54,0	No
		4.3.2014 18:00	18:00:00	22,7	54,9	22,6	54,4	No
		4.3.2014 18:15	18:15:00	22,8	55	22,7	54,6	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	4.3.2014 18:30	18:30:00	22,8	55,7	22,8	55,1	No
		4.3.2014 18:45	18:45:00	22,7	51,6	22,8	54,3	No
		4.3.2014 19:00	19:00:00	22,3	56,6	22,7	54,7	No
		4.3.2014 19:15	19:15:00	22,4	55,8	22,6	54,9	No
		4.3.2014 19:30	19:30:00	22,5	55	22,5	54,8	No
		4.3.2014 19:45	19:45:00	22,5	55,1	22,4	55,6	No
		4.3.2014 20:00	20:00:00	22,5	54,2	22,5	55,0	No
		4.3.2014 20:15	20:15:00	22,5	54,7	22,5	54,8	No
		4.3.2014 20:30	20:30:00	22,5	54,5	22,5	54,6	No
		4.3.2014 20:45	20:45:00	22,6	54,5	22,5	54,5	No
		4.3.2014 21:00	21:00:00	22,6	54,7	22,6	54,6	No
		4.3.2014 21:15	21:15:00	22,6	54	22,6	54,4	No
		4.3.2014 21:30	21:30:00	22,5	54,7	22,6	54,5	No
		4.3.2014 21:45	21:45:00	22,6	54,5	22,6	54,5	No
		4.3.2014 22:00	22:00:00	22,5	55,2	22,6	54,6	No
		4.3.2014 22:15	22:15:00	22,5	54,1	22,5	54,6	Yes
		4.3.2014 22:30	22:30:00	22,4	53,8	22,5	54,4	Yes
		4.3.2014 22:45	22:45:00	22,4	55,2	22,5	54,6	Yes
		4.3.2014 23:00	23:00:00	22,4	54,2	22,4	54,3	Yes
		4.3.2014 23:15	23:15:00	22,4	53,5	22,4	54,2	Yes
		4.3.2014 23:30	23:30:00	22,3	54,7	22,4	54,4	Yes
		4.3.2014 23:45	23:45:00	22,3	53,8	22,4	54,1	Yes
		5.3.2014 0:00	0:00:00	22,3	54,7	22,3	54,2	Yes
		5.3.2014 0:15	0:15:00	22,3	54,3	22,3	54,4	Yes
		5.3.2014 0:30	0:30:00	22,3	54,7	22,3	54,4	Yes
		5.3.2014 0:45	0:45:00	22,3	54,7	22,3	54,6	Yes
		5.3.2014 1:00	1:00:00	22,3	54,5	22,3	54,6	Yes
		5.3.2014 1:15	1:15:00	22,3	55,1	22,3	54,8	Yes
		5.3.2014 1:30	1:30:00	22,3	54,4	22,3	54,7	Yes
		5.3.2014 1:45	1:45:00	22,3	55,1	22,3	54,8	Yes
		5.3.2014 2:00	2:00:00	22,3	54,5	22,3	54,8	Yes
		5.3.2014 2:15	2:15:00	22,3	55,2	22,3	54,8	Yes
		5.3.2014 2:30	2:30:00	22,2	54,9	22,3	54,9	Yes
		5.3.2014 2:45	2:45:00	22,3	54,4	22,3	54,8	Yes
		5.3.2014 3:00	3:00:00	22,2	55	22,3	54,9	Yes
		5.3.2014 3:15	3:15:00	22,3	54,3	22,3	54,7	Yes
		5.3.2014 3:30	3:30:00	22,3	55,7	22,3	54,9	Yes
		5.3.2014 3:45	3:45:00	22	50,8	22,2	54,0	Yes
		5.3.2014 4:00	4:00:00	22	56,9	22,2	54,4	Yes
		5.3.2014 4:15	4:15:00	22,2	55,7	22,1	54,8	Yes
		5.3.2014 4:30	4:30:00	22,3	56,3	22,1	54,9	Yes
		5.3.2014 4:45	4:45:00	22,2	51,8	22,2	55,2	Yes
		5.3.2014 5:00	5:00:00	21,9	57,3	22,2	55,3	Yes
		5.3.2014 5:15	5:15:00	22,1	55,1	22,1	55,1	Yes
		5.3.2014 5:30	5:30:00	22,2	55,3	22,1	54,9	Yes
		5.3.2014 5:45	5:45:00	22,1	51,1	22,1	54,7	Yes
		5.3.2014 6:00	6:00:00	22	47,9	22,1	52,4	Yes
		5.3.2014 6:15	6:15:00	22,1	53,3	22,1	51,9	Yes
		5.3.2014 6:30	6:30:00	22,1	47,5	22,1	50,0	Yes
		5.3.2014 6:45	6:45:00	22	45,3	22,1	48,5	Yes
		5.3.2014 7:00	7:00:00	22,2	47,4	22,1	48,4	Yes
		5.3.2014 7:15	7:15:00	22,1	44,9	22,1	46,3	Yes
		5.3.2014 7:30	7:30:00	22,1	44,3	22,1	45,5	Yes
		5.3.2014 7:45	7:45:00	22,1	44,2	22,1	45,2	Yes
		5.3.2014 8:00	8:00:00	22,1	41,7	22,1	43,8	Yes
		5.3.2014 8:15	8:15:00	22	44,2	22,1	43,6	Yes
		5.3.2014 8:30	8:30:00	22,2	45,1	22,1	43,8	Yes
		5.3.2014 8:45	8:45:00	21,8	44,5	22,0	43,9	Yes
		5.3.2014 9:00	9:00:00	22,3	47,4	22,1	45,3	Yes
		5.3.2014 9:15	9:15:00	22,1	42,7	22,1	44,9	Yes
		5.3.2014 9:30	9:30:00	21,9	40,2	22,0	43,7	Yes
		5.3.2014 9:45	9:45:00	21,9	40,7	22,1	42,8	Yes
		5.3.2014 10:00	10:00:00	22,2	45,8	22,0	42,4	Yes
		5.3.2014 10:15	10:15:00	22,1	42,3	22,0	42,3	Yes
		5.3.2014 10:30	10:30:00	22	40,9	22,1	42,4	Yes
		5.3.2014 10:45	10:45:00	22	40,8	22,1	42,5	Yes
		5.3.2014 11:00	11:00:00	22,3	44,5	22,1	42,1	Yes
		5.3.2014 11:15	11:15:00	22,2	40,3	22,1	41,6	Yes
		5.3.2014 11:30	11:30:00	21,9	39,1	22,1	41,2	Yes
		5.3.2014 11:45	11:45:00	22	41,2	22,1	41,3	Yes
		5.3.2014 12:00	12:00:00	22,3	41,3	22,1	40,5	Yes
		5.3.2014 12:15	12:15:00	22,1	39,9	22,1	40,4	Yes
		5.3.2014 12:30	12:30:00	22,1	38,3	22,1	40,2	Yes
		5.3.2014 12:45	12:45:00	22,1	38,2	22,2	39,4	Yes
		5.3.2014 13:00	13:00:00	22,4	45,7	22,2	40,5	Yes
		5.3.2014 13:15	13:15:00	22,3	39,6	22,2	40,5	Yes
		5.3.2014 13:30	13:30:00	22,2	38,7	22,3	40,6	Yes
		5.3.2014 13:45	13:45:00	22,1	38,9	22,3	40,7	Yes
		5.3.2014 14:00	14:00:00	22,3	38,4	22,2	38,9	Yes
		5.3.2014 14:15	14:15:00	22,5	39,7	22,3	38,9	Yes
		5.3.2014 14:30	14:30:00	22,4	37,4	22,3	38,6	Yes
		5.3.2014 14:45	14:45:00	22,5	40,7	22,4	39,1	Yes
		5.3.2014 15:00	15:00:00	22,5	38	22,5	39,0	Yes
		5.3.2014 15:15	15:15:00	22,3	37,4	22,4	38,4	Yes
		5.3.2014 15:30	15:30:00	22,4	39,1	22,4	38,8	Yes
		5.3.2014 15:45	15:45:00	22,6	40,1	22,5	38,7	Yes
		5.3.2014 16:00	16:00:00	22,2	38,7	22,4	38,8	Yes
		5.3.2014 16:15	16:15:00	22,2	43,2	22,4	40,3	Yes
		5.3.2014 16:30	16:30:00	22,6	45	22,4	41,8	Yes
PHI_14_I0092 PHI_14_I0102 PHI_14_I0113 PHI_14_I0122	PHI_14_I0096 PHI_14_I0107 PHI_14_I0117 PHI_14_I0126							

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	5.3.2014 16:45	16:45:00	22,2	47,6	22,3	43,6	No
	5.3.2014 17:00	17:00:00	22,5	50,5	22,4	46,6	No
	5.3.2014 17:15	17:15:00	22,7	51,2	22,5	48,6	No
	5.3.2014 17:30	17:30:00	22,8	51,5	22,6	50,2	No
	5.3.2014 17:45	17:45:00	22,8	49,4	22,7	50,7	No
	5.3.2014 18:00	18:00:00	22,3	52,8	22,7	51,2	No
	5.3.2014 18:15	18:15:00	22,5	53,1	22,6	51,7	No
	5.3.2014 18:30	18:30:00	22,6	53,6	22,6	52,2	No
	5.3.2014 18:45	18:45:00	22,7	54,3	22,5	53,5	No
	5.3.2014 19:00	19:00:00	22,7	53,8	22,6	53,7	No
	5.3.2014 19:15	19:15:00	22,8	53,9	22,7	53,9	No
	5.3.2014 19:30	19:30:00	22,7	53,4	22,7	53,9	No
	5.3.2014 19:45	19:45:00	22,7	53,7	22,7	53,7	No
	5.3.2014 20:00	20:00:00	22,7	53,6	22,7	53,7	No
	5.3.2014 20:15	20:15:00	22,7	54,7	22,7	53,9	No
	5.3.2014 20:30	20:30:00	22,7	55,1	22,7	54,3	No
	5.3.2014 20:45	20:45:00	22,7	55,6	22,7	54,8	No
	5.3.2014 21:00	21:00:00	22,7	56,1	22,7	55,4	No
	5.3.2014 21:15	21:15:00	22,6	55,3	22,7	55,5	No
	5.3.2014 21:30	21:30:00	22,6	55,9	22,7	55,7	No
	5.3.2014 21:45	21:45:00	22,6	56,4	22,6	55,9	No
	5.3.2014 22:00	22:00:00	22,6	55,3	22,6	55,7	Yes
	5.3.2014 22:15	22:15:00	22,6	54,8	22,6	55,6	Yes
	5.3.2014 22:30	22:30:00	22,5	54	22,6	55,1	Yes
	5.3.2014 22:45	22:45:00	22,4	55,5	22,5	54,9	Yes
	5.3.2014 23:00	23:00:00	22,4	54,8	22,5	54,8	Yes
	5.3.2014 23:15	23:15:00	22,4	53,4	22,4	54,4	Yes
	5.3.2014 23:30	23:30:00	22,3	54,7	22,4	54,6	Yes
	5.3.2014 23:45	23:45:00	22,3	55,1	22,4	54,5	Yes
	6.3.2014 0:00	0:00:00	22,3	55,1	22,3	54,6	Yes
	6.3.2014 0:15	0:15:00	22,3	54,6	22,3	54,9	Yes
	6.3.2014 0:30	0:30:00	22,3	55,3	22,3	55,0	Yes
	6.3.2014 0:45	0:45:00	22,3	54,7	22,3	54,9	Yes
	6.3.2014 1:00	1:00:00	22,3	53,4	22,3	54,5	Yes
	6.3.2014 1:15	1:15:00	22,3	54,5	22,3	54,5	Yes
	6.3.2014 1:30	1:30:00	22,3	55,1	22,3	54,4	Yes
	6.3.2014 1:45	1:45:00	22,3	54,3	22,3	54,3	Yes
	6.3.2014 2:00	2:00:00	22,3	55	22,3	54,7	Yes
	6.3.2014 2:15	2:15:00	22,2	54,7	22,3	54,8	Yes
	6.3.2014 2:30	2:30:00	22,2	55,6	22,3	54,9	Yes
	6.3.2014 2:45	2:45:00	22,2	54,4	22,2	54,9	Yes
	6.3.2014 3:00	3:00:00	22,2	55,5	22,2	55,1	Yes
	6.3.2014 3:15	3:15:00	22,2	54,7	22,2	55,1	Yes
	6.3.2014 3:30	3:30:00	22,3	54,9	22,2	54,9	Yes
	6.3.2014 3:45	3:45:00	22,3	55,7	22,3	55,2	Yes
	6.3.2014 4:00	4:00:00	22,1	51,5	22,2	54,2	Yes
	6.3.2014 4:15	4:15:00	22	56,8	22,2	54,7	Yes
	6.3.2014 4:30	4:30:00	22,2	54,9	22,2	54,7	Yes
	6.3.2014 4:45	4:45:00	22,2	55,7	22,1	54,7	Yes
	6.3.2014 5:00	5:00:00	22,3	55,6	22,2	55,8	Yes
	6.3.2014 5:15	5:15:00	22,4	55,2	22,3	55,4	Yes
	6.3.2014 5:30	5:30:00	22,5	55,8	22,4	55,6	Yes
	6.3.2014 5:45	5:45:00	22,2	54,4	22,4	55,3	Yes
	6.3.2014 6:00	6:00:00	22,3	52,1	22,4	54,4	Yes
	6.3.2014 6:15	6:15:00	22,1	47,4	22,3	52,4	Yes
	6.3.2014 6:30	6:30:00	21,9	44,7	22,1	49,7	Yes
	6.3.2014 6:45	6:45:00	22	51,7	22,1	49,0	Yes
	6.3.2014 7:00	7:00:00	22	45,7	22,0	47,4	Yes
	6.3.2014 7:15	7:15:00	22	44,1	22,0	46,6	Yes
	6.3.2014 7:30	7:30:00	22,1	46,6	22,0	47,0	Yes
	6.3.2014 7:45	7:45:00	22,2	43,7	22,1	45,0	Yes
	6.3.2014 8:00	8:00:00	22,1	42,6	22,1	44,3	Yes
	6.3.2014 8:15	8:15:00	22,1	40,4	22,1	43,3	Yes
	6.3.2014 8:30	8:30:00	22,4	46,1	22,2	43,2	Yes
	6.3.2014 8:45	8:45:00	22,1	41	22,2	42,5	Yes
	6.3.2014 9:00	9:00:00	22	39	22,2	41,6	Yes
	6.3.2014 9:15	9:15:00	22,2	43,5	22,2	42,4	Yes
	6.3.2014 9:30	9:30:00	22,4	41	22,2	41,1	Yes
	6.3.2014 9:45	9:45:00	22,3	40,4	22,2	41,0	Yes
	6.3.2014 10:00	10:00:00	22,5	42,7	22,4	41,9	Yes
	6.3.2014 10:15	10:15:00	22,4	40,4	22,4	41,1	Yes
	6.3.2014 10:30	10:30:00	22,3	40	22,4	40,9	Yes
	6.3.2014 10:45	10:45:00	22,5	41,2	22,4	41,1	Yes
	6.3.2014 11:00	11:00:00	22,4	40,7	22,4	40,6	Yes
	6.3.2014 11:15	11:15:00	22,5	39,6	22,4	40,4	Yes
	6.3.2014 11:30	11:30:00	22,6	43,5	22,5	41,3	Yes
	6.3.2014 11:45	11:45:00	22,7	41,7	22,6	41,4	Yes
	6.3.2014 12:00	12:00:00	22,5	40	22,6	41,2	Yes
	6.3.2014 12:15	12:15:00	22,5	39,8	22,6	41,3	Yes
	6.3.2014 12:30	12:30:00	22,8	42,2	22,6	40,9	Yes
	6.3.2014 12:45	12:45:00	22,9	40,5	22,7	40,6	Yes
	6.3.2014 13:00	13:00:00	23	41,5	22,8	41,0	Yes
	6.3.2014 13:15	13:15:00	23,1	39,5	23,0	40,9	Yes
	6.3.2014 13:30	13:30:00	23,2	40,2	23,1	40,4	Yes
	6.3.2014 13:45	13:45:00	23,1	38,4	23,1	39,9	Yes
	6.3.2014 14:00	14:00:00	23,2	39,4	23,2	39,4	Yes
	6.3.2014 14:15	14:15:00	23,3	38,4	23,2	39,1	Yes
	6.3.2014 14:30	14:30:00	23,2	37,6	23,2	38,5	Yes
	6.3.2014 14:45	14:45:00	23,3	39,2	23,3	38,7	Yes
PHI_14_I0093 PHI_14_I0103 PHI_14_I0113 PHI_14_I0122	PHI_14_I0097 PHI_14_I0108 PHI_14_I0118 PHI_14_I0127						

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	6.3.2014 15:00	15:00:00	23,2	36,6	23,3	38,0	Yes
		6.3.2014 15:15	15:15:00	23,4	38,2	23,3	37,9	Yes
		6.3.2014 15:30	15:30:00	23,2	36,3	23,3	37,6	Yes
		6.3.2014 15:45	15:45:00	23,3	37,6	23,3	37,2	Yes
		6.3.2014 16:00	16:00:00	23,2	37,2	23,3	37,3	Yes
		6.3.2014 16:15	16:15:00	23,2	37,1	23,2	37,1	Yes
		6.3.2014 16:30	16:30:00	23,1	37,7	23,2	37,4	Yes
		6.3.2014 16:45	16:45:00	23	39	23,1	37,8	No
		6.3.2014 17:00	17:00:00	22,4	37,1	22,9	37,7	No
		6.3.2014 17:15	17:15:00	22,6	45,5	22,8	39,8	No
		6.3.2014 17:30	17:30:00	22,8	47	22,7	42,2	No
		6.3.2014 17:45	17:45:00	22,9	48,2	22,7	44,5	No
		6.3.2014 18:00	18:00:00	23,1	49,3	22,9	47,5	No
		6.3.2014 18:15	18:15:00	23,1	49,8	23,0	48,6	No
		6.3.2014 18:30	18:30:00	23,1	50,4	23,1	49,4	No
		6.3.2014 18:45	18:45:00	23,2	51	23,1	50,1	No
		6.3.2014 19:00	19:00:00	23,2	51,7	23,2	50,7	No
		6.3.2014 19:15	19:15:00	23,2	52	23,2	51,3	No
		6.3.2014 19:30	19:30:00	23,2	51,1	23,2	51,5	No
		6.3.2014 19:45	19:45:00	23,1	51,4	23,2	51,6	No
		6.3.2014 20:00	20:00:00	23,1	51,2	23,2	51,4	No
		6.3.2014 20:15	20:15:00	23,1	51,9	23,1	51,4	No
		6.3.2014 20:30	20:30:00	23	52,7	23,1	51,8	No
		6.3.2014 20:45	20:45:00	23	53,6	23,1	52,4	No
		6.3.2014 21:00	21:00:00	22,9	53,7	23,0	53,0	No
		6.3.2014 21:15	21:15:00	22,9	53,6	23,0	53,4	No
		6.3.2014 21:30	21:30:00	22,9	54,2	22,9	53,8	No
		6.3.2014 21:45	21:45:00	22,9	54,8	22,9	54,1	No
		6.3.2014 22:00	22:00:00	22,8	55,2	22,9	54,5	Yes
		6.3.2014 22:15	22:15:00	22,8	54,2	22,9	54,6	Yes
		6.3.2014 22:30	22:30:00	22,7	54,1	22,8	54,6	Yes
		6.3.2014 22:45	22:45:00	22,7	54,8	22,8	54,6	Yes
		6.3.2014 23:00	23:00:00	22,7	55,5	22,7	54,7	Yes
		6.3.2014 23:15	23:15:00	22,6	53,1	22,7	54,4	Yes
		6.3.2014 23:30	23:30:00	22,5	54,3	22,6	54,4	Yes
		6.3.2014 23:45	23:45:00	22,5	55,2	22,6	54,5	Yes
		7.3.2014 0:00	0:00:00	22,5	54,7	22,5	54,3	Yes
		7.3.2014 0:15	0:15:00	22,5	55,3	22,5	54,9	Yes
		7.3.2014 0:30	0:30:00	22,5	54,4	22,5	54,9	Yes
		7.3.2014 0:45	0:45:00	22,4	55,5	22,5	55,0	Yes
		7.3.2014 1:00	1:00:00	22,4	54,4	22,5	54,9	Yes
		7.3.2014 1:15	1:15:00	22,4	55,5	22,4	55,0	Yes
		7.3.2014 1:30	1:30:00	22,4	54,7	22,4	55,0	Yes
		7.3.2014 1:45	1:45:00	22,4	55,3	22,4	55,0	Yes
		7.3.2014 2:00	2:00:00	22,3	54,7	22,4	55,1	Yes
		7.3.2014 2:15	2:15:00	22,3	55,2	22,4	55,0	Yes
		7.3.2014 2:30	2:30:00	22,3	55,7	22,3	55,2	Yes
		7.3.2014 2:45	2:45:00	22,3	54,2	22,3	55,0	Yes
		7.3.2014 3:00	3:00:00	22,3	55,4	22,3	55,1	Yes
		7.3.2014 3:15	3:15:00	22,4	54,6	22,3	55,0	Yes
		7.3.2014 3:30	3:30:00	22,2	52,8	22,3	54,3	Yes
		7.3.2014 3:45	3:45:00	22,2	55,6	22,3	54,6	Yes
		7.3.2014 4:00	4:00:00	22,4	55,2	22,3	54,6	Yes
		7.3.2014 4:15	4:15:00	22,5	55,8	22,3	54,9	Yes
		7.3.2014 4:30	4:30:00	22,4	50,6	22,4	54,3	Yes
		7.3.2014 4:45	4:45:00	22,2	56,3	22,4	54,5	Yes
		7.3.2014 5:00	5:00:00	22,4	54,5	22,4	54,3	Yes
		7.3.2014 5:15	5:15:00	22,4	54,8	22,4	54,1	Yes
		7.3.2014 5:30	5:30:00	22,5	55,5	22,4	55,3	Yes
		7.3.2014 5:45	5:45:00	22,5	51,1	22,5	54,0	Yes
		7.3.2014 6:00	6:00:00	22,3	46,2	22,4	51,9	Yes
		7.3.2014 6:15	6:15:00	22,4	47,6	22,4	50,1	Yes
		7.3.2014 6:30	6:30:00	22,4	46,2	22,4	47,8	Yes
		7.3.2014 6:45	6:45:00	22,5	45,2	22,4	46,3	Yes
		7.3.2014 7:00	7:00:00	22,4	47,8	22,4	46,7	Yes
		7.3.2014 7:15	7:15:00	22,4	43,3	22,4	45,6	Yes
		7.3.2014 7:30	7:30:00	22,5	45,6	22,5	45,5	Yes
		7.3.2014 7:45	7:45:00	22,4	42,3	22,4	44,8	Yes
		7.3.2014 8:00	8:00:00	22,5	43,2	22,5	43,6	Yes
		7.3.2014 8:15	8:15:00	22,5	40,9	22,5	43,0	Yes
		7.3.2014 8:30	8:30:00	22,5	41,5	22,5	42,0	Yes
		7.3.2014 8:45	8:45:00	22,4	39	22,5	41,2	Yes
		7.3.2014 9:00	9:00:00	22,5	40	22,5	40,4	Yes
		7.3.2014 9:15	9:15:00	22,5	41,1	22,5	40,4	Yes
		7.3.2014 9:30	9:30:00	22,5	39,8	22,5	40,0	Yes
		7.3.2014 9:45	9:45:00	22,6	42,6	22,5	40,9	Yes
		7.3.2014 10:00	10:00:00	22,8	41	22,6	41,1	Yes
		7.3.2014 10:15	10:15:00	23	39,5	22,7	40,7	Yes
		7.3.2014 10:30	10:30:00	23,1	38,8	22,9	40,5	Yes
		7.3.2014 10:45	10:45:00	23,3	40,4	23,1	39,9	Yes
		7.3.2014 11:00	11:00:00	23,3	38,5	23,2	39,3	Yes
		7.3.2014 11:15	11:15:00	23,4	37,7	23,3	38,9	Yes
		7.3.2014 11:30	11:30:00	23,5	37,4	23,4	38,5	Yes
		7.3.2014 11:45	11:45:00	23,4	37,4	23,4	37,8	Yes
		7.3.2014 12:00	12:00:00	23,4	37,7	23,4	37,6	Yes
		7.3.2014 12:15	12:15:00	23,7	40,9	23,5	38,4	Yes
		7.3.2014 12:30	12:30:00	23,8	41,2	23,6	39,3	Yes
		7.3.2014 12:45	12:45:00	23,8	41,2	23,7	40,3	Yes
		7.3.2014 13:00	13:00:00	23,8	38,9	23,8	40,6	Yes
PHI_14_10093 PHI_14_10103 PHI_14_10113 PHI_14_10114 PHI_14_10122 PHI_14_10123	PHI_14_10097 PHI_14_10108 PHI_14_10118 PHI_14_10127	7.3.2014 13:00	13:00:00	23,8	38,9	23,8	40,6	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

	7.3.2014 13:15	13:15:00	23,9	41,4	23,8	40,7	Yes
	7.3.2014 13:30	13:30:00	23,9	41,9	23,9	40,9	Yes
	7.3.2014 13:45	13:45:00	23,9	38,7	23,9	40,2	Yes
	7.3.2014 14:00	14:00:00	24	40	23,9	40,5	Yes
	7.3.2014 14:15	14:15:00	24,1	40,3	24,0	40,2	Yes
	7.3.2014 14:30	14:30:00	24	41,1	24,0	40,0	Yes
	7.3.2014 14:45	14:45:00	23,6	37,1	23,9	39,6	No
	7.3.2014 15:00	15:00:00	23,3	38,4	23,8	39,2	No
	7.3.2014 15:15	15:15:00	22,9	39,2	23,5	39,0	No
	7.3.2014 15:30	15:30:00	23	44,6	23,2	39,8	No
	7.3.2014 15:45	15:45:00	22,6	42,8	23,0	41,3	No
	7.3.2014 16:00	16:00:00	22,7	47,8	22,8	43,6	No
	7.3.2014 16:15	16:15:00	22,9	48,9	22,8	46,0	No
	7.3.2014 16:30	16:30:00	23	49,3	22,8	47,2	No
	7.3.2014 16:45	16:45:00	23,1	50,1	22,9	49,0	No
	7.3.2014 17:00	17:00:00	23,1	50	23,0	49,6	No
	7.3.2014 17:15	17:15:00	23,2	50,8	23,1	50,1	No
	7.3.2014 17:30	17:30:00	23,2	51,6	23,2	50,6	No
	7.3.2014 17:45	17:45:00	23,2	51,9	23,2	51,1	No
	7.3.2014 18:00	18:00:00	23,2	52,8	23,2	51,8	No
	7.3.2014 18:15	18:15:00	23,2	53	23,2	52,3	No
	7.3.2014 18:30	18:30:00	23,2	53,9	23,2	52,9	No
	7.3.2014 18:45	18:45:00	23,2	54,4	23,2	53,5	No
	7.3.2014 19:00	19:00:00	23,1	54,8	23,2	54,0	No
	7.3.2014 19:15	19:15:00	23,1	55,4	23,2	54,6	No
	7.3.2014 19:30	19:30:00	23,1	54,9	23,1	54,9	No
	7.3.2014 19:45	19:45:00	23	54,7	23,1	55,0	No
	7.3.2014 20:00	20:00:00	23	54,3	23,1	54,8	No
	7.3.2014 20:15	20:15:00	23	54,1	23,0	54,5	No
	7.3.2014 20:30	20:30:00	22,9	55,3	23,0	54,6	No
	7.3.2014 20:45	20:45:00	22,9	55,8	23,0	54,9	No
	7.3.2014 21:00	21:00:00	22,9	56,1	22,9	55,3	No
	7.3.2014 21:15	21:15:00	22,8	55,8	22,9	55,8	No
	7.3.2014 21:30	21:30:00	22,8	55,8	22,9	55,9	No
	7.3.2014 21:45	21:45:00	22,8	55,9	22,8	55,9	No
	7.3.2014 22:00	22:00:00	22,8	55,2	22,8	55,7	No
	7.3.2014 22:15	22:15:00	22,7	55,6	22,8	55,6	No
	7.3.2014 22:30	22:30:00	22,7	55,5	22,8	55,6	No
	7.3.2014 22:45	22:45:00	22,6	56,1	22,7	55,6	No
	7.3.2014 23:00	23:00:00	22,6	56,1	22,7	55,8	No
	7.3.2014 23:15	23:15:00	22,5	55,1	22,6	55,7	No
	7.3.2014 23:30	23:30:00	22,5	55,8	22,6	55,8	No
	7.3.2014 23:45	23:45:00	22,5	56	22,5	55,8	No
	8.3.2014 0:00	0:00:00	22,4	56,5	22,5	55,9	No
	8.3.2014 0:15	0:15:00	22,4	56	22,5	56,1	No
	8.3.2014 0:30	0:30:00	22,4	56,4	22,4	56,2	No
	8.3.2014 0:45	0:45:00	22,4	54,8	22,4	55,9	No
	8.3.2014 1:00	1:00:00	22,4	55,3	22,4	55,6	No
	8.3.2014 1:15	1:15:00	22,4	54,1	22,4	55,2	No
	8.3.2014 1:30	1:30:00	22,4	55,6	22,4	55,0	No
	8.3.2014 1:45	1:45:00	22,3	54,8	22,4	55,0	No
	8.3.2014 2:00	2:00:00	22,3	54	22,4	54,6	No
	8.3.2014 2:15	2:15:00	22,3	54,9	22,3	54,8	No
	8.3.2014 2:30	2:30:00	22,2	54,4	22,3	54,5	No
	8.3.2014 2:45	2:45:00	22,2	54,6	22,3	54,5	No
	8.3.2014 3:00	3:00:00	22,2	53,9	22,2	54,5	No
	8.3.2014 3:15	3:15:00	22,2	55,7	22,2	54,7	No
	8.3.2014 3:30	3:30:00	22,1	53,3	22,2	54,4	No
	8.3.2014 3:45	3:45:00	22,1	54,9	22,2	54,5	No
	8.3.2014 4:00	4:00:00	22,1	53,8	22,1	54,4	No
	8.3.2014 4:15	4:15:00	22,1	54,9	22,1	54,2	No
	8.3.2014 4:30	4:30:00	22	55,5	22,1	54,8	No
	8.3.2014 4:45	4:45:00	22	56,2	22,1	55,1	No
	8.3.2014 5:00	5:00:00	22	55	22,0	55,4	No
	8.3.2014 5:15	5:15:00	22	55,2	22,0	55,5	No
	8.3.2014 5:30	5:30:00	21,9	56,2	22,0	55,7	No
	8.3.2014 5:45	5:45:00	21,9	54,5	22,0	55,2	No
	8.3.2014 6:00	6:00:00	21,9	55,3	21,9	55,3	No
	8.3.2014 6:15	6:15:00	21,9	54,4	21,9	55,1	No
	8.3.2014 6:30	6:30:00	21,8	55,2	21,9	54,9	No
	8.3.2014 6:45	6:45:00	21,7	55,6	21,8	55,1	No
	8.3.2014 7:00	7:00:00	21,7	55,2	21,8	55,1	No
	8.3.2014 7:15	7:15:00	21,7	55	21,7	55,3	No
	8.3.2014 7:30	7:30:00	21,7	56	21,7	55,5	No
	8.3.2014 7:45	7:45:00	21,7	54,4	21,7	55,2	No
	8.3.2014 8:00	8:00:00	21,6	55,3	21,7	55,2	No
	8.3.2014 8:15	8:15:00	21,6	55,7	21,7	55,4	No
	8.3.2014 8:30	8:30:00	21,6	56,5	21,6	55,5	No
	8.3.2014 8:45	8:45:00	21,6	55	21,6	55,6	No
	8.3.2014 9:00	9:00:00	21,5	56,2	21,6	55,9	No
	8.3.2014 9:15	9:15:00	21,5	56	21,6	55,9	No
	8.3.2014 9:30	9:30:00	21,5	55,8	21,5	55,8	No
	8.3.2014 9:45	9:45:00	21,5	55,7	21,5	55,9	No
	8.3.2014 10:00	10:00:00	21,5	56,2	21,5	55,9	No
	8.3.2014 10:15	10:15:00	21,4	56	21,5	55,9	No
	8.3.2014 10:30	10:30:00	21,4	55,8	21,5	55,9	No
	8.3.2014 10:45	10:45:00	21,4	55,8	21,4	56,0	No
	8.3.2014 11:00	11:00:00	21,4	56,2	21,4	56,0	No
	8.3.2014 11:15	11:15:00	21,4	55,8	21,4	55,9	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	8.3.2014 11:30	11:30:00	21,4	56,6	21,4	56,1	No
		8.3.2014 11:45	11:45:00	21,4	56,4	21,4	56,3	No
		8.3.2014 12:00	12:00:00	21,4	55,6	21,4	56,1	No
		8.3.2014 12:15	12:15:00	21,4	54,6	21,4	55,8	No
		8.3.2014 12:30	12:30:00	21,4	54,8	21,4	55,4	No
		8.3.2014 12:45	12:45:00	21,4	55,7	21,4	55,2	No
		8.3.2014 13:00	13:00:00	21,4	56,4	21,4	55,4	No
		8.3.2014 13:15	13:15:00	21,5	57,1	21,4	56,0	No
		8.3.2014 13:30	13:30:00	21,5	55,5	21,5	56,2	No
		8.3.2014 13:45	13:45:00	21,5	56	21,5	56,3	No
		8.3.2014 14:00	14:00:00	21,6	56,5	21,5	56,3	No
		8.3.2014 14:15	14:15:00	21,6	55,7	21,6	55,9	No
		8.3.2014 14:30	14:30:00	21,7	56,1	21,6	56,1	No
		8.3.2014 14:45	14:45:00	21,7	56,5	21,7	56,2	No
		8.3.2014 15:00	15:00:00	21,8	56,6	21,7	56,2	No
		8.3.2014 15:15	15:15:00	21,8	56,4	21,8	56,4	No
		8.3.2014 15:30	15:30:00	21,9	56,4	21,8	56,5	No
		8.3.2014 15:45	15:45:00	21,9	56,4	21,9	56,5	No
		8.3.2014 16:00	16:00:00	22	56,2	21,9	56,4	No
		8.3.2014 16:15	16:15:00	22	56	22,0	56,3	No
		8.3.2014 16:30	16:30:00	22	55,7	22,0	56,1	No
		8.3.2014 16:45	16:45:00	22,1	55,7	22,0	55,9	No
		8.3.2014 17:00	17:00:00	22,1	55,7	22,1	55,8	No
		8.3.2014 17:15	17:15:00	22,1	55,6	22,1	55,7	No
		8.3.2014 17:30	17:30:00	22,1	55,8	22,1	55,7	No
		8.3.2014 17:45	17:45:00	22,2	56	22,1	55,8	No
		8.3.2014 18:00	18:00:00	22,2	56	22,2	55,9	No
		8.3.2014 18:15	18:15:00	22,2	55,9	22,2	55,9	No
		8.3.2014 18:30	18:30:00	22,2	54,8	22,2	55,7	No
		8.3.2014 18:45	18:45:00	22,2	55,2	22,2	55,5	No
		8.3.2014 19:00	19:00:00	22,2	55,4	22,2	55,3	No
		8.3.2014 19:15	19:15:00	22,1	56	22,2	55,4	No
		8.3.2014 19:30	19:30:00	22,2	54,6	22,2	55,3	No
		8.3.2014 19:45	19:45:00	22,1	55,3	22,2	55,3	No
		8.3.2014 20:00	20:00:00	22,1	54,6	22,1	55,1	No
		8.3.2014 20:15	20:15:00	22	55,1	22,1	54,9	No
		8.3.2014 20:30	20:30:00	22	55,5	22,1	55,1	No
		8.3.2014 20:45	20:45:00	22	55,6	22,0	55,2	No
		8.3.2014 21:00	21:00:00	22	55,7	22,0	55,5	No
		8.3.2014 21:15	21:15:00	22	54,5	22,0	55,3	No
		8.3.2014 21:30	21:30:00	21,9	54,9	22,0	55,2	No
		8.3.2014 21:45	21:45:00	21,9	54,9	22,0	55,0	No
		8.3.2014 22:00	22:00:00	21,9	54,6	21,9	54,7	No
		8.3.2014 22:15	22:15:00	21,9	54,2	21,9	54,7	No
		8.3.2014 22:30	22:30:00	21,8	54,3	21,9	54,5	No
		8.3.2014 22:45	22:45:00	21,8	54,8	21,9	54,5	No
		8.3.2014 23:00	23:00:00	21,8	54,7	21,8	54,5	No
		8.3.2014 23:15	23:15:00	21,7	54,4	21,8	54,6	No
		8.3.2014 23:30	23:30:00	21,7	53,8	21,8	54,4	No
		8.3.2014 23:45	23:45:00	21,6	53,4	21,7	54,1	No
		9.3.2014 0:00	0:00:00	21,6	53,9	21,7	53,9	No
		9.3.2014 0:15	0:15:00	21,6	54,1	21,6	53,8	No
		9.3.2014 0:30	0:30:00	21,6	54,4	21,6	54,0	No
		9.3.2014 0:45	0:45:00	21,5	54,7	21,6	54,3	No
		9.3.2014 1:00	1:00:00	21,5	54	21,6	54,3	No
		9.3.2014 1:15	1:15:00	21,4	54,4	21,5	54,4	No
		9.3.2014 1:30	1:30:00	21,4	54,9	21,5	54,5	No
		9.3.2014 1:45	1:45:00	21,4	55,1	21,4	54,6	No
		9.3.2014 2:00	2:00:00	21,4	55,1	21,4	54,9	No
		9.3.2014 2:15	2:15:00	21,3	55,1	21,4	55,1	No
		9.3.2014 2:30	2:30:00	21,3	55,1	21,4	55,1	No
		9.3.2014 2:45	2:45:00	21,3	55	21,3	55,1	No
		9.3.2014 3:00	3:00:00	21,3	55	21,3	55,1	No
		9.3.2014 3:15	3:15:00	21,3	55,1	21,3	55,1	No
		9.3.2014 3:30	3:30:00	21,3	54,8	21,3	55,0	No
		9.3.2014 3:45	3:45:00	21,2	54,5	21,3	54,9	No
		9.3.2014 4:00	4:00:00	21,2	54,1	21,3	54,6	No
		9.3.2014 4:15	4:15:00	21,3	53,2	21,3	54,2	No
		9.3.2014 4:30	4:30:00	21,3	54,9	21,3	54,2	No
		9.3.2014 4:45	4:45:00	21,6	55,2	21,4	54,4	No
		9.3.2014 5:00	5:00:00	21,8	55,2	21,5	54,6	No
		9.3.2014 5:15	5:15:00	21,9	55,1	21,7	55,1	No
		9.3.2014 5:30	5:30:00	21,8	51,6	21,8	54,3	No
		9.3.2014 5:45	5:45:00	21,8	55,8	21,8	54,4	No
		9.3.2014 6:00	6:00:00	22	56,1	21,9	54,7	No
		9.3.2014 6:15	6:15:00	21,8	50,8	21,9	53,6	No
		9.3.2014 6:30	6:30:00	21,8	55,4	21,9	54,5	No
		9.3.2014 6:45	6:45:00	22	56,4	21,9	54,7	No
		9.3.2014 7:00	7:00:00	22,1	56,7	21,9	54,8	No
		9.3.2014 7:15	7:15:00	22,1	56,9	22,0	56,4	No
		9.3.2014 7:30	7:30:00	22,2	57,1	22,1	56,8	No
		9.3.2014 7:45	7:45:00	22,2	57	22,2	56,9	No
		9.3.2014 8:00	8:00:00	22,2	56,9	22,2	57,0	No
		9.3.2014 8:15	8:15:00	22,2	56,6	22,2	56,9	No
		9.3.2014 8:30	8:30:00	22,1	56,5	22,2	56,8	No
		9.3.2014 8:45	8:45:00	22,1	56,7	22,2	56,7	No
		9.3.2014 9:00	9:00:00	22	56,7	22,1	56,6	No
		9.3.2014 9:15	9:15:00	22	56,6	22,1	56,6	No
		9.3.2014 9:30	9:30:00	21,9	56,2	22,0	56,6	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

	9.3.2014 9:45	9:45:00	21,9	55,4	22,0	56,2	No
	9.3.2014 10:00	10:00:00	21,9	55,6	21,9	56,0	No
	9.3.2014 10:15	10:15:00	21,8	55,5	21,9	55,7	No
	9.3.2014 10:30	10:30:00	21,8	55,3	21,9	55,5	No
	9.3.2014 10:45	10:45:00	21,8	54,4	21,8	55,2	No
	9.3.2014 11:00	11:00:00	21,7	55,3	21,8	55,1	No
	9.3.2014 11:15	11:15:00	21,7	55,7	21,8	55,2	No
	9.3.2014 11:30	11:30:00	21,7	56,2	21,7	55,4	No
	9.3.2014 11:45	11:45:00	21,7	56,5	21,7	55,9	No
	9.3.2014 12:00	12:00:00	21,7	56,1	21,7	56,1	No
	9.3.2014 12:15	12:15:00	21,7	56,5	21,7	56,3	No
	9.3.2014 12:30	12:30:00	21,8	57,1	21,7	56,6	No
	9.3.2014 12:45	12:45:00	21,8	56,4	21,8	56,5	No
	9.3.2014 13:00	13:00:00	21,8	57	21,8	56,8	No
	9.3.2014 13:15	13:15:00	21,9	56,3	21,8	56,7	No
	9.3.2014 13:30	13:30:00	21,9	56,7	21,9	56,6	No
	9.3.2014 13:45	13:45:00	21,9	56,9	21,9	56,7	No
	9.3.2014 14:00	14:00:00	22	56,8	21,9	56,7	No
	9.3.2014 14:15	14:15:00	22,1	54,5	22,0	56,2	No
	9.3.2014 14:30	14:30:00	22,1	55,3	22,0	55,9	No
	9.3.2014 14:45	14:45:00	22,1	55,9	22,1	55,6	No
	9.3.2014 15:00	15:00:00	22,2	55,9	22,1	55,4	No
	9.3.2014 15:15	15:15:00	22,2	56,4	22,2	55,9	No
	9.3.2014 15:30	15:30:00	22,3	56,5	22,2	56,2	No
	9.3.2014 15:45	15:45:00	22,3	56,4	22,3	56,3	No
	9.3.2014 16:00	16:00:00	22,4	56,5	22,3	56,5	No
	9.3.2014 16:15	16:15:00	22,4	56,3	22,4	56,4	No
	9.3.2014 16:30	16:30:00	22,5	55,8	22,4	56,3	No
	9.3.2014 16:45	16:45:00	22,5	55,5	22,5	56,0	No
	9.3.2014 17:00	17:00:00	22,5	55,4	22,5	55,8	No
	9.3.2014 17:15	17:15:00	22,5	55,2	22,5	55,5	No
	9.3.2014 17:30	17:30:00	22,6	54,9	22,5	55,3	No
	9.3.2014 17:45	17:45:00	22,6	53,6	22,6	54,8	No
	9.3.2014 18:00	18:00:00	22,6	52,5	22,6	54,1	No
	9.3.2014 18:15	18:15:00	22,7	51,6	22,6	53,2	No
	9.3.2014 18:30	18:30:00	22,6	50,8	22,6	52,1	No
	9.3.2014 18:45	18:45:00	22,6	50,1	22,6	51,3	No
	9.3.2014 19:00	19:00:00	22,6	49,4	22,6	50,5	No
	9.3.2014 19:15	19:15:00	22,6	48,8	22,6	49,8	No
	9.3.2014 19:30	19:30:00	22,6	48,2	22,6	49,1	No
	9.3.2014 19:45	19:45:00	22,6	47,7	22,6	48,5	No
	9.3.2014 20:00	20:00:00	22,5	46,5	22,6	47,8	No
	9.3.2014 20:15	20:15:00	22,5	44,9	22,6	46,8	No
	9.3.2014 20:30	20:30:00	22,1	45,9	22,4	46,3	No
	9.3.2014 20:45	20:45:00	22	42,9	22,3	45,1	No
	9.3.2014 21:00	21:00:00	22	46,7	22,2	45,1	No
	9.3.2014 21:15	21:15:00	22,2	45,6	22,1	45,3	No
	9.3.2014 21:30	21:30:00	21,9	43,8	22,0	44,8	No
	9.3.2014 21:45	21:45:00	22,1	46,2	22,1	45,6	No
	9.3.2014 22:00	22:00:00	22,3	45,6	22,1	45,3	Yes
	9.3.2014 22:15	22:15:00	22,3	45,2	22,2	45,2	Yes
	9.3.2014 22:30	22:30:00	22	44,4	22,2	45,4	Yes
	9.3.2014 22:45	22:45:00	22	45,4	22,2	45,2	Yes
	9.3.2014 23:00	23:00:00	22,2	45,1	22,1	45,0	Yes
	9.3.2014 23:15	23:15:00	22,2	43,3	22,1	44,6	Yes
	9.3.2014 23:30	23:30:00	22,3	43,2	22,2	44,3	Yes
	9.3.2014 23:45	23:45:00	22,3	41,1	22,3	43,2	Yes
	10.3.2014 0:00	0:00:00	21,9	43,6	22,2	42,8	Yes
	10.3.2014 0:15	0:15:00	22,1	43,6	22,2	42,9	Yes
	10.3.2014 0:30	0:30:00	22,2	43,3	22,1	42,9	Yes
	10.3.2014 0:45	0:45:00	22,3	43,1	22,1	43,4	Yes
	10.3.2014 1:00	1:00:00	22,2	40,3	22,2	42,6	Yes
	10.3.2014 1:15	1:15:00	21,9	42	22,2	42,2	Yes
	10.3.2014 1:30	1:30:00	22,1	43,2	22,1	42,2	Yes
	10.3.2014 1:45	1:45:00	22,3	42,7	22,1	42,1	Yes
	10.3.2014 2:00	2:00:00	21,9	40,5	22,1	42,1	Yes
	10.3.2014 2:15	2:15:00	22,1	42,8	22,1	42,3	Yes
	10.3.2014 2:30	2:30:00	22,2	42,4	22,1	42,1	Yes
	10.3.2014 2:45	2:45:00	22,1	38,7	22,1	41,1	Yes
	10.3.2014 3:00	3:00:00	22	41,9	22,1	41,5	Yes
	10.3.2014 3:15	3:15:00	22,2	42	22,1	41,3	Yes
	10.3.2014 3:30	3:30:00	22,2	39,6	22,1	40,6	Yes
	10.3.2014 3:45	3:45:00	22	41,2	22,1	41,2	Yes
	10.3.2014 4:00	4:00:00	22,1	41,5	22,1	41,1	Yes
	10.3.2014 4:15	4:15:00	22,3	41,1	22,2	40,9	Yes
	10.3.2014 4:30	4:30:00	22,1	37,9	22,1	40,4	Yes
	10.3.2014 4:45	4:45:00	22	41	22,1	40,4	Yes
	10.3.2014 5:00	5:00:00	22,2	40,8	22,2	40,2	Yes
	10.3.2014 5:15	5:15:00	22,3	40	22,2	39,9	Yes
	10.3.2014 5:30	5:30:00	22	40	22,1	40,5	Yes
	10.3.2014 5:45	5:45:00	22,3	41,5	22,2	40,6	Yes
	10.3.2014 6:00	6:00:00	22,3	45,2	22,2	41,7	Yes
	10.3.2014 6:15	6:15:00	22,4	41,1	22,3	42,0	Yes
	10.3.2014 6:30	6:30:00	22,5	42,7	22,4	42,6	Yes
	10.3.2014 6:45	6:45:00	22,6	41,9	22,5	42,7	Yes
	10.3.2014 7:00	7:00:00	22,6	41,7	22,5	41,9	Yes
	10.3.2014 7:15	7:15:00	22,7	40,1	22,6	41,6	Yes
	10.3.2014 7:30	7:30:00	22,8	40,1	22,7	41,0	Yes
	10.3.2014 7:45	7:45:00	22,9	39,7	22,8	40,4	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0093 PHI_14_I0103 PHI_14_I0114 PHI_14_I0123	PHI_14_I0097 PHI_14_I0108 PHI_14_I0119 PHI_14_I0127 PHI_14_I0128	10.3.2014 8:00	8:00:00	23	41,3	22,9	40,3	Yes
		10.3.2014 8:15	8:15:00	23	42,1	22,9	40,8	Yes
		10.3.2014 8:30	8:30:00	23,1	42,2	23,0	41,3	Yes
		10.3.2014 8:45	8:45:00	23	37,1	23,0	40,7	Yes
		10.3.2014 9:00	9:00:00	23,1	38,2	23,1	39,9	Yes
		10.3.2014 9:15	9:15:00	23,2	38,7	23,1	39,1	Yes
		10.3.2014 9:30	9:30:00	23,3	41,1	23,2	38,8	Yes
		10.3.2014 9:45	9:45:00	23,4	40,5	23,3	39,6	Yes
		10.3.2014 10:00	10:00:00	23,5	41,2	23,4	40,4	Yes
		10.3.2014 10:15	10:15:00	23,5	42,6	23,4	41,4	Yes
		10.3.2014 10:30	10:30:00	23,7	42,5	23,5	41,7	Yes
		10.3.2014 10:45	10:45:00	23,8	44,4	23,6	42,7	Yes
		10.3.2014 11:00	11:00:00	24	42,6	23,8	43,0	Yes
		10.3.2014 11:15	11:15:00	24,1	42,7	23,9	43,1	Yes
		10.3.2014 11:30	11:30:00	24,1	41,8	24,0	42,9	Yes
		10.3.2014 11:45	11:45:00	24,1	41	24,1	42,0	Yes
		10.3.2014 12:00	12:00:00	24,1	40,3	24,1	41,5	Yes
		10.3.2014 12:15	12:15:00	24,1	39,7	24,1	40,7	Yes
		10.3.2014 12:30	12:30:00	24,1	40,5	24,1	40,4	Yes
		10.3.2014 12:45	12:45:00	24	41,1	24,1	40,4	Yes
		10.3.2014 13:00	13:00:00	24,1	41,1	24,1	40,6	Yes
		10.3.2014 13:15	13:15:00	24,3	41,9	24,1	41,2	Yes
		10.3.2014 13:30	13:30:00	24,3	41,4	24,2	41,4	Yes
		10.3.2014 13:45	13:45:00	24,4	42	24,3	41,6	Yes
		10.3.2014 14:00	14:00:00	24,4	41,5	24,4	41,7	Yes
		10.3.2014 14:15	14:15:00	24,4	41,5	24,4	41,6	Yes
		10.3.2014 14:30	14:30:00	24,5	40,4	24,4	41,4	Yes
		10.3.2014 14:45	14:45:00	24	38,9	24,3	40,6	No
		10.3.2014 15:00	15:00:00	23,6	36,3	24,1	39,3	No
		10.3.2014 15:15	15:15:00	23,2	38,9	23,8	38,6	No
		10.3.2014 15:30	15:30:00	23,1	38,5	23,5	38,2	No
		10.3.2014 15:45	15:45:00	22,8	42,4	23,2	39,0	No
		10.3.2014 16:00	16:00:00	23	44,4	23,0	41,1	No
		10.3.2014 16:15	16:15:00	23,1	46,8	23,0	43,0	No
		10.3.2014 16:30	16:30:00	23	41,6	23,0	43,8	No
		10.3.2014 16:45	16:45:00	22,7	44,7	23,0	44,4	No
		10.3.2014 17:00	17:00:00	22,9	48,9	22,9	45,5	No
		10.3.2014 17:15	17:15:00	23	49,7	22,9	46,2	No
		10.3.2014 17:30	17:30:00	23,1	50,1	22,9	48,4	No
		10.3.2014 17:45	17:45:00	23,2	49,5	23,1	49,6	No
		10.3.2014 18:00	18:00:00	22,6	41,8	23,0	47,8	No
		10.3.2014 18:15	18:15:00	22,6	48,3	22,9	47,4	No
		10.3.2014 18:30	18:30:00	22,7	49,6	22,8	47,3	No
		10.3.2014 18:45	18:45:00	22,8	50,5	22,7	47,6	No
		10.3.2014 19:00	19:00:00	22,9	50,9	22,8	49,8	No
		10.3.2014 19:15	19:15:00	22,9	50,7	22,8	50,4	No
		10.3.2014 19:30	19:30:00	22,9	52	22,9	51,0	No
		10.3.2014 19:45	19:45:00	22,9	51,4	22,9	51,3	No
		10.3.2014 20:00	20:00:00	22,9	51,5	22,9	51,4	No
		10.3.2014 20:15	20:15:00	22,9	52	22,9	51,7	No
		10.3.2014 20:30	20:30:00	22,9	52,4	22,9	51,8	No
		10.3.2014 20:45	20:45:00	22,9	52,9	22,9	52,2	No
		10.3.2014 21:00	21:00:00	22,9	53,3	22,9	52,7	No
		10.3.2014 21:15	21:15:00	22,9	52,8	22,9	52,9	No
		10.3.2014 21:30	21:30:00	22,8	52,9	22,9	53,0	No
		10.3.2014 21:45	21:45:00	22,8	53,6	22,9	53,2	No
		10.3.2014 22:00	22:00:00	22,8	53,8	22,8	53,3	Yes
		10.3.2014 22:15	22:15:00	22,7	53	22,8	53,3	Yes
		10.3.2014 22:30	22:30:00	22,7	52,4	22,8	53,2	Yes
		10.3.2014 22:45	22:45:00	22,7	52,7	22,7	53,0	Yes
		10.3.2014 23:00	23:00:00	22,6	53	22,7	52,8	Yes
		10.3.2014 23:15	23:15:00	22,6	52,6	22,7	52,7	Yes
		10.3.2014 23:30	23:30:00	22,5	53	22,6	52,8	Yes
		10.3.2014 23:45	23:45:00	22,5	53,2	22,6	53,0	Yes
		11.3.2014 0:00	0:00:00	22,5	53,7	22,5	53,1	Yes
		11.3.2014 0:15	0:15:00	22,4	53,7	22,5	53,4	Yes
		11.3.2014 0:30	0:30:00	22,4	54,2	22,5	53,7	Yes
		11.3.2014 0:45	0:45:00	22,4	54,2	22,4	54,0	Yes
		11.3.2014 1:00	1:00:00	22,4	53,9	22,4	54,0	Yes
		11.3.2014 1:15	1:15:00	22,4	53,7	22,4	54,0	Yes
		11.3.2014 1:30	1:30:00	22,3	54,2	22,4	54,0	Yes
		11.3.2014 1:45	1:45:00	22,3	54,2	22,4	54,0	Yes
		11.3.2014 2:00	2:00:00	22,3	54,4	22,3	54,1	Yes
		11.3.2014 2:15	2:15:00	22,3	54	22,3	54,2	Yes
		11.3.2014 2:30	2:30:00	22,3	54	22,3	54,2	Yes
		11.3.2014 2:45	2:45:00	22,3	54,6	22,3	54,3	Yes
		11.3.2014 3:00	3:00:00	22,2	54,8	22,3	54,4	Yes
		11.3.2014 3:15	3:15:00	22,1	50,1	22,2	53,4	Yes
		11.3.2014 3:30	3:30:00	22,1	55,7	22,2	53,8	Yes
		11.3.2014 3:45	3:45:00	22,3	56,7	22,2	54,3	Yes
		11.3.2014 4:00	4:00:00	22,1	50,1	22,2	53,2	Yes
		11.3.2014 4:15	4:15:00	22,1	54,4	22,2	54,2	Yes
		11.3.2014 4:30	4:30:00	22,2	55,6	22,2	54,2	Yes
		11.3.2014 4:45	4:45:00	22,3	56,5	22,2	54,2	Yes
		11.3.2014 5:00	5:00:00	22,4	54,8	22,3	55,3	Yes
		11.3.2014 5:15	5:15:00	22,1	52,9	22,3	55,0	Yes
		11.3.2014 5:30	5:30:00	22,1	55,2	22,2	54,9	Yes
		11.3.2014 5:45	5:45:00	22,3	49,9	22,2	53,2	Yes
		11.3.2014 6:00	6:00:00	22,4	48,6	22,2	51,7	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_0093 PHI_14_0114 PHI_14_0123	PHI_14_0098 PHI_14_0109 PHI_14_0110 PHI_14_0119 PHI_14_0128	11.3.2014 6:15	6:15:00	22,5	50,3	22,3	51,0	Yes
		11.3.2014 6:30	6:30:00	22,6	46,9	22,5	48,9	Yes
		11.3.2014 6:45	6:45:00	22,7	45,2	22,6	47,8	Yes
		11.3.2014 7:00	7:00:00	22,7	44,8	22,6	46,8	Yes
		11.3.2014 7:15	7:15:00	22,9	43,9	22,7	45,2	Yes
		11.3.2014 7:30	7:30:00	22,9	41,8	22,8	43,9	Yes
		11.3.2014 7:45	7:45:00	23	40,6	22,9	42,8	Yes
		11.3.2014 8:00	8:00:00	23	41,7	23,0	42,0	Yes
		11.3.2014 8:15	8:15:00	23,1	40,3	23,0	41,1	Yes
		11.3.2014 8:30	8:30:00	23,1	41,5	23,1	41,0	Yes
		11.3.2014 8:45	8:45:00	23,2	44,1	23,1	41,9	Yes
		11.3.2014 9:00	9:00:00	23,3	43	23,2	42,2	Yes
		11.3.2014 9:15	9:15:00	23,4	42,8	23,3	42,9	Yes
		11.3.2014 9:30	9:30:00	23,4	44	23,3	43,5	Yes
		11.3.2014 9:45	9:45:00	23,5	43,9	23,4	43,4	Yes
		11.3.2014 10:00	10:00:00	23,4	42,1	23,4	43,2	Yes
		11.3.2014 10:15	10:15:00	23,5	43,5	23,5	43,4	Yes
		11.3.2014 10:30	10:30:00	23,5	43,3	23,5	43,2	Yes
		11.3.2014 10:45	10:45:00	23,6	42,6	23,5	42,9	Yes
		11.3.2014 11:00	11:00:00	23,6	43,5	23,6	43,2	Yes
		11.3.2014 11:15	11:15:00	23,6	42,5	23,6	43,0	Yes
		11.3.2014 11:30	11:30:00	23,7	43,6	23,6	43,1	Yes
		11.3.2014 11:45	11:45:00	23,8	41	23,7	42,7	Yes
		11.3.2014 12:00	12:00:00	23,9	42,7	23,8	42,5	Yes
		11.3.2014 12:15	12:15:00	23,9	40,4	23,8	41,9	Yes
		11.3.2014 12:30	12:30:00	24	41,2	23,9	41,3	Yes
		11.3.2014 12:45	12:45:00	24	41,6	24,0	41,5	Yes
		11.3.2014 13:00	13:00:00	24,2	42,2	24,0	41,4	Yes
		11.3.2014 13:15	13:15:00	24,2	40,4	24,1	41,4	Yes
		11.3.2014 13:30	13:30:00	24,2	41,3	24,2	41,4	Yes
		11.3.2014 13:45	13:45:00	24,3	41,9	24,2	41,5	Yes
		11.3.2014 14:00	14:00:00	24,3	41,4	24,3	41,3	Yes
		11.3.2014 14:15	14:15:00	24,4	41,3	24,3	41,5	Yes
		11.3.2014 14:30	14:30:00	24,4	39,1	24,4	40,9	Yes
		11.3.2014 14:45	14:45:00	24,4	41,3	24,4	40,8	Yes
		11.3.2014 15:00	15:00:00	24,4	39,3	24,4	40,3	Yes
		11.3.2014 15:15	15:15:00	24,3	39,9	24,4	39,9	Yes
		11.3.2014 15:30	15:30:00	24,4	37,8	24,4	39,6	Yes
		11.3.2014 15:45	15:45:00	24,4	40,5	24,4	39,4	Yes
		11.3.2014 16:00	16:00:00	24,4	38,6	24,4	39,2	Yes
		11.3.2014 16:15	16:15:00	24,4	39,7	24,4	39,2	Yes
		11.3.2014 16:30	16:30:00	24,4	37,4	24,4	39,1	Yes
		11.3.2014 16:45	16:45:00	24,1	39,9	24,3	38,9	No
		11.3.2014 17:00	17:00:00	23,8	38,8	24,2	39,0	No
		11.3.2014 17:15	17:15:00	23,3	38,7	23,9	38,7	No
		11.3.2014 17:30	17:30:00	23,5	46,8	23,7	41,1	No
		11.3.2014 17:45	17:45:00	23,4	42,4	23,5	41,7	No
		11.3.2014 18:00	18:00:00	23,2	48	23,4	44,0	No
		11.3.2014 18:15	18:15:00	23,4	50,1	23,4	46,8	No
		11.3.2014 18:30	18:30:00	23,5	47,9	23,4	47,1	No
		11.3.2014 18:45	18:45:00	23,1	51	23,3	49,3	No
		11.3.2014 19:00	19:00:00	23,2	52,9	23,3	50,5	No
		11.3.2014 19:15	19:15:00	23,4	55,1	23,3	51,7	No
		11.3.2014 19:30	19:30:00	23,5	54	23,3	53,3	No
		11.3.2014 19:45	19:45:00	23,5	54,1	23,4	54,0	No
		11.3.2014 20:00	20:00:00	23,5	54,3	23,5	54,4	No
		11.3.2014 20:15	20:15:00	23,5	52,7	23,5	53,8	No
		11.3.2014 20:30	20:30:00	23,1	53,9	23,4	53,8	No
		11.3.2014 20:45	20:45:00	23,1	55,9	23,3	54,2	No
		11.3.2014 21:00	21:00:00	23,2	57,8	23,2	55,1	No
		11.3.2014 21:15	21:15:00	23,2	54,9	23,2	55,6	No
		11.3.2014 21:30	21:30:00	23,2	56,4	23,2	56,3	No
		11.3.2014 21:45	21:45:00	23,2	57,2	23,2	56,6	No
		11.3.2014 22:00	22:00:00	23,3	58,2	23,2	56,7	Yes
		11.3.2014 22:15	22:15:00	23,2	54,9	23,2	56,7	Yes
		11.3.2014 22:30	22:30:00	23,2	53,9	23,2	56,1	Yes
		11.3.2014 22:45	22:45:00	23,1	55,6	23,2	55,7	Yes
		11.3.2014 23:00	23:00:00	23,1	56,7	23,2	55,3	Yes
		11.3.2014 23:15	23:15:00	23,1	53,6	23,1	55,0	Yes
		11.3.2014 23:30	23:30:00	23,1	55,7	23,1	55,4	Yes
		11.3.2014 23:45	23:45:00	23,1	57,1	23,1	55,8	Yes
		12.3.2014 0:00	0:00:00	23,1	57,5	23,1	56,0	Yes
		12.3.2014 0:15	0:15:00	23,1	58,6	23,1	57,2	Yes
		12.3.2014 0:30	0:30:00	23	59,2	23,1	58,1	Yes
		12.3.2014 0:45	0:45:00	23	59,4	23,1	58,7	Yes
		12.3.2014 1:00	1:00:00	23	59,2	23,0	59,1	Yes
		12.3.2014 1:15	1:15:00	23	59,3	23,0	59,3	Yes
		12.3.2014 1:30	1:30:00	23	59,3	23,0	59,3	Yes
		12.3.2014 1:45	1:45:00	23	59,4	23,0	59,3	Yes
		12.3.2014 2:00	2:00:00	23	59,5	23,0	59,4	Yes
		12.3.2014 2:15	2:15:00	22,9	59,8	23,0	59,5	Yes
		12.3.2014 2:30	2:30:00	22,9	59,9	23,0	59,7	Yes
		12.3.2014 2:45	2:45:00	22,9	59,9	22,9	59,8	Yes
		12.3.2014 3:00	3:00:00	22,9	59,8	22,9	59,9	Yes
		12.3.2014 3:15	3:15:00	22,5	51	22,8	57,7	Yes
		12.3.2014 3:30	3:30:00	22,2	57,3	22,6	57,0	Yes
		12.3.2014 3:45	3:45:00	22,3	60,2	22,5	57,1	Yes
		12.3.2014 4:00	4:00:00	22,2	50,5	22,3	54,8	Yes
		12.3.2014 4:15	4:15:00	22	58,8	22,2	56,7	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0114 PHI_14_I0123	PHI_14_I0098 PHI_14_I0099 PHI_14_I0110 PHI_14_I0119 PHI_14_I0120 PHI_14_I0128 PHI_14_I0129	12.3.2014 4:30	4:30:00	22,2	60,4	22,2	57,5	Yes
		12.3.2014 4:45	4:45:00	22,1	52,6	22,1	55,6	Yes
		12.3.2014 5:00	5:00:00	21,9	58,3	22,1	57,5	Yes
		12.3.2014 5:15	5:15:00	22	61,3	22,1	58,2	Yes
		12.3.2014 5:30	5:30:00	22,1	61,7	22,0	58,5	Yes
		12.3.2014 5:45	5:45:00	22,2	53,4	22,1	58,7	Yes
		12.3.2014 6:00	6:00:00	22,3	52,5	22,2	57,2	Yes
		12.3.2014 6:15	6:15:00	22,3	47,3	22,2	53,7	Yes
		12.3.2014 6:30	6:30:00	22,5	46,9	22,3	50,0	Yes
		12.3.2014 6:45	6:45:00	22,6	47	22,4	48,4	Yes
N/A	N/A	12.3.2014 7:00	7:00:00	22,7	46,2	22,5	46,9	Yes
		12.3.2014 7:15	7:15:00	22,8	42,7	22,7	45,7	Yes
		12.3.2014 7:30	7:30:00	22,9	42,1	22,8	44,5	Yes
		12.3.2014 7:45	7:45:00	22,9	40,8	22,8	43,0	Yes
		12.3.2014 8:00	8:00:00	23,1	40,1	22,9	41,4	Yes
		12.3.2014 8:15	8:15:00	23,3	41,4	23,1	41,1	Yes
		12.3.2014 8:30	8:30:00	23,3	39	23,2	40,3	Yes
		12.3.2014 8:45	8:45:00	23,3	37	23,3	39,4	Yes
		12.3.2014 9:00	9:00:00	23,4	37,7	23,3	38,8	Yes
		12.3.2014 9:15	9:15:00	23,4	38,5	23,4	38,1	Yes
N/A	N/A	12.3.2014 9:30	9:30:00	23,5	40,9	23,4	38,5	Yes
		12.3.2014 9:45	9:45:00	23,5	41,1	23,5	39,6	Yes
		12.3.2014 10:00	10:00:00	23,7	41,5	23,5	40,5	Yes
		12.3.2014 10:15	10:15:00	23,8	41	23,6	41,1	Yes
		12.3.2014 10:30	10:30:00	23,8	41,3	23,7	41,2	Yes
		12.3.2014 10:45	10:45:00	24	42,9	23,8	41,7	Yes
		12.3.2014 11:00	11:00:00	24	41	23,9	41,6	Yes
		12.3.2014 11:15	11:15:00	24,2	43,1	24,0	42,1	Yes
		12.3.2014 11:30	11:30:00	24,1	38,3	24,1	41,3	Yes
		12.3.2014 11:45	11:45:00	24,1	41,5	24,1	41,0	Yes
N/A	N/A	12.3.2014 12:00	12:00:00	24,2	40,2	24,2	40,8	Yes
		12.3.2014 12:15	12:15:00	24,2	41,8	24,2	40,5	Yes
		12.3.2014 12:30	12:30:00	24,2	38,5	24,2	40,5	Yes
		12.3.2014 12:45	12:45:00	24,3	40,5	24,2	40,3	Yes
		12.3.2014 13:00	13:00:00	24,3	37,7	24,3	39,6	Yes
		12.3.2014 13:15	13:15:00	24,4	37,3	24,3	38,5	Yes
		12.3.2014 13:30	13:30:00	24,4	37,3	24,4	38,2	Yes
		12.3.2014 13:45	13:45:00	24,4	39	24,4	37,8	Yes
		12.3.2014 14:00	14:00:00	24,5	38,9	24,4	38,1	Yes
		12.3.2014 14:15	14:15:00	24,6	39	24,5	38,6	Yes
N/A	N/A	12.3.2014 14:30	14:30:00	24,7	36,9	24,6	38,5	Yes
		12.3.2014 14:45	14:45:00	24,4	35,2	24,6	37,5	No
		12.3.2014 15:00	15:00:00	23,8	34	24,4	36,3	No
		12.3.2014 15:15	15:15:00	23,5	36,4	24,1	35,6	No
		12.3.2014 15:30	15:30:00	23,2	35,3	23,7	35,2	No
		12.3.2014 15:45	15:45:00	23	38,4	23,4	36,0	No
		12.3.2014 16:00	16:00:00	23,1	39,8	23,2	37,5	No
		12.3.2014 16:15	16:15:00	22,6	37,5	23,0	37,8	No
		12.3.2014 16:30	16:30:00	22,9	44	22,9	39,9	No
		12.3.2014 16:45	16:45:00	23,1	44,6	22,9	41,5	No
N/A	N/A	12.3.2014 17:00	17:00:00	22,7	41,9	22,8	42,0	No
		12.3.2014 17:15	17:15:00	22,9	48	22,9	44,6	No
		12.3.2014 17:30	17:30:00	23,2	47,8	23,0	45,6	No
		12.3.2014 17:45	17:45:00	22,7	46,5	22,9	46,1	No
		12.3.2014 18:00	18:00:00	22,9	50,3	22,9	48,2	No
		12.3.2014 18:15	18:15:00	23,1	49,2	23,0	48,5	No
		12.3.2014 18:30	18:30:00	22,9	43,8	22,9	47,5	No
		12.3.2014 18:45	18:45:00	22,8	50,6	22,9	48,5	No
		12.3.2014 19:00	19:00:00	23	51	23,0	48,7	No
		12.3.2014 19:15	19:15:00	23,1	50,6	23,0	49,0	No
N/A	N/A	12.3.2014 19:30	19:30:00	23,2	51,6	23,0	51,0	No
		12.3.2014 19:45	19:45:00	22,8	49,7	23,0	50,7	No
		12.3.2014 20:00	20:00:00	22,8	52,1	23,0	51,0	No
		12.3.2014 20:15	20:15:00	23	51,5	23,0	51,2	No
		12.3.2014 20:30	20:30:00	23	51,6	22,9	51,2	No
		12.3.2014 20:45	20:45:00	23,1	52,4	23,0	51,9	No
		12.3.2014 21:00	21:00:00	23,1	52,3	23,1	52,0	No
		12.3.2014 21:15	21:15:00	23,1	51,4	23,1	51,9	No
		12.3.2014 21:30	21:30:00	23,1	52,2	23,1	52,1	No
		12.3.2014 21:45	21:45:00	23,1	49,2	23,1	51,3	No
N/A	N/A	12.3.2014 22:00	22:00:00	22,6	52,3	23,0	51,3	Yes
		12.3.2014 22:15	22:15:00	22,6	51,4	22,9	51,3	Yes
		12.3.2014 22:30	22:30:00	22,7	50,2	22,8	50,8	Yes
		12.3.2014 22:45	22:45:00	22,7	51,5	22,7	51,4	Yes
		12.3.2014 23:00	23:00:00	22,8	53,1	22,7	51,6	Yes
		12.3.2014 23:15	23:15:00	22,7	50,4	22,7	51,3	Yes
		12.3.2014 23:30	23:30:00	22,8	51,7	22,8	51,7	Yes
		12.3.2014 23:45	23:45:00	22,8	53,4	22,8	52,2	Yes
		13.3.2014 0:00	0:00:00	22,8	53,8	22,8	52,3	Yes
		13.3.2014 0:15	0:15:00	22,8	53,2	22,8	53,0	Yes
N/A	N/A	13.3.2014 0:30	0:30:00	22,8	53	22,8	53,4	Yes
		13.3.2014 0:45	0:45:00	22,9	53	22,8	53,3	Yes
		13.3.2014 1:00	1:00:00	22,9	52,2	22,9	52,9	Yes
		13.3.2014 1:15	1:15:00	22,9	52,7	22,9	52,7	Yes
		13.3.2014 1:30	1:30:00	22,9	53,4	22,9	52,8	Yes
		13.3.2014 1:45	1:45:00	22,9	54,4	22,9	53,2	Yes
		13.3.2014 2:00	2:00:00	22,9	55	22,9	53,9	Yes
		13.3.2014 2:15	2:15:00	22,9	54,6	22,9	54,4	Yes
		13.3.2014 2:30	2:30:00	22,9	55,6	22,9	54,9	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_I0114 PHI_14_I0123	PHI_14_I0099 PHI_14_I0110 PHI_14_I0120 PHI_14_I0129	13.3.2014 2:45	2:45:00	22,6	50,7	22,8	54,0	Yes
		13.3.2014 3:00	3:00:00	22,5	55,8	22,7	54,2	Yes
		13.3.2014 3:15	3:15:00	22,6	54,6	22,7	54,2	Yes
		13.3.2014 3:30	3:30:00	22,7	55,9	22,6	54,3	Yes
		13.3.2014 3:45	3:45:00	22,3	51,5	22,5	54,5	Yes
		13.3.2014 4:00	4:00:00	22,3	55,6	22,5	54,4	Yes
		13.3.2014 4:15	4:15:00	22,5	55,3	22,5	54,6	Yes
		13.3.2014 4:30	4:30:00	22,6	56,6	22,4	54,8	Yes
		13.3.2014 4:45	4:45:00	22,3	50,8	22,4	54,6	Yes
		13.3.2014 5:00	5:00:00	22,3	56,2	22,4	54,7	Yes
N/A	N/A	13.3.2014 5:15	5:15:00	22,4	54,8	22,4	54,6	Yes
		13.3.2014 5:30	5:30:00	22,5	55,4	22,4	54,3	Yes
		13.3.2014 5:45	5:45:00	22,5	51,5	22,4	54,5	Yes
		13.3.2014 6:00	6:00:00	22,5	49,8	22,5	52,9	Yes
		13.3.2014 6:15	6:15:00	22,4	47,3	22,5	51,0	Yes
		13.3.2014 6:30	6:30:00	22,5	45,7	22,5	48,6	Yes
		13.3.2014 6:45	6:45:00	22,5	44,6	22,5	46,9	Yes
		13.3.2014 7:00	7:00:00	22,7	43,9	22,5	45,4	Yes
		13.3.2014 7:15	7:15:00	22,8	44,2	22,6	44,6	Yes
		13.3.2014 7:30	7:30:00	22,9	42,3	22,7	43,8	Yes
N/A	N/A	13.3.2014 7:45	7:45:00	22,9	41,6	22,8	43,0	Yes
		13.3.2014 8:00	8:00:00	22,9	39,6	22,9	41,9	Yes
		13.3.2014 8:15	8:15:00	23	40,4	22,9	41,0	Yes
		13.3.2014 8:30	8:30:00	23,1	42,2	23,0	41,0	Yes
		13.3.2014 8:45	8:45:00	23,1	42,4	23,0	41,2	Yes
		13.3.2014 9:00	9:00:00	23,3	41,2	23,1	41,6	Yes
		13.3.2014 9:15	9:15:00	23,4	43,2	23,2	42,3	Yes
		13.3.2014 9:30	9:30:00	23,5	42,4	23,3	42,3	Yes
		13.3.2014 9:45	9:45:00	23,6	42,4	23,5	42,3	Yes
		13.3.2014 10:00	10:00:00	23,6	41,3	23,5	42,3	Yes
N/A	N/A	13.3.2014 10:15	10:15:00	23,7	41,4	23,6	41,9	Yes
		13.3.2014 10:30	10:30:00	23,5	38,4	23,6	40,9	Yes
		13.3.2014 10:45	10:45:00	23,6	40	23,6	40,3	Yes
		13.3.2014 11:00	11:00:00	23,6	39,9	23,6	39,9	Yes
		13.3.2014 11:15	11:15:00	23,8	41,6	23,6	40,0	Yes
		13.3.2014 11:30	11:30:00	23,8	39,6	23,7	40,3	Yes
		13.3.2014 11:45	11:45:00	23,8	39,9	23,8	40,3	Yes
		13.3.2014 12:00	12:00:00	23,9	40,5	23,8	40,4	Yes
		13.3.2014 12:15	12:15:00	24,1	41,1	23,9	40,3	Yes
		13.3.2014 12:30	12:30:00	24,2	40,6	24,0	40,5	Yes
N/A	N/A	13.3.2014 12:45	12:45:00	24,4	42,7	24,2	41,2	Yes
		13.3.2014 13:00	13:00:00	24,5	41,1	24,3	41,4	Yes
		13.3.2014 13:15	13:15:00	24,5	40,6	24,4	41,3	Yes
		13.3.2014 13:30	13:30:00	24,7	40,8	24,5	41,3	Yes
		13.3.2014 13:45	13:45:00	24,7	39,9	24,6	40,6	Yes
		13.3.2014 14:00	14:00:00	24,8	40,8	24,7	40,5	Yes
		13.3.2014 14:15	14:15:00	24,9	40,4	24,8	40,5	Yes
		13.3.2014 14:30	14:30:00	25	38,1	24,9	39,8	Yes
		13.3.2014 14:45	14:45:00	24,8	36,9	24,9	39,1	No
		13.3.2014 15:00	15:00:00	24,4	35	24,8	37,6	No
N/A	N/A	13.3.2014 15:15	15:15:00	24,1	33,8	24,6	36,0	No
		13.3.2014 15:30	15:30:00	23,7	34,4	24,3	35,0	No
		13.3.2014 15:45	15:45:00	23,3	33,4	23,9	34,2	No
		13.3.2014 16:00	16:00:00	23,4	38,5	23,6	35,0	No
		13.3.2014 16:15	16:15:00	22,9	35,4	23,3	35,4	No
		13.3.2014 16:30	16:30:00	22,9	40,4	23,1	36,9	No
		13.3.2014 16:45	16:45:00	23,2	41,7	23,1	39,0	No
		13.3.2014 17:00	17:00:00	22,6	38,2	22,9	38,9	No
		13.3.2014 17:15	17:15:00	22,9	43,9	22,9	41,1	No
		13.3.2014 17:30	17:30:00	23,1	45,4	23,0	42,3	No
N/A	N/A	13.3.2014 17:45	17:45:00	22,8	40,7	22,9	42,1	No
		13.3.2014 18:00	18:00:00	22,9	48,1	22,9	44,5	No
		13.3.2014 18:15	18:15:00	23,1	48,3	23,0	45,6	No
		13.3.2014 18:30	18:30:00	23,2	45,6	23,0	45,7	No
		13.3.2014 18:45	18:45:00	22,7	44,7	23,0	46,7	No
		13.3.2014 19:00	19:00:00	22,8	51,1	23,0	47,4	No
		13.3.2014 19:15	19:15:00	23	51,5	22,9	48,2	No
		13.3.2014 19:30	19:30:00	23,1	51	22,9	49,6	No
		13.3.2014 19:45	19:45:00	23,2	51,9	23,0	51,4	No
		13.3.2014 20:00	20:00:00	23,3	49,7	23,2	51,0	No
N/A	N/A	13.3.2014 20:15	20:15:00	22,7	48,3	23,1	50,2	No
		13.3.2014 20:30	20:30:00	22,8	53,6	23,0	50,9	No
		13.3.2014 20:45	20:45:00	23	54,1	23,0	51,4	No
		13.3.2014 21:00	21:00:00	23	54,7	22,9	52,7	No
		13.3.2014 21:15	21:15:00	23,1	53,7	23,0	54,0	No
		13.3.2014 21:30	21:30:00	23,2	54,3	23,1	54,2	No
		13.3.2014 21:45	21:45:00	23,2	54,9	23,1	54,4	No
		13.3.2014 22:00	22:00:00	23,2	55,1	23,2	54,5	Yes
		13.3.2014 22:15	22:15:00	23,2	53,8	23,2	54,5	Yes
		13.3.2014 22:30	22:30:00	23,2	54,3	23,2	54,5	Yes
N/A	N/A	13.3.2014 22:45	22:45:00	23,2	54,3	23,2	54,4	Yes
		13.3.2014 23:00	23:00:00	23,2	54,4	23,2	54,2	Yes
		13.3.2014 23:15	23:15:00	23,3	53,3	23,2	54,1	Yes
		13.3.2014 23:30	23:30:00	23,2	54,7	23,2	54,2	Yes
		13.3.2014 23:45	23:45:00	23,2	55,3	23,2	54,4	Yes
		14.3.2014 0:00	0:00:00	23,2	55,2	23,2	54,6	Yes
		14.3.2014 0:15	0:15:00	23,2	55,5	23,2	55,2	Yes
		14.3.2014 0:30	0:30:00	23,2	55,7	23,2	55,4	Yes
		14.3.2014 0:45	0:45:00	23,2	55,6	23,2	55,5	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

PHI_14_0123		PHI_14_0099 PHI_14_0100 PHI_14_0111 PHI_14_0120 PHI_14_0129	14.3.2014 1:00	1:00:00	23.1	55.6	23.2	55.6	Yes
			14.3.2014 1:15	1:15:00	22.6	51.4	23.0	54.6	Yes
			14.3.2014 1:30	1:30:00	22.6	54.8	22.9	54.4	Yes
			14.3.2014 1:45	1:45:00	22.7	57	22.8	54.7	Yes
			14.3.2014 2:00	2:00:00	22.8	56.3	22.7	54.9	Yes
			14.3.2014 2:15	2:15:00	22.8	55.9	22.7	56.0	Yes
			14.3.2014 2:30	2:30:00	22.4	52.8	22.7	55.5	Yes
			14.3.2014 2:45	2:45:00	22.5	55.8	22.6	55.2	Yes
			14.3.2014 3:00	3:00:00	22.5	57	22.6	55.4	Yes
			14.3.2014 3:15	3:15:00	22.6	57.9	22.5	55.9	Yes
		N/A	14.3.2014 3:30	3:30:00	22.3	51.2	22.5	55.5	Yes
			14.3.2014 3:45	3:45:00	22.3	55.6	22.4	55.4	Yes
			14.3.2014 4:00	4:00:00	22.5	54.7	22.4	54.9	Yes
			14.3.2014 4:15	4:15:00	22.2	54.2	22.3	53.9	Yes
			14.3.2014 4:30	4:30:00	22.4	55.4	22.4	55.0	Yes
			14.3.2014 4:45	4:45:00	22.2	50.4	22.3	53.7	Yes
			14.3.2014 5:00	5:00:00	22.2	55.3	22.3	53.8	Yes
			14.3.2014 5:15	5:15:00	22.4	54.8	22.3	54.0	Yes
			14.3.2014 5:30	5:30:00	22.1	53.1	22.2	53.4	Yes
			14.3.2014 5:45	5:45:00	22.4	50.1	22.3	53.3	Yes
			14.3.2014 6:00	6:00:00	22.4	50.6	22.3	52.2	Yes
			14.3.2014 6:15	6:15:00	22.6	48.8	22.4	50.7	Yes
			14.3.2014 6:30	6:30:00	22.7	48.4	22.5	49.5	Yes
			14.3.2014 6:45	6:45:00	22.8	47.8	22.6	48.9	Yes
			14.3.2014 7:00	7:00:00	22.9	47.1	22.8	48.0	Yes
			14.3.2014 7:15	7:15:00	22.9	47.1	22.8	47.6	Yes
			14.3.2014 7:30	7:30:00	22.9	46	22.9	47.0	Yes
			14.3.2014 7:45	7:45:00	22.9	44.5	22.9	46.2	Yes
			14.3.2014 8:00	8:00:00	23	43.5	22.9	45.3	Yes
			14.3.2014 8:15	8:15:00	23	44.7	23.0	44.7	Yes
			14.3.2014 8:30	8:30:00	22.9	43.3	23.0	44.0	Yes
			14.3.2014 8:45	8:45:00	22.9	43.4	23.0	43.7	Yes
			14.3.2014 9:00	9:00:00	23	44.9	23.0	44.1	Yes
			14.3.2014 9:15	9:15:00	23.1	42.4	23.0	43.5	Yes
			14.3.2014 9:30	9:30:00	23.1	44	23.0	43.7	Yes
			14.3.2014 9:45	9:45:00	23.1	41.9	23.1	43.3	Yes
			14.3.2014 10:00	10:00:00	23.2	42.9	23.1	42.8	Yes
			14.3.2014 10:15	10:15:00	23.2	42.2	23.2	42.8	Yes
			14.3.2014 10:30	10:30:00	23.1	41.2	23.2	42.1	Yes
			14.3.2014 10:45	10:45:00	23	41.7	23.1	42.0	Yes
			14.3.2014 11:00	11:00:00	23.1	41.9	23.1	41.8	Yes
			14.3.2014 11:15	11:15:00	22.9	39.8	23.0	41.2	Yes
			14.3.2014 11:30	11:30:00	23.1	44.8	23.0	42.1	Yes
			14.3.2014 11:45	11:45:00	23	40.6	23.0	41.8	Yes
			14.3.2014 12:00	12:00:00	23.1	40.7	23.0	41.5	Yes
			14.3.2014 12:15	12:15:00	23.3	42	23.1	42.0	Yes
			14.3.2014 12:30	12:30:00	23.3	41.5	23.2	41.2	Yes
			14.3.2014 12:45	12:45:00	23.4	41.2	23.3	41.4	Yes
			14.3.2014 13:00	13:00:00	23.4	40.5	23.4	41.3	Yes
			14.3.2014 13:15	13:15:00	23.5	40.7	23.4	41.0	Yes
			14.3.2014 13:30	13:30:00	23.7	40.8	23.5	40.8	Yes
			14.3.2014 13:45	13:45:00	23.6	39.9	23.6	40.5	Yes
			14.3.2014 14:00	14:00:00	23.9	41.7	23.7	40.8	Yes
			14.3.2014 14:15	14:15:00	24	40.4	23.8	40.7	Yes
			14.3.2014 14:30	14:30:00	24	39.2	23.9	40.3	Yes
			14.3.2014 14:45	14:45:00	23.8	40.7	23.9	40.5	No
			14.3.2014 15:00	15:00:00	23.2	37.3	23.8	39.4	No
			14.3.2014 15:15	15:15:00	22.8	36.2	23.5	38.4	No
			14.3.2014 15:30	15:30:00	22.9	43.5	23.2	39.4	No
			14.3.2014 15:45	15:45:00	23.3	44.5	23.1	40.4	No
			14.3.2014 16:00	16:00:00	22.8	40.4	23.0	41.2	No
			14.3.2014 16:15	16:15:00	23	47.2	23.0	43.9	No
			14.3.2014 16:30	16:30:00	23.2	48.5	23.1	45.2	No
			14.3.2014 16:45	16:45:00	22.8	42.9	23.0	44.8	No
			14.3.2014 17:00	17:00:00	23	49.1	23.0	46.9	No
			14.3.2014 17:15	17:15:00	23.2	50.3	23.1	47.7	No
			14.3.2014 17:30	17:30:00	23.2	46.8	23.1	47.3	No
			14.3.2014 17:45	17:45:00	22.8	48.8	23.1	48.8	No
			14.3.2014 18:00	18:00:00	23	51.9	23.1	49.5	No
			14.3.2014 18:15	18:15:00	23.2	51.8	23.1	49.8	No
			14.3.2014 18:30	18:30:00	23.3	51.9	23.1	51.1	No
			14.3.2014 18:45	18:45:00	22.8	45.5	23.1	50.3	No
			14.3.2014 19:00	19:00:00	22.8	53.4	23.0	50.7	No
			14.3.2014 19:15	19:15:00	22.9	50.7	23.0	50.4	No
			14.3.2014 19:30	19:30:00	22.7	50.4	22.8	50.0	No
			14.3.2014 19:45	19:45:00	22.8	52.5	22.8	51.8	No
			14.3.2014 20:00	20:00:00	22.9	50.5	22.8	51.0	No
			14.3.2014 20:15	20:15:00	22.9	52.8	22.8	51.6	No
			14.3.2014 20:30	20:30:00	23	53.9	22.9	52.4	No
			14.3.2014 20:45	20:45:00	23.1	54.7	23.0	53.0	No
			14.3.2014 21:00	21:00:00	23.1	54.9	23.0	54.1	No
			14.3.2014 21:15	21:15:00	23.1	51.8	23.1	53.8	No
			14.3.2014 21:30	21:30:00	23.1	53.7	23.1	53.8	No
			14.3.2014 21:45	21:45:00	23.1	54	23.1	53.6	No
			14.3.2014 22:00	22:00:00	23.1	54.2	23.1	53.4	Yes
			14.3.2014 22:15	22:15:00	23.1	50.3	23.1	53.1	Yes
			14.3.2014 22:30	22:30:00	23.1	50.9	23.1	52.4	Yes
			14.3.2014 22:45	22:45:00	23	52.1	23.1	51.9	Yes
			14.3.2014 23:00	23:00:00	23.1	52.9	23.1	51.6	Yes

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

			14.3.2014 23:15	23:15:00	23	49,9	23,1	51,5	Yes
			14.3.2014 23:30	23:30:00	23	51,7	23,0	51,7	Yes
			14.3.2014 23:45	23:45:00	23	52,6	23,0	51,8	Yes
			15.3.2014 0:00	0:00:00	23	52,7	23,0	51,7	Yes
			15.3.2014 0:15	0:15:00	23	53,3	23,0	52,6	Yes
			15.3.2014 0:30	0:30:00	23	53,4	23,0	53,0	Yes
			15.3.2014 0:45	0:45:00	22,9	53,9	23,0	53,3	Yes
			15.3.2014 1:00	1:00:00	22,9	54,1	23,0	53,7	Yes
			15.3.2014 1:15	1:15:00	22,9	53,8	22,9	53,8	Yes
			15.3.2014 1:30	1:30:00	22,9	54,2	22,9	54,0	Yes
			15.3.2014 1:45	1:45:00	22,9	54,5	22,9	54,2	Yes
			15.3.2014 2:00	2:00:00	22,9	54,4	22,9	54,2	Yes
			15.3.2014 2:15	2:15:00	22,8	54,9	22,9	54,5	Yes
			15.3.2014 2:30	2:30:00	22,8	54,3	22,9	54,5	Yes
			15.3.2014 2:45	2:45:00	22,8	54,2	22,8	54,5	Yes
			15.3.2014 3:00	3:00:00	22,7	55,3	22,8	54,7	Yes
			15.3.2014 3:15	3:15:00	22,8	55,4	22,8	54,8	Yes
			15.3.2014 3:30	3:30:00	22,3	52,1	22,7	54,3	Yes
			15.3.2014 3:45	3:45:00	22,4	57,4	22,6	55,1	Yes
			15.3.2014 4:00	4:00:00	22,5	57	22,5	55,5	Yes
			15.3.2014 4:15	4:15:00	22,6	57	22,5	55,9	Yes
			15.3.2014 4:30	4:30:00	22,2	53,1	22,4	56,1	Yes
			15.3.2014 4:45	4:45:00	22,3	57,2	22,4	56,1	Yes
			15.3.2014 5:00	5:00:00	22,4	57	22,4	56,1	Yes
			15.3.2014 5:15	5:15:00	22,3	52,2	22,3	54,9	Yes
			15.3.2014 5:30	5:30:00	22,2	56,4	22,3	55,7	Yes
			15.3.2014 5:45	5:45:00	22,3	55,5	22,3	55,3	Yes
			15.3.2014 6:00	6:00:00	22,4	49,9	22,3	53,5	Yes
			15.3.2014 6:15	6:15:00	22,3	53,2	22,3	53,8	Yes
			15.3.2014 6:30	6:30:00	22,3	47,2	22,3	51,5	Yes
			15.3.2014 6:45	6:45:00	22,5	49,1	22,4	49,9	Yes
			15.3.2014 7:00	7:00:00	22,5	47,7	22,4	49,3	Yes
			15.3.2014 7:15	7:15:00	22,5	49	22,5	48,3	Yes
			15.3.2014 7:30	7:30:00	22,6	50,3	22,5	49,0	Yes
			15.3.2014 7:45	7:45:00	22,6	46,6	22,6	48,4	Yes
			15.3.2014 8:00	8:00:00	22,7	47,8	22,6	48,4	Yes
			15.3.2014 8:15	8:15:00	22,7	45,4	22,7	47,5	Yes
			15.3.2014 8:30	8:30:00	22,7	43,6	22,7	45,9	Yes
			15.3.2014 8:45	8:45:00	22,7	46,9	22,7	45,9	Yes
			15.3.2014 9:00	9:00:00	22,7	44,2	22,7	45,0	Yes
			15.3.2014 9:15	9:15:00	22,7	43,9	22,7	44,7	Yes
			15.3.2014 9:30	9:30:00	22,7	44,5	22,7	44,9	Yes
			15.3.2014 9:45	9:45:00	22,6	41,8	22,7	43,6	Yes
			15.3.2014 10:00	10:00:00	22,7	46,4	22,7	44,2	Yes
			15.3.2014 10:15	10:15:00	22,6	42	22,7	43,7	Yes
			15.3.2014 10:30	10:30:00	22,8	47,4	22,7	44,4	Yes
			15.3.2014 10:45	10:45:00	22,6	41,7	22,7	44,4	Yes
			15.3.2014 11:00	11:00:00	22,8	48,3	22,7	44,9	Yes
			15.3.2014 11:15	11:15:00	22,7	43	22,7	45,1	Yes
			15.3.2014 11:30	11:30:00	22,6	42,7	22,7	43,9	Yes
			15.3.2014 11:45	11:45:00	22,7	49,5	22,7	45,9	Yes
			15.3.2014 12:00	12:00:00	22,5	48	22,6	45,8	Yes
			15.3.2014 12:15	12:15:00	22,6	53,5	22,6	48,4	Yes
			15.3.2014 12:30	12:30:00	22,8	55,5	22,7	51,6	Yes
			15.3.2014 12:45	12:45:00	22,5	52,2	22,6	52,3	Yes
			15.3.2014 13:00	13:00:00	22,6	56,7	22,6	54,5	Yes
			15.3.2014 13:15	13:15:00	22,7	54,1	22,7	54,6	Yes
			15.3.2014 13:30	13:30:00	22,4	48,5	22,6	52,9	Yes
			15.3.2014 13:45	13:45:00	22,7	51,8	22,6	52,8	Yes
			15.3.2014 14:00	14:00:00	22,5	45,6	22,6	50,0	Yes
			15.3.2014 14:15	14:15:00	22,5	50,7	22,5	49,2	No
			15.3.2014 14:30	14:30:00	22,6	53,8	22,6	50,5	No
			15.3.2014 14:45	14:45:00	22,3	51,6	22,5	50,4	No
			15.3.2014 15:00	15:00:00	22,4	54,7	22,5	52,7	No
			15.3.2014 15:15	15:15:00	22,5	54,4	22,5	53,6	No
			15.3.2014 15:30	15:30:00	22,6	53,6	22,5	53,6	No
			15.3.2014 15:45	15:45:00	22,6	54,8	22,5	54,4	No
			15.3.2014 16:00	16:00:00	22,7	53,6	22,6	54,1	No
			15.3.2014 16:15	16:15:00	22,7	53,7	22,7	53,9	No
			15.3.2014 16:30	16:30:00	22,7	54	22,7	54,0	No
			15.3.2014 16:45	16:45:00	22,7	53,3	22,7	53,7	No
			15.3.2014 17:00	17:00:00	22,7	54,4	22,7	53,9	No
			15.3.2014 17:15	17:15:00	22,7	53,4	22,7	53,8	No
			15.3.2014 17:30	17:30:00	22,7	53,5	22,7	53,7	No
			15.3.2014 17:45	17:45:00	22,7	54,1	22,7	53,9	No
			15.3.2014 18:00	18:00:00	22,7	53,3	22,7	53,6	No
			15.3.2014 18:15	18:15:00	22,7	54,4	22,7	53,8	No
			15.3.2014 18:30	18:30:00	22,7	53,7	22,7	53,9	No
			15.3.2014 18:45	18:45:00	22,7	53,4	22,7	53,7	No
			15.3.2014 19:00	19:00:00	22,7	54,2	22,7	53,9	No
			15.3.2014 19:15	19:15:00	22,7	53,6	22,7	53,7	No
			15.3.2014 19:30	19:30:00	22,7	53,7	22,7	53,7	No
			15.3.2014 19:45	19:45:00	22,7	54,1	22,7	53,9	No
			15.3.2014 20:00	20:00:00	22,6	52,1	22,7	53,4	No
			15.3.2014 20:15	20:15:00	22,5	53,4	22,6	53,3	No
			15.3.2014 20:30	20:30:00	22,5	51,8	22,6	52,9	No
			15.3.2014 20:45	20:45:00	22,5	53,7	22,5	52,8	No
			15.3.2014 21:00	21:00:00	22,5	54,2	22,5	53,3	No
			15.3.2014 21:15	21:15:00	22,5	53,5	22,5	53,3	No

PHI_14_10100
PHI_14_10111
PHI_14_10120
PHI_14_10129

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

N/A	N/A	N/A	15.3.2014 21:30	21:30:00	22,5	53,8	22,5	53,8	No
			15.3.2014 21:45	21:45:00	22,4	54	22,5	53,9	No
			15.3.2014 22:00	22:00:00	22,4	53,5	22,5	53,7	No
			15.3.2014 22:15	22:15:00	22,4	53,8	22,4	53,8	No
			15.3.2014 22:30	22:30:00	22,4	54,3	22,4	53,9	No
			15.3.2014 22:45	22:45:00	22,4	53,3	22,4	53,7	No
			15.3.2014 23:00	23:00:00	22,3	54,3	22,4	53,9	No
			15.3.2014 23:15	23:15:00	22,3	52	22,4	53,5	No
			15.3.2014 23:30	23:30:00	22,2	53,6	22,3	53,3	No
			15.3.2014 23:45	23:45:00	22,2	53,8	22,3	53,4	No
			16.3.2014 0:00	0:00:00	22,1	54	22,2	53,4	No
			16.3.2014 0:15	0:15:00	22,1	54,7	22,2	54,0	No
			16.3.2014 0:30	0:30:00	22,1	54,9	22,1	54,4	No
			16.3.2014 0:45	0:45:00	22,1	54,5	22,1	54,5	No
			16.3.2014 1:00	1:00:00	22,1	54,4	22,1	54,6	No
			16.3.2014 1:15	1:15:00	22,1	54,6	22,1	54,6	No
			16.3.2014 1:30	1:30:00	22,1	54,7	22,1	54,6	No
			16.3.2014 1:45	1:45:00	22,1	55,2	22,1	54,7	No
			16.3.2014 2:00	2:00:00	22,1	53,8	22,1	54,6	No
			16.3.2014 2:15	2:15:00	22,1	55	22,1	54,7	No
			16.3.2014 2:30	2:30:00	22,1	55,3	22,1	54,8	No
			16.3.2014 2:45	2:45:00	22	54,2	22,1	54,6	No
			16.3.2014 3:00	3:00:00	22	55,7	22,1	55,1	No
			16.3.2014 3:15	3:15:00	22	55,4	22,0	55,2	No
			16.3.2014 3:30	3:30:00	22	55,2	22,0	55,1	No
			16.3.2014 3:45	3:45:00	22	55,7	22,0	55,5	No
			16.3.2014 4:00	4:00:00	22	54,9	22,0	55,3	No
			16.3.2014 4:15	4:15:00	22	55,1	22,0	55,2	No
			16.3.2014 4:30	4:30:00	22,1	56,2	22,0	55,5	No
			16.3.2014 4:45	4:45:00	22,2	55,3	22,1	55,4	No
			16.3.2014 5:00	5:00:00	22,2	56,5	22,1	55,8	No
			16.3.2014 5:15	5:15:00	22,3	55,3	22,2	55,8	No
			16.3.2014 5:30	5:30:00	22,2	56,3	22,2	55,9	No
			16.3.2014 5:45	5:45:00	22,3	55,2	22,3	55,8	No
			16.3.2014 6:00	6:00:00	22,3	56	22,3	55,7	No
			16.3.2014 6:15	6:15:00	22,3	55,5	22,3	55,8	No
			16.3.2014 6:30	6:30:00	22,3	54,1	22,3	55,2	No
			16.3.2014 6:45	6:45:00	22,1	52,1	22,3	54,4	No
			16.3.2014 7:00	7:00:00	22	56,3	22,2	54,5	No
			16.3.2014 7:15	7:15:00	22,1	55,6	22,1	54,5	No
			16.3.2014 7:30	7:30:00	22,1	56,6	22,1	55,2	No
			16.3.2014 7:45	7:45:00	22,1	55,3	22,1	56,0	No
			16.3.2014 8:00	8:00:00	22,1	56,5	22,1	56,0	No
			16.3.2014 8:15	8:15:00	22,2	55,5	22,1	56,0	No
			16.3.2014 8:30	8:30:00	22,1	56,4	22,1	55,9	No
			16.3.2014 8:45	8:45:00	22,1	54,7	22,1	55,8	No
			16.3.2014 9:00	9:00:00	22,1	55,8	22,1	55,6	No
			16.3.2014 9:15	9:15:00	22,1	55,4	22,1	55,6	No
			16.3.2014 9:30	9:30:00	22	55,5	22,1	55,4	No
			16.3.2014 9:45	9:45:00	22	55,7	22,1	55,6	No
			16.3.2014 10:00	10:00:00	22	56,1	22,0	55,7	No
			16.3.2014 10:15	10:15:00	22	55,9	22,0	55,8	No
			16.3.2014 10:30	10:30:00	22	56,1	22,0	56,0	No
			16.3.2014 10:45	10:45:00	21,9	54,6	22,0	55,7	No
			16.3.2014 11:00	11:00:00	21,9	55,6	22,0	55,6	No
			16.3.2014 11:15	11:15:00	21,9	55,5	21,9	55,5	No
			16.3.2014 11:30	11:30:00	21,9	55,8	21,9	55,4	No
			16.3.2014 11:45	11:45:00	21,9	55,6	21,9	55,6	No
			16.3.2014 12:00	12:00:00	21,8	55,9	21,9	55,7	No
			16.3.2014 12:15	12:15:00	21,8	55,9	21,9	55,8	No
			16.3.2014 12:30	12:30:00	21,8	56	21,8	55,9	No
			16.3.2014 12:45	12:45:00	21,8	55,6	21,8	55,9	No
			16.3.2014 13:00	13:00:00	21,8	55,6	21,8	55,8	No
			16.3.2014 13:15	13:15:00	21,7	55,9	21,8	55,8	No
			16.3.2014 13:30	13:30:00	21,7	56,1	21,8	55,8	No
			16.3.2014 13:45	13:45:00	21,7	55,9	21,7	55,9	No
			16.3.2014 14:00	14:00:00	21,7	55,9	21,7	56,0	No
			16.3.2014 14:15	14:15:00	21,7	55,1	21,7	55,8	No
			16.3.2014 14:30	14:30:00	21,7	55,1	21,7	55,5	No
			16.3.2014 14:45	14:45:00	21,7	56	21,7	55,5	No
			16.3.2014 15:00	15:00:00	21,6	55,9	21,7	55,5	No
			16.3.2014 15:15	15:15:00	21,6	56,5	21,7	55,9	No
			16.3.2014 15:30	15:30:00	21,6	56,2	21,6	56,2	No
			16.3.2014 15:45	15:45:00	21,6	55,9	21,6	56,1	No
			16.3.2014 16:00	16:00:00	21,6	56,2	21,6	56,2	No
			16.3.2014 16:15	16:15:00	21,6	56,5	21,6	56,2	No
			16.3.2014 16:30	16:30:00	21,6	56,4	21,6	56,3	No
			16.3.2014 16:45	16:45:00	21,6	55,7	21,6	56,2	No
			16.3.2014 17:00	17:00:00	21,6	57,1	21,6	56,4	No
			16.3.2014 17:15	17:15:00	21,6	57,8	21,6	56,8	No
			16.3.2014 17:30	17:30:00	21,6	58,2	21,6	57,2	No
			16.3.2014 17:45	17:45:00	21,6	58,3	21,6	57,9	No
			16.3.2014 18:00	18:00:00	21,5	58,7	21,6	58,3	No
			16.3.2014 18:15	18:15:00	21,5	59	21,6	58,6	No
			16.3.2014 18:30	18:30:00	21,5	59,3	21,5	58,8	No
			16.3.2014 18:45	18:45:00	21,5	59,6	21,5	59,2	No
			16.3.2014 19:00	19:00:00	21,5	59,8	21,5	59,4	No
			16.3.2014 19:15	19:15:00	21,5	60,3	21,5	59,8	No
			16.3.2014 19:30	19:30:00	21,5	59,6	21,5	59,8	No

TEST REPORT

Creation date of report: 27.3.2014

Signature: stavmi

TYPE OF TEST: ISO 19798

			16.3.2014 19:45	19:45:00	21,5	60,6	21,5	60,1	No
			16.3.2014 20:00	20:00:00	21,4	59,1	21,5	59,9	No
			16.3.2014 20:15	20:15:00	21,5	60,2	21,5	59,9	No
			16.3.2014 20:30	20:30:00	21,5	60,9	21,5	60,2	No
			16.3.2014 20:45	20:45:00	21,6	61,5	21,5	60,4	No
			16.3.2014 21:00	21:00:00	21,6	61,8	21,6	61,1	No
			16.3.2014 21:15	21:15:00	21,6	62,1	21,6	61,6	No
			16.3.2014 21:30	21:30:00	21,7	62,1	21,6	61,9	No
			16.3.2014 21:45	21:45:00	21,7	62,2	21,7	62,1	No
			16.3.2014 22:00	22:00:00	21,8	62,8	21,7	62,3	Yes
			16.3.2014 22:15	22:15:00	21,8	62,7	21,8	62,5	Yes
			16.3.2014 22:30	22:30:00	21,8	62,5	21,8	62,6	Yes
			16.3.2014 22:45	22:45:00	21,9	62,6	21,8	62,7	Yes
			16.3.2014 23:00	23:00:00	21,9	62,7	21,9	62,6	Yes
			16.3.2014 23:15	23:15:00	21,9	60,9	21,9	62,2	Yes
			16.3.2014 23:30	23:30:00	22	61,7	21,9	62,0	Yes
			16.3.2014 23:45	23:45:00	22	62,1	22,0	61,9	Yes
			17.3.2014 0:00	0:00:00	22	62,5	22,0	61,8	Yes
			17.3.2014 0:15	0:15:00	22	62,3	22,0	62,2	Yes
			17.3.2014 0:30	0:30:00	22	62,3	22,0	62,3	Yes
			17.3.2014 0:45	0:45:00	22	62,7	22,0	62,5	Yes
			17.3.2014 1:00	1:00:00	22	62,7	22,0	62,5	Yes
			17.3.2014 1:15	1:15:00	22	63	22,0	62,7	Yes
			17.3.2014 1:30	1:30:00	22	63,1	22,0	62,9	Yes
			17.3.2014 1:45	1:45:00	22	63,1	22,0	63,0	Yes
			17.3.2014 2:00	2:00:00	22	63,3	22,0	63,1	Yes
			17.3.2014 2:15	2:15:00	22	63,4	22,0	63,2	Yes
			17.3.2014 2:30	2:30:00	22	63,6	22,0	63,4	Yes
			17.3.2014 2:45	2:45:00	21,9	63,7	22,0	63,5	Yes
			17.3.2014 3:00	3:00:00	22	63,8	22,0	63,6	Yes
			17.3.2014 3:15	3:15:00	21,9	63,9	22,0	63,8	Yes
			17.3.2014 3:30	3:30:00	22	64,1	22,0	63,9	Yes
			17.3.2014 3:45	3:45:00	22	64,3	22,0	64,0	Yes
			17.3.2014 4:00	4:00:00	22	64,5	22,0	64,2	Yes
			17.3.2014 4:15	4:15:00	22	64,4	22,0	64,3	Yes
			17.3.2014 4:30	4:30:00	22	64,6	22,0	64,5	Yes
			17.3.2014 4:45	4:45:00	22	64,3	22,0	64,5	Yes
			17.3.2014 5:00	5:00:00	22	64,6	22,0	64,5	Yes
			17.3.2014 5:15	5:15:00	22	64,4	22,0	64,5	Yes
			17.3.2014 5:30	5:30:00	22	64,2	22,0	64,4	Yes
			17.3.2014 5:45	5:45:00	22,1	57,8	22,0	62,8	Yes
			17.3.2014 6:00	6:00:00	22	60,3	22,0	61,7	Yes
			17.3.2014 6:15	6:15:00	22,3	61,7	22,1	61,0	Yes
			17.3.2014 6:30	6:30:00	22	52,2	22,1	58,0	Yes
			17.3.2014 6:45	6:45:00	21,9	56	22,1	57,6	Yes
			17.3.2014 7:00	7:00:00	22	50,4	22,1	55,1	Yes
			17.3.2014 7:15	7:15:00	21,9	52,8	22,0	52,9	Yes
			17.3.2014 7:30	7:30:00	22,1	49,4	22,0	52,2	Yes
			17.3.2014 7:45	7:45:00	22,2	57,2	22,1	52,5	Yes
			17.3.2014 8:00	8:00:00	22	47,8	22,1	51,8	Yes
			17.3.2014 8:15	8:15:00	22,2	57,1	22,1	52,9	Yes
			17.3.2014 8:30	8:30:00	22,3	50,8	22,2	53,2	Yes
			17.3.2014 8:45	8:45:00	22,2	55,7	22,2	52,9	Yes
			17.3.2014 9:00	9:00:00	22,4	51,8	22,3	53,9	Yes
			17.3.2014 9:15	9:15:00	22,4	49,9	22,3	52,1	Yes
			17.3.2014 9:30	9:30:00	22,5	49,7	22,4	51,8	Yes
			17.3.2014 9:45	9:45:00	22,4	54,4	22,4	51,5	Yes
			17.3.2014 10:00	10:00:00	22,7	53,7	22,5	51,9	Yes
			17.3.2014 10:15	10:15:00	22,5	55,1	22,5	53,2	Yes
			17.3.2014 10:30	10:30:00	22,7	53,3	22,6	54,1	Yes
			17.3.2014 10:45	10:45:00	22,4	48,7	22,6	52,7	Yes
			17.3.2014 11:00	11:00:00	22,5	53,7	22,5	52,7	Yes
			17.3.2014 11:15	11:15:00	22,7	53,3	22,6	52,3	Yes
			17.3.2014 11:30	11:30:00	22,3	46,6	22,5	50,6	Yes
			17.3.2014 11:45	11:45:00	22,3	50,5	22,5	51,0	Yes
			17.3.2014 12:00	12:00:00	22,2	46,2	22,4	49,2	Yes
			17.3.2014 12:15	12:15:00	22,1	45,3	22,2	47,2	Yes
			17.3.2014 12:30	12:30:00	22,5	46,8	22,3	47,2	Yes
			17.3.2014 12:45	12:45:00	22,2	42,5	22,3	45,2	Yes
			17.3.2014 13:00	13:00:00	22,4	48,8	22,3	45,9	Yes
			17.3.2014 13:15	13:15:00	22,3	44,5	22,4	45,7	Yes
			17.3.2014 13:30	13:30:00	22,3	44	22,3	45,0	Yes
			17.3.2014 13:45	13:45:00	22,5	46,2	22,4	45,9	Yes
			17.3.2014 14:00	14:00:00	22,2	42,3	22,3	44,3	Yes
			17.3.2014 14:15	14:15:00	22,5	46	22,4	44,6	Yes
			17.3.2014 14:30	14:30:00	22,5	42,5	22,4	44,3	Yes
			17.3.2014 14:45	14:45:00	22,5	47,8	22,4	44,7	No
			17.3.2014 15:00	15:00:00	22,2	43,4	22,4	44,9	No
			17.3.2014 15:15	15:15:00	22,4	51,2	22,4	46,2	No
			17.3.2014 15:30	15:30:00	22,6	51,7	22,4	48,5	No
			17.3.2014 15:45	15:45:00	22,7	52,8	22,5	49,8	No
			17.3.2014 16:00	16:00:00	22,7	53,6	22,6	52,3	No
			17.3.2014 16:15	16:15:00	22,5	54,2	22,6	53,1	No
			17.3.2014 16:30	16:30:00	22,6	53,2	22,6	53,5	No
			17.3.2014 16:45	16:45:00	22,6	53,6	22,6	53,7	No
			17.3.2014 17:00	17:00:00	22,6	53,9	22,6	53,7	No
			17.3.2014 17:15	17:15:00	22,6	54,1	22,6	53,7	No
			17.3.2014 17:30	17:30:00	22,5	53,8	22,6	53,9	No
			17.3.2014 17:45	17:45:00	22,5	54,5	22,6	54,1	No

PHI_14_10100
PHI_14_10111
PHI_14_10129

