



SPECIFICATIONS

DEFIBRILLATION ELECTRIC SHOCK		DATA BACKUP AND COMMUNICATION	
Waveform:	Biphasic Truncated Exponential (BTE) waveform	Standard SD card, IR communication port	
	(Impedance compensation)		
Energy:	Adult: 185 to 200J (±5%)	ACCESSORY SPECIFICATIONS	
	Paediatric: 50J (±5%)	ADULT/PAEDIATRIC	
Operating Mode:	Semi-Auto	Standby Life:	2 years from date of manufacture
Analysis/Charge Time:		Electrodes:	Disposable pads
Shock Time:	Delivers 200 shocks with full battery	Pad Life:	30 Months
		Placement:	Adult: Anterior-lateral
ECG			Paediatric: Anterior-posterior
Lead:	II (RA, LL)	Minimum active gel area: 80cm2 +/-5%	
Patient Impedance:	25 to 200 ohm	Cable Length:	Approx. 1.8m
Heart Rate:	20 to 300 per min	BATTERY	
Accuracy:	1 per min	Туре:	LiMnO2, disposable, long-life primary cell
Detection V/F	more than 200μV	Voltage/Capacity:	15V, 4200 mAh
V/T	more than 160 per min	Standby Life:	5 years from date of manufacture
Lead off detection:	Detected and voiced		(inserted in the AED)
Filter:	0.5 to 30Hz	Discharge:	A minimum of 200 shocks
INDICATION			(except the CPR period between the
INDICATION		-	defibrillation therapy) or 10 hours of
CONTROLS			operating time at 20°C
Standard:	Open (power) switch, shock button,	ENVIRONMENTAL CONDITIONS	
	Adult/Paediatric mode switch		
INDICATORS		Temperature:	Operation: 0 to 43°C (32 to 109.4°F)
Visible:	ICON Indicator, Status LCD (unit status, battery		Storage: 0 to 43°C (32 to 109.4°F)
	status, temperature status),	Relative Humidity:	5 to 95% RH (non-condensing)
	LED (Adult/Paediatric mode switch LED)	Altitude:	0 to 4,575m
Audible:	Speaker (voice prompt), Beep (CPR indication)	Shock:	Acceleration: 100G (+/-10%)
PHYSICAL		Vibration:	Frequency: 10Hz to 2000Hz
Dimensions:	294H x 240W x 95mmD		Acceleration: 10Hz to 100Hz: 5,0 (m/s²)²/Hz
Weight:	Approx. 2.65kg (including battery,		100Hz to 200Hz: -7 dB per octave
	excluding pads)		100Hz to 2000Hz: 1,0 (m/s²)²/Hz
		Drop Height:	1m
SELF TEST		Water and Dust Resi	stance: IP54 (Class 2)
Cycle:	Every 24 hours, 1 week, 1 month		
	Power on self test, battery insertion self test	Information correct at time of going to press. Subject to change without notice.	
Test Result:	Status LCD displays 'O' / 'X'		

www.medianadefib.co.uk

- t. +44 (0) 08456 448808
- e. sales@reliancemedical.co.uk
- w. www.medianadefib.co.uk



